

INSTRUMENT PROCESSING SHEET

Agency Southern County S/N 800000768
 Date In 9/21/17 Date Out 9/26/17 Ship P/U H/D CMI EE

Intake Performed By [Signature]

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
ok Case ok Handle
ok Dry Gas Holder ok Feet
ok Keyboard/Plug ok Back/Plugs
ok Screws tight ok Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By [Signature]

Lab Temp °C 21.9
 Breath Tube Screen
 Replace External O-Rings
 Instrument Set Up Verified
 R-Value 199
 Flow Verification (L/s)
 Flow Column # A7P103
 32mm 148 (.139 - .169)
 36mm 164 (.156 - .190)
 53mm 238 (.228 - .278)
 103mm 496 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 26932

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD3962	201707D 7/25/19
0.08	SD1013	201707E 7/25/19
0.20	DR23856	201707C 7/24/19
0.08 DGS	N/A	AG619605 7/14/18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # SEP
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Quality Checks Cont. Performed By [Signature]

Simulator Temperatures °C
 External Digital Therm. ID#: 300505
 34°C +/- .2 Serial #: SD3962
 34°C +/- .2 Serial #: SD1013
 34°C +/- .2 Serial #: DR23856

Calibration Adjustment Performed By _____

Calibration Adjustment N/A
 Calibration Adjustment Complete

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By [Signature]

Barometric Pressure 1012 Gauge
 ID# 26932 1010 Instrument

Mouth Alcohol Solution Lot # 2016 C
 Acetone Stock Solution Lot # 2017 A

Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	SD3962
0.08	SD1013
0.20	DR23856

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Calibration Adjustment
 Post-Stability Tests
 Other Form 40

Notes/Suggested Service:
Modern is fine. May need to hold phone
line in place while uploading for
a tyler connection.
N compliance
QA/QC/AS 9/26/17
 Quality Control Review _____ Date 9/26/17

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

80-000768
 Stability Checks
 9/26/17

INTOXILYZER 8000
 Instrument Initialization
 08:18 09/26/2017

SUWANNEE CO SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000768
 09/26/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:52
Control Test	0.079	08:52
Air Blank	0.000	08:53
Control Test	0.079	08:53
Air Blank	0.000	08:54
Control Test	0.079	08:54
Air Blank	0.000	08:55
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS



Operator's Signature

9/26/17

SUWANNEE CO SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000768
 09/26/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:02
Control Test	0.049	09:02
Air Blank	0.000	09:03
Control Test	0.049	09:04
Air Blank	0.000	09:04
Control Test	0.048	09:05
Air Blank	0.000	09:05
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	



Operator's Signature

9/26/17

SUWANNEE CO SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000768
 09/26/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:07
Control Test	0.197	09:07
Air Blank	0.000	09:08
Control Test	0.195	09:09
Air Blank	0.000	09:09
Control Test	0.196	09:10
Air Blank	0.000	09:10
Control Test Stats		
Average	0.1960	
Std Dev	0.0010	
Rel Std Dev(%)	0.5102	



Operator's Signature

INTOXILYZER 8000
Instrument Initialization
11:57 09/21/2017

SUWANNEE CO 50
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000768
09/21/2017
Software: 8100.27

***** Printer Test *****

40 characters per line
abcdefghijklmnopqrstuwxz1234567890-=:
ABCDEFGHIJKLMNopqRSTUUVWXYZ!@#%&^*()_+?

Double Char/line
abcdefghijklmnopqrst
1234567890-=:!@#%&^*
ABCDEFGHIJKLmNOPQRST

WIDE Char/line
abcdefghijklmnopqrst
1234567890-=:!@#%&^*
ABCDEFGHIJKLmNOPQRST

TALL Char/line
abcdefghijklmnopqrstuwxz1234567890-=:
ABCDEFGHIJKLMNopqRSTUUVWXYZ!@#%&^*()_+?

***** Test End *****

DiagnosTics
+
Printer Test

OK ✓

SUWANNEE CO 50
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000768
09/21/2017
Software: 8100.27

: DIAGNOSTICS :

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

9/21/17