



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Santa Rosa County

S/N 80-000772

Date In 10/2/17

Date Out 10/3/2017

Ship P/U H/D CMI EE

Intake Performed By JSB

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By JSB

Lab Temp °C 22.3
 Breath Tube Screen
 Replace External O-Rings
 Instrument Set Up Verified
 R-Value 212
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm .156 (.139 - .169)
 36mm .167 (.156 - .190)
 53mm .242 (.228 - .278)
 103mm .503 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 26932

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>SD3962</u>	<u>201707D</u> <u>7/25/19</u>
0.08	<u>SD1013</u>	<u>201707E</u> <u>7/25/19</u>
0.20	<u>DR3856</u>	<u>201707C</u> <u>7/24/19</u>
0.08 DGS	N/A	<u>AG708807</u> <u>3/29/19</u>

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Quality Checks Cont. Performed By JSB

Simulator Temperatures °C
 External Digital Therm. ID#: 208505
 34°C +/- .2 Serial #: SD3962
 34°C +/- .2 Serial #: SD1013
 34°C +/- .2 Serial #: DR3856

Calibration Adjustment Performed By _____

Calibration Adjustment N/A
 Calibration Adjustment Complete
 Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By JSB

Barometric Pressure 1019 Gauge
 ID# 26932 1020 Instrument

Mouth Alcohol Solution Lot # 2016-C
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	<u>SD1019</u>
Interferent	<u>SD1021</u>
0.05	<u>SD3962</u>
0.08	<u>SD1013</u>
0.20	<u>DR3856</u>

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Calibration Adjustment
 Post-Stability Tests
 Other _____

Notes/Suggested Service:

QA/QC OK SP

[Signature] 10/5/17
Quality Control Review Date

Instrument Complies with Chapter 11D-8, FAC

Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

SANTA ROSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000772
10/03/2017
Software: 8100.27

Test	g/210L	Time	
Air Blank	0.000	09:38	
Control Test	0.049	09:39	
Air Blank	0.000	09:39	
Control Test	0.050	09:40	
Air Blank	0.000	09:40	
Control Test	0.050	09:41	
Air Blank	0.000	09:41	
Control Test Stats			
Average	0.0497		
Std Dev	0.0006		
Rel Std Dev(%)	1.1625		

P Murphy
Operator's Signature

SANTA ROSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000772
10/03/2017
Software: 8100.27

Test	g/210L	Time	
Air Blank	0.000	09:43	
Control Test	0.082	09:43	
Air Blank	0.000	09:44	
Control Test	0.083	09:44	
Air Blank	0.000	09:45	
Control Test	0.082	09:46	
Air Blank	0.000	09:46	
Control Test Stats			
Average	0.0823		
Std Dev	0.0006		
Rel Std Dev(%)	0.7012		

P Murphy
Operator's Signature

SANTA ROSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000772
10/03/2017
Software: 8100.27

Test	g/210L	Time	
Air Blank	0.000	09:47	
Control Test	0.200	09:48	
Air Blank	0.000	09:48	
Control Test	0.201	09:49	
Air Blank	0.000	09:50	
Control Test	0.202	09:50	
Air Blank	0.000	09:51	
Control Test Stats			
Average	0.2010		
Std Dev	0.0010		
Rel Std Dev(%)	0.4975		

P Murphy
Operator's Signature

SP

SANTA ROSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000772
10/03/2017
Software: 8100.27

Test	g/210L	Time	
Air Blank	0.000	09:52	
Control Test	0.081	09:52	
Air Blank	0.000	09:53	
Control Test	0.082	09:53	
Air Blank	0.000	09:54	
Control Test	0.083	09:54	
Air Blank	0.000	09:54	
Control Test Stats			
Average	0.0820		
Std Dev	0.0010		
Rel Std Dev(%)	1.2195		

DGS

P Murphy
Operator's Signature

10/5/17
[Signature]