

## INSTRUMENT PROCESSING SHEET

Agency Lafayette County S/N 80-000773  
 Date In 7/28/17 Date Out 8/30/17  Ship  P/U  H/D  CMI  EE

**Intake** Performed By [Signature]

Registration  
 Annual  
 Return from CMI  
 Return from Enforcement Electronics  
 Other \_\_\_\_\_

Visual Inspection:  
 Case  Handle  
 Dry Gas Holder  Feet  
 Keyboard/Plug  Back/Plugs  
 Screws tight  Breath Hose

Other Equipment:  
 Power cord  
 Printer Cable  
 Other Static Bag

Notes: \_\_\_\_\_

**Quality Checks** Performed By [Signature]

Lab Temp °C 21.9  
 Breath Tube Screen  
 Replace External O-Rings  
 Instrument Set Up Verified  
 R-Value 210  
 Flow Verification (L/s)  
 Flow Column # A7P102  
 32mm 148 (.139 - .169)  
 36mm 164 (.156 - .190)  
 53mm 242 (.228 - .278)  
 103mm 499 (.447 - .547)

Barometric Pressure Check  
 Gauge ID # 26932

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>SD3962</u>	<u>201707D</u> <u>7/25/19</u>
0.08	<u>SD1013</u>	<u>201707E</u> <u>7/25/19</u>
0.20	<u>SD3933</u>	<u>201707C</u> <u>7/24/19</u>
0.08 DGS	N/A	<u>AG-619605</u> <u>7/14/18</u>

**Flow Calibration** Performed By \_\_\_\_\_

Flow Calibration N/A  
 Flow Calibration Complete  
 Flow Column # \_\_\_\_\_  
 5L/min - 17mm  
 15L/min - 53mm  
 30L/min - 103mm

R-Value \_\_\_\_\_  
 Post Calibration Verification (L/s)  
 Flow Column # \_\_\_\_\_  
 32mm \_\_\_\_\_ (.139 - .169)  
 36mm \_\_\_\_\_ (.156 - .190)  
 53mm \_\_\_\_\_ (.228 - .278)  
 103mm \_\_\_\_\_ (.447 - .547)

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Quality Checks Cont.** Performed By [Signature]  
**Simulator Temperatures °C**  
 External Digital Therm. ID#: 300505  
 34°C +/- .2 Serial #: SD3962  
 34°C +/- .2 Serial #: SD1013  
 34°C +/- .2 Serial #: SD3933

RECEIVED  
 AUG 30 2017  
 FDLE  
 Alcohol Testing Program

**Calibration Adjustment** Performed By \_\_\_\_\_

Calibration Adjustment N/A  
 Calibration Adjustment Complete

Barometric Pressure Gauge \_\_\_\_\_ ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

**Department Inspection** Performed By [Signature]

Barometric Pressure 1013 Gauge  
 ID# 26932 1013 Instrument

Mouth Alcohol Solution Lot # 2016-C  
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	<u>SD1019</u>
Interferent	<u>SD1021</u>
0.05	<u>SD3962</u>
0.08	<u>SD1013</u>
0.20	<u>SD3933</u>

**Attachments**

Form 41  Calibration Adjustment  
 Pre-Stability Tests  Post-Stability Tests  
 Flow Calibration  Other \_\_\_\_\_

Notes/Suggested Service:  
On/oc ok GPM 8/30/2017

[Signature] 8/30/17  
 Quality Control Review Date

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use

88-000773  
 Stability Checks  
 8/30/17

PPM

INTOXILYZER 8000  
 Instrument Initialization  
 07:39 08/30/2017

LAFAYETTE CO SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000  
 08/30/2017  
 Software: 8100.27  
 SN 80-000773

Test	g/210L	Time
Air Blank	0.000	08:40
Control Test	0.078	08:41
Air Blank	0.000	08:41
Control Test	0.079	08:41
Air Blank	0.000	08:42
Control Test	0.078	08:42
Air Blank	0.000	08:43
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

DS

Operator Signature

8/30/17

LAFAYETTE CO SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000  
 08/30/2017  
 Software: 8100.27  
 SN 80-000773

Test	g/210L	Time
Air Blank	0.000	08:46
Control Test	0.050	08:47
Air Blank	0.000	08:47
Control Test	0.050	08:48
Air Blank	0.000	08:49
Control Test	0.050	08:49
Air Blank	0.000	08:50
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator Signature

LAFAYETTE CO SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000  
 08/30/2017  
 Software: 8100.27  
 SN 80-000773

Test	g/210L	Time
Air Blank	0.000	08:51
Control Test	0.199	08:52
Air Blank	0.000	08:52
Control Test	0.199	08:53
Air Blank	0.000	08:53
Control Test	0.201	08:54
Air Blank	0.000	08:54
Control Test Stats		
Average	0.1997	
Std Dev	0.0012	
Rel Std Dev(%)	0.5783	

Operator Signature

LAFAYETTE CO SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000  
 08/30/2017  
 Software: 8100.27  
 SN 80-000773

Test	g/210L	Time
Air Blank	0.000	08:57
Control Test	0.081	08:57
Air Blank	0.000	08:58
Control Test	0.080	08:59
Air Blank	0.000	08:59
Control Test	0.081	09:00
Air Blank	0.000	09:00
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

Operator Signature