

Agency Lafayette County

S/N 80-000776

MAR 28 2017

Date In 3/19/17

Date Out 3/28/17

Ship P/U H/D CMI FDLE

Alcohol Testing Program

<p>Intake Performed By _____</p> <p><input type="checkbox"/> Registration</p> <p><input checked="" type="checkbox"/> Annual</p> <p><input type="checkbox"/> Return from CMI</p> <p><input type="checkbox"/> Return from Enforcement Electronics</p> <p><input type="checkbox"/> Other _____</p> <p>Visual Inspection:</p> <p><u>ok</u> Case <u>ok</u> Handle</p> <p><u>ok</u> Dry Gas Holder <u>ok</u> Feet</p> <p><u>ok</u> Keyboard/Plug <u>ok</u> Back/Plugs</p> <p><u>ok</u> Screws tight <u>ok</u> Breath Hose</p> <p>Other Equipment:</p> <p><input checked="" type="checkbox"/> Power cord</p> <p><input type="checkbox"/> Printer Cable</p> <p><input checked="" type="checkbox"/> Other <u>Static Bag</u></p> <p>Notes: _____</p>	<p>Quality Checks Performed By <u>②</u></p> <p><input checked="" type="checkbox"/> Breath Tube Screen</p> <p><input checked="" type="checkbox"/> Replace O-Rings</p> <p><input checked="" type="checkbox"/> Instrument Set Up Verified</p> <p><input checked="" type="checkbox"/> R-Value <u>198</u></p> <p><input checked="" type="checkbox"/> Flow Verification (L/s)</p> <p>Flow Column # <u>ATP105</u></p> <p>32mm <u>149</u> (.139 - .169)</p> <p>36mm <u>163</u> (.156 - .190)</p> <p>53mm <u>241</u> (.228 - .278)</p> <p>103mm <u>498</u> (.447 - .547)</p> <p><input checked="" type="checkbox"/> Barometric Pressure Check</p> <p>Gauge ID # <u>28421</u></p> <p><input checked="" type="checkbox"/> Stability Checks</p> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td><u>DR2035</u></td> <td><u>201603D</u> <u>3/18/18</u></td> </tr> <tr> <td>0.08</td> <td><u>SD3964</u></td> <td><u>201601F</u> <u>1/26/18</u></td> </tr> <tr> <td>0.20</td> <td><u>SD1025</u></td> <td><u>201601C</u> <u>1/5/18</u></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td><u>A6626604</u> <u>9/22/18</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	<u>DR2035</u>	<u>201603D</u> <u>3/18/18</u>	0.08	<u>SD3964</u>	<u>201601F</u> <u>1/26/18</u>	0.20	<u>SD1025</u>	<u>201601C</u> <u>1/5/18</u>	0.08 DGS	N/A	<u>A6626604</u> <u>9/22/18</u>	<p>Flow Calibration Performed By _____</p> <p><input checked="" type="checkbox"/> Flow Calibration N/A</p> <p><input type="checkbox"/> Flow Calibration Complete</p> <p>Flow Column # _____</p> <p><input type="checkbox"/> 5L/min - 17mm</p> <p><input type="checkbox"/> 15L/min - 53mm</p> <p><input type="checkbox"/> 30L/min - 103mm</p> <p><input type="checkbox"/> R-Value _____</p> <p><input type="checkbox"/> Post Calibration Verification (L/s)</p> <p>Flow Column # _____</p> <p>32mm _____ (.139 - .169)</p> <p>36mm _____ (.156 - .190)</p> <p>53mm _____ (.228 - .278)</p> <p>103mm _____ (.447 - .547)</p> <p>Maintenance Performed By _____</p> <p><input type="checkbox"/> Battery Replacement</p> <p><input type="checkbox"/> Dry Gas Regulator Replacement</p> <p><input type="checkbox"/> Breath Tube Replacement</p> <p><input type="checkbox"/> Other _____</p> <p>Suggested Service</p> <p>_____</p> <p>_____</p>
Simulator	Serial #	Lot #/Exp															
0.05	<u>DR2035</u>	<u>201603D</u> <u>3/18/18</u>															
0.08	<u>SD3964</u>	<u>201601F</u> <u>1/26/18</u>															
0.20	<u>SD1025</u>	<u>201601C</u> <u>1/5/18</u>															
0.08 DGS	N/A	<u>A6626604</u> <u>9/22/18</u>															

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A

Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By ②

Barometric Pressure

ID# 28421

Gauge 1014

Instrument 1015

Mouth Alcohol Solution Lot # 2016-A

Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	<u>SD1019</u>
Interferent	<u>SD1021</u>
0.05	<u>DR2035</u>
0.08	<u>SD3964</u>
0.20	<u>SD1025</u>

Attachments

Form 41

Pre-Stability Tests

Flow Calibration

Optical Bench Cal

Post-Stability Tests

Other _____

Notes: DN/OC OK @ pm 3/28/2017

Instrument Complies with Chapter 11D-8, FAC

Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use

Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Brett K. ...

Quality Control Review

3/28/17

Date

80-000776
 Stability Checks
 3/28/17

INTOXILYZER 8000
 Instrument Initialization
 07:32 03/28/2017

LAFAYETTE CO SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000776
 03/28/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:05
Control Test	0.050	08:05
Air Blank	0.000	08:06
Control Test	0.050	08:07
Air Blank	0.000	08:07
Control Test	0.050	08:08
Air Blank	0.000	08:08
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator Signature

LAFAYETTE CO SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000776
 03/28/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:09
Control Test	0.080	08:09
Air Blank	0.000	08:10
Control Test	0.080	08:10
Air Blank	0.000	08:11
Control Test	0.080	08:11
Air Blank	0.000	08:11
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

Operator Signature

LAFAYETTE CO SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000776
 03/28/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:14
Control Test	0.201	08:14
Air Blank	0.000	08:15
Control Test	0.202	08:16
Air Blank	0.000	08:16
Control Test	0.203	08:17
Air Blank	0.000	08:17
Control Test Stats		
Average	0.2020	
Std Dev	0.0010	
Rel Std Dev(%)	0.4950	

Operator Signature

DGS

Operator's Signature

Operator's Signature