



Alcohol Testing Program

### INSTRUMENT PROCESSING SHEET

Agency Suwannee County S.C. S/N 80-000780

Date In 12/27/16 Date Out 2/2/17  Ship  P/U  H/D  CMI  EE

**Intake** Performed By [Signature]

Registration  
 Annual  
 Return from CMI  
 Return from Enforcement Electronics  
 Other \_\_\_\_\_

Visual Inspection:  
OK Case OK Handle  
OK Dry Gas Holder OK Feet  
OK Keyboard/Plug OK Back/Plugs  
OK Screws tight OK Breath Hose

Other Equipment:  
 Power cord  
 Printer Cable  
 Other \_\_\_\_\_

Notes: \_\_\_\_\_

**Quality Checks** Performed By [Signature]

Breath Tube Screen  
 Replace O-Rings  
 Instrument Set Up Verified  
 R-Value 230  
 Flow Verification (L/s)  
 Flow Column # ATP102  
 32mm 160 (.139 - .169)  
 36mm 181 (.156 - .190)  
 53mm 262 (.228 - .278)  
 103mm 504 (.447 - .547)

Barometric Pressure Check  
 Gauge ID # 28421

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	G11739	201603 D 318118
0.08	G8149	201601 F 1/26/18
0.20	G11621	201604 C 4/5/18
0.08 DGS	N/A	AG-619605 7/1/18

**Flow Calibration** Performed By \_\_\_\_\_

Flow Calibration N/A  
 Flow Calibration Complete  
 Flow Column # \_\_\_\_\_  
 5L/min - 17mm  
 15L/min - 53mm  
 30L/min - 103mm

R-Value \_\_\_\_\_  
 Post Calibration Verification  
 Flow Column # \_\_\_\_\_  
 32mm \_\_\_\_\_ (.139 - .169)  
 36mm \_\_\_\_\_ (.156 - .190)  
 53mm \_\_\_\_\_ (.228 - .278)  
 103mm \_\_\_\_\_ (.447 - .547)

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Suggested Service**

**RECEIVED**  
FEB 02 2017  
FDLE  
Alcohol Testing Program

**Optical Bench Calibration** Performed By \_\_\_\_\_

Optical Bench Calibration N/A  
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

**Department Inspection** Performed By [Signature]

Barometric Pressure 1020 Gauge  
 ID# 28421 1019 Instrument

Mouth Alcohol Solution Lot # 2016-A  
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	G2879
Interferent	G8144
0.05	G11739
0.08	<del>G8149</del> G8149
0.20	G11621

**Attachments**

Form 41  Optical Bench Cal  
 Pre-Stability Tests  Post-Stability Tests  
 Flow Calibration  Other Form 40

Notes: QC DGS

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use

Brett Highland

Quality Control Review

2/2/17  
Date

80-000 780  
 Stable by check  
 2/2/17

INTOXILYZER 8000  
 Instrument Initialization  
 08:24 02/02/2017

SUWANNEE CO SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000780  
 02/02/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:17
Control Test	0.078	09:18
Air Blank	0.000	09:18
Control Test	0.079	09:19
Air Blank	0.000	09:19
Control Test	0.079	09:20
Air Blank	0.000	09:21
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

*DS*

Operator's Signature

SUWANNEE CO SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000780  
 02/02/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:11
Control Test	0.078	09:12
Air Blank	0.000	09:13
Control Test	0.079	09:13
Air Blank	0.000	09:14
Control Test	0.079	09:15
Air Blank	0.000	09:15
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

Operator's Signature

SUWANNEE CO SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000780  
 02/02/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:06
Control Test	0.198	09:07
Air Blank	0.000	09:08
Control Test	0.198	09:08
Air Blank	0.000	09:09
Control Test	0.198	09:10
Air Blank	0.000	09:10
Control Test Stats		
Average	0.1980	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

*BK*  
*ALDS*

Operator's Signature

Operator's Signature