

WASHINGTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000781
10/27/2017
Software: 8100.27

Test	g/210L	Time	
Air Blank	0.000	09:52	
Control Test	0.049	09:52	
Air Blank	0.000	09:53	
Control Test	0.050	09:53	
Air Blank	0.000	09:54	
Control Test	0.050	09:55	
Air Blank	0.000	09:55	
Control Test Stats			
Average	0.0497		
Std Dev	0.0006		
Rel Std Dev(%)	1.1625		

P Murphy
Operator's Signature

WASHINGTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000781
10/27/2017
Software: 8100.27

Test	g/210L	Time	
Air Blank	0.000	10:03	
Control Test	0.081	10:04	
Air Blank	0.000	10:04	
Control Test	0.080	10:05	
Air Blank	0.000	10:05	
Control Test	0.080	10:06	
Air Blank	0.000	10:07	
Control Test Stats			
Average	0.0803		
Std Dev	0.0006		
Rel Std Dev(%)	0.7187		

P Murphy
Operator's Signature

WASHINGTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000781
10/27/2017
Software: 8100.27

Test	g/210L	Time	
Air Blank	0.000	09:58	
Control Test	0.201	09:59	
Air Blank	0.000	09:59	
Control Test	0.201	10:00	
Air Blank	0.000	10:00	
Control Test	0.201	10:01	
Air Blank	0.000	10:02	
Control Test Stats			
Average	0.2010		
Std Dev	0.0000		
Rel Std Dev(%)	0.0000		

P Murphy
Operator's Signature

WASHINGTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000781
10/27/2017
Software: 8100.27

Test	g/210L	Time	
Air Blank	0.000	10:08	
Control Test	0.077	10:09	
Air Blank	0.000	10:09	
Control Test	0.077	10:09	
Air Blank	0.000	10:10	
Control Test	0.078	10:10	
Air Blank	0.000	10:11	
Control Test Stats			
Average	0.0773		
Std Dev	0.0006		
Rel Std Dev(%)	0.7466		

D65

P Murphy
Operator's Signature

WASHINGTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000781
10/27/2017
Software: 8100.27

Flow Rate Calibration*****
1: Rate (Liters/min) = 5
SQRT(Diff)) = 4.000
2: Rate (Liters/min) = 15
SQRT(Diff)) = 9.164
3: Rate (Liters/min) = 30
SQRT(Diff)) = 19.746
Dependent Data Scale Factor = 100000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 610
Rounded Intercept = -47352
Correlation = 0.99671

Pgan

11/15/17
JD