



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Washington County

S/N 80-000782

Date In 7/25/17

Date Out 7/28/2017

Ship P/U H/D CMI EE

RECEIVED
AUG 01 2017
FDLE
Alcohol Testing Program

Intake Performed By [Signature]

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By [Signature]

Lab Temp °C 20.5
 Breath Tube Screen
 Replace External O-Rings
 Instrument Set Up Verified
 R-Value 211
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm .148 (.139 - .169)
 36mm .167 (.156 - .190)
 53mm .230 (.228 - .278)
 103mm .503 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 26932

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>62078</u>	<u>201603D</u> <u>3/8/18</u>
0.08	<u>68149</u>	<u>201611B</u> <u>11/15/18</u>
0.20	<u>611621</u>	<u>201702B</u> <u>2/23/19</u>
0.08 DGS	N/A	<u>AG619605</u> <u>7/14/18</u>

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Quality Checks Cont. Performed By [Signature]

Simulator Temperatures °C
 External Digital Therm. ID#: 300504
 34°C +/- .2 Serial #: 62078
 34°C +/- .2 Serial #: 68149
 34°C +/- .2 Serial #: 611621

Calibration Adjustment Performed By _____

Calibration Adjustment N/A
 Calibration Adjustment Complete
 Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By [Signature]

Barometric Pressure 1014 Gauge
 ID# 26932 1013 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	<u>62879</u>
Interferent	<u>68144</u>
0.05	<u>62078</u>
0.08	<u>68149</u>
0.20	<u>611621</u>

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Calibration Adjustment
 Post-Stability Tests
 Other _____

Notes/Suggested Service:

[Signature] 8/1/17 [Signature] 8/1/17
 Quality Control Review Date

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

WASHINGTON COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000782
 07/28/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:58
Control Test	0.049	07:58
Air Blank	0.000	07:59
Control Test	0.049	08:00
Air Blank	0.000	08:00
Control Test	0.050	08:01
Air Blank	0.000	08:01
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

P Murphy
 Operator's Signature

WASHINGTON COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000782
 07/28/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:50
Control Test	0.079	07:51
Air Blank	0.000	07:51
Control Test	0.079	07:52
Air Blank	0.000	07:53
Control Test	0.080	07:53
Air Blank	0.000	07:54
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

P Murphy
 Operator's Signature

WASHINGTON COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000782
 07/28/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:03
Control Test	0.198	08:04
Air Blank	0.000	08:04
Control Test	0.197	08:05
Air Blank	0.000	08:05
Control Test	0.198	08:06
Air Blank	0.000	08:07
Control Test Stats		
Average	0.1977	
Std Dev	0.0006	
Rel Std Dev(%)	0.2921	

P Murphy
 Operator's Signature

WASHINGTON COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000782
 07/28/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:12
Control Test	0.080	08:13
Air Blank	0.000	08:13
Control Test	0.079	08:13
Air Blank	0.000	08:14
Control Test	0.080	08:14
Air Blank	0.000	08:15
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

DGS

P Murphy
 Operator's Signature

JE