



INSTRUMENT PROCESSING SHEET

Agency Gulf Breeze PDS/N 80-000787

Florida Department of Law Enforcement

Date In 10/26/17DI Completion Date 10/26/17 Ship P/U H/D CMI EEIntake Performed By [Signature]Quality Checks Performed By [Signature]Flow Calibration Performed By [Signature]**RECEIVED**
OCT 26 2017FDLE
Alcohol Testing

- Annual
- Registration
- Return from CMI / EE

Visual Inspection:

- Case Handle
- Keyboard Dry Gas Shelf
- Feet Breath Tube
- Ports Screws Tight

Other Equipment/ Accessories:

- Power cord Printer Cable
- Static Bag 12V DC Cable

Notes: _____

- Breath Tube Screen
- Replace External O-Rings
- Instrument Set Up Verified
- R-Value 119
- Flow Verification (L/s)
- Flow Column # ATP105
- 32 mm .156 (.139 - .169)
- 36 mm .171 (.156 - .190)
- 53 mm .242 (.228 - .278)
- 103 mm .515 (.447 - .547)
- Barometric Pressure Check
- Gauge ID # 28421
- Stability Checks

- Flow Column # _____
- 5L/min - 17mm
- 15L/min - 53mm
- 30L/min - 103mm
- R-Value _____
- Post Calibration Verification (L/s)
- Flow Column # _____
- 32 mm _____ (.139 - .169)
- 36 mm _____ (.156 - .190)
- 53 mm _____ (.228 - .278)
- 103 mm _____ (.447 - .547)

Maintenance Performed By _____

- Battery Replacement
- Dry Gas Regulator Replacement
- Breath Tube Replacement
- Other _____

Temperature Checks Performed By [Signature]

- Lab Temp °C 22.2
- External Digital Therm. ID#: 300504
- 34°C +/- .2 Serial #: 62835
- 34°C +/- .2 Serial #: DR1279
- 34°C +/- .2 Serial #: SD1025

Final Release Date

RECEIVED

OCT 26 2017

FDLE
Alcohol Testing

Simulator	Serial #	Lot #/Exp
0.050	<u>62835</u>	<u>201707D</u> <u>7/25/19</u>
0.080	<u>DR1279</u>	<u>201707E</u> <u>7/25/19</u>
0.200	<u>SD1025</u>	<u>201707C</u> <u>7/24/19</u>
0.080 DGS	N/A	<u>AG626605</u> <u>9/22/18</u>

Calibration Adjustment Performed By _____

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

 Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By [Signature]Barometric Pressure ID# 28421Gauge 2841 1016 Instrument 1014Mouth Alcohol Solution Lot # 2016-CAcetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.000	<u>62880</u>
Interferent	<u>62840</u>
0.050	<u>62835</u>
0.080	<u>DR1279</u>
0.200	<u>SD1025</u>

Attachments

- Form 41
- Stability Checks
- Calibration Adjustment
- Post-Stability Checks
- Flow Calibration
- Form 40
- Other _____

Notes/Suggested Service: _____

- Instrument Complies with Chapter 11D-8, FAC
- Instrument Does Not Comply with Chapter 11D-8, FAC
- Return to/Place into Evidentiary Use
- Remain Out of Evidentiary Use
- Conduct an Agency Inspection Before Evidentiary Use

[Signature] 10/26/17 [Signature] 10/26/17
Tech Review / Date Admin Review / Date

GULF BREEZE PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000787
 10/26/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:30
Control Test	0.049	09:30
Air Blank	0.000	09:31
Control Test	0.049	09:32
Air Blank	0.000	09:32
Control Test	0.049	09:33
Air Blank	0.000	09:33
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P Murphy
 Operator's Signature

GULF BREEZE PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000787
 10/26/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:35
Control Test	0.079	09:35
Air Blank	0.000	09:36
Control Test	0.079	09:36
Air Blank	0.000	09:37
Control Test	0.079	09:38
Air Blank	0.000	09:38
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P Murphy
 Operator's Signature

GULF BREEZE PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000787
 10/26/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:40
Control Test	0.200	09:41
Air Blank	0.000	09:41
Control Test	0.201	09:42
Air Blank	0.000	09:42
Control Test	0.201	09:43
Air Blank	0.000	09:44
Control Test Stats		
Average	0.2007	
Std Dev	0.0006	
Rel Std Dev(%)	0.2877	

P Murphy
 Operator's Signature

AWB

GULF BREEZE PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000787
 10/26/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:59
Control Test	0.081	10:00
Air Blank	0.000	10:00
Control Test	0.082	10:00
Air Blank	0.000	10:01
Control Test	0.081	10:01
Air Blank	0.000	10:02
Control Test Stats		
Average	0.0813	
Std Dev	0.0006	
Rel Std Dev(%)	0.7099	

DGS

P Murphy
 Operator's Signature

10/26/17
ee