



INSTRUMENT PROCESSING SHEET

Agency Gulf County S/N 80-000795

Florida Department of Law Enforcement

Date In 10/27/17 DI Completion Date 11/3/17 Ship P/U H/D CMI EE

Intake Performed By <u>[Signature]</u>	Quality Checks Performed By <u>[Signature]</u>	Flow Calibration Performed By _____
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input type="checkbox"/> Dry Gas Shelf <input type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable	<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>238</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP105</u> 32 mm <u>.148</u> (.139 - .169) 36 mm <u>.166</u> (.156 - .190) 53 mm <u>.243</u> (.228 - .278) 103 mm <u>.503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28421</u>	Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)

Notes: Dry Gas shipped w/ inst.
Dry gas shelf not attached. 2
feet missing. 4 feet and 4
users loose in static Bag.

Final Release Date

NOV 06 2017

Stability Checks		
Simulator	Serial #	Lot #/Exp
0.050	G2835	201707D 7/25/19
0.080	DR1279	201707E 7/25/19
0.200	SD1025	201707C 7/24/19
0.080 DGS	N/A	AG626605 9/22/18

Maintenance Performed By _____
<input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____
Temperature Checks Performed By <u>[Signature]</u>
<input checked="" type="checkbox"/> Lab Temp °C <u>22.3</u> External Digital Therm. ID#: <u>300504</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>G2835</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>DR1279</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>SD1025</u>

Calibration Adjustment Performed By _____																												
Barometric Pressure Gauge _____ ID # _____																												
<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A		
Simulator	Serial Number	Lot Number	Expiration																									
0.000		N/A	N/A																									
0.040																												
0.100																												
0.200																												
0.300																												
0.080 DGS	N/A																											
<input type="checkbox"/> Post Calibration Adjustment Stability Checks																												
<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A										
Simulator	Serial Number	Lot Number	Expiration																									
0.050																												
0.080																												
0.200																												
0.080 DGS	N/A																											

Department Inspection Performed By <u>[Signature]</u>												
Barometric Pressure ID# <u>28421</u>												
Gauge <u>1019</u> Instrument <u>1021</u>												
Mouth Alcohol Solution Lot # <u>2016-C</u>												
Acetone Stock Solution Lot # <u>2017-A</u>												
<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G2880</td> </tr> <tr> <td>Interferent</td> <td>G2840</td> </tr> <tr> <td>0.050</td> <td>G2835</td> </tr> <tr> <td>0.080</td> <td>DR1279</td> </tr> <tr> <td>0.200</td> <td>SD1025</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	G2880	Interferent	G2840	0.050	G2835	0.080	DR1279	0.200	SD1025
Simulator	Serial Number											
0.000	G2880											
Interferent	G2840											
0.050	G2835											
0.080	DR1279											
0.200	SD1025											
Attachments												
<input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____												

Notes/Suggested Service: _____

<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use
<u>[Signature]</u> 11/6/17 Tech Review / Date
<u>[Signature]</u> 11/6/17 Admin Review / Date

GULF COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000795
11/02/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:26
Control Test	0.049	13:27
Air Blank	0.000	13:27
Control Test	0.049	13:28
Air Blank	0.000	13:29
Control Test	0.050	13:29
Air Blank	0.000	13:30
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

P Murphy

Operator's Signature

GULF COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000795
11/02/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:20
Control Test	0.080	13:21
Air Blank	0.000	13:22
Control Test	0.080	13:22
Air Blank	0.000	13:23
Control Test	0.080	13:23
Air Blank	0.000	13:24
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P Murphy

Operator's Signature

GULF COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000795
11/02/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:15
Control Test	0.199	13:16
Air Blank	0.000	13:17
Control Test	0.201	13:17
Air Blank	0.000	13:18
Control Test	0.201	13:19
Air Blank	0.000	13:19
Control Test Stats		
Average	0.2003	
Std Dev	0.0012	
Rel Std Dev(%)	0.5764	

P Murphy

Operator's Signature

GULF COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000795
11/02/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:58
Control Test	0.079	13:58
Air Blank	0.000	13:59
Control Test	0.079	13:59
Air Blank	0.000	14:00
Control Test	0.078	14:00
Air Blank	0.000	14:00
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

DGS

P Murphy

Operator's Signature

AMS

11/6/17
JL