



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Dixie County

S/N 80-000809

Date In 1/25/17

Date Out 2/8/17

Ship  P/U  H/D  CMI  EE

**Intake** Performed By [Signature]

Registration  
 Annual  
 Return from CMI  
 Return from Enforcement Electronics  
 Other \_\_\_\_\_

Visual Inspection:  
 Case  Handle  
 Dry Gas Holder  Feet  
 Keyboard/Plug  Back/Plugs  
 Screws tight  Breath Hose

Other Equipment:  
 Power cord  
 Printer Cable  
 Other Static Bag

Notes: \_\_\_\_\_

**Quality Checks** Performed By [Signature]

Breath Tube Screen  
 Replace O-Rings  
 Instrument Set Up Verified  
 R-Value 200  
 Flow Verification (L/s)  
 Flow Column # ATP102  
 32mm 151 (.139 - .169)  
 36mm 177 (.156 - .190)  
 53mm 264 (.228 - .278)  
 103mm 507 (.447 - .547)

Barometric Pressure Check  
 Gauge ID # 28421

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	G11739	201603D 3/8/18
0.08	G8149	201601F 1/24/18
0.20	G11621	201604C 4/5/18
0.08 DGS	N/A	AGG 26605 9/22/18

**Flow Calibration** Performed By [Signature]

Flow Calibration N/A  
 Flow Calibration Complete  
 Flow Column # \_\_\_\_\_  
 5L/min - 17mm  
 15L/min - 53mm  
 30L/min - 103mm

R-Value \_\_\_\_\_  
 Post Calibration Verification (L/s)  
 Flow Column # \_\_\_\_\_  
 32mm \_\_\_\_\_ (.139 - .169)  
 36mm \_\_\_\_\_ (.156 - .190)  
 53mm \_\_\_\_\_ (.228 - .278)  
 103mm \_\_\_\_\_ (.447 - .547)

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Suggested Service**

RECEIVED  
FEB 10 2017  
FDLE  
Alcohol Testing Program

**Optical Bench Calibration** Performed By \_\_\_\_\_

Optical Bench Calibration N/A  
 Optical Bench Calibration Complete  
 Barometric Pressure Gauge \_\_\_\_\_ ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

**Department Inspection** Performed By [Signature]

Barometric Pressure 1009 Gauge  
 ID# 28421 1010 Instrument

Mouth Alcohol Solution Lot # 2016-C  
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	G2879
Interferent	G8144
0.05	G-11739
0.08	G8149
0.20	G-11621

**Attachments**

Form 41  
 Pre-Stability Tests  
 Flow Calibration

Optical Bench Cal  
 Post-Stability Tests  
 Other \_\_\_\_\_

Notes: Printer communication issue. Will not print to multiple external printers. Will only attempt to print to internal and error out. Recommend it being sent back to CMI since it just came back from there.

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use

Quality Control Review

Date

80-000809  
 Strab. Hwy. devices  
 2/8/17  
 RK

INTOXILYZER 8000  
 Instrument Initialization  
 07:18 02/08/2017

DIXIE COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000809  
 02/08/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:56
Control Test	0.048	08:57
Air Blank	0.000	08:57
Control Test	0.049	08:58
Air Blank	0.000	08:59
Control Test	0.049	08:59
Air Blank	0.000	09:00
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1653	

Operator's Signature

DIXIE COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000809  
 02/08/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:02
Control Test	0.079	09:02
Air Blank	0.000	09:03
Control Test	0.079	09:04
Air Blank	0.000	09:04
Control Test	0.079	09:05
Air Blank	0.000	09:05
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

DIXIE COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000809  
 02/08/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:06
Control Test	0.199	09:07
Air Blank	0.000	09:08
Control Test	0.200	09:08
Air Blank	0.000	09:09
Control Test	0.200	09:09
Air Blank	0.000	09:10
Control Test Stats		
Average	0.1997	
Std Dev	0.0006	
Rel Std Dev(%)	0.2892	

Operator's Signature

SR

DIXIE COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000809  
 02/08/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:47
Control Test	0.080	08:48
Air Blank	0.000	08:48
Control Test	0.080	08:48
Air Blank	0.000	08:48
Control Test	0.080	08:49
Air Blank	0.000	08:49
Control Test	0.000	08:50
Control Test Stats		
Average	0.0600	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

Operator's Signature