



Alcohol Testing Program

### INSTRUMENT PROCESSING SHEET

Agency Dixie County

S/N 80-000810

Date In 3/9/17

Date Out 3/16/17

Ship  P/U  H/D  CMI  EE

**Intake** Performed By TD

Registration  
 Annual  
 Return from CMI  
 Return from Enforcement Electronics  
 Other \_\_\_\_\_

Visual Inspection:  
ok Case ok Handle  
ok Dry Gas Holder ok Feet  
ok Keyboard/Plug ok Back/Plugs  
ok Screws tight ok Breath Hose

Other Equipment:  
 Power cord  
 Printer Cable  
 Other Static Bag

Notes: \_\_\_\_\_

**Quality Checks** Performed By Q

Breath Tube Screen  
 Replace O-Rings  
 Instrument Set Up Verified  
 R-Value 165  
 Flow Verification (L/s)  
 Flow Column # ATP105  
 32mm 144 (.139 - .169)  
 36mm 160 (.156 - .190)  
 53mm 241 (.228 - .278)  
 103mm 493 (.447 - .547)

Barometric Pressure Check  
 Gauge ID # 26932

Stability Checks

| Simulator | Serial #       | Lot #/Exp                         |
|-----------|----------------|-----------------------------------|
| 0.05      | <u>DR 2035</u> | <u>201603D</u><br><u>3/8/18</u>   |
| 0.08      | <u>SD3964</u>  | <u>201601F</u><br><u>1/26/18</u>  |
| 0.20      | <u>DR3855</u>  | <u>201604C</u><br><u>4/5/18</u>   |
| 0.08 DGS  | N/A            | <u>AB626604</u><br><u>9/22/18</u> |

**Flow Calibration** Performed By \_\_\_\_\_

Flow Calibration N/A  
 Flow Calibration Complete  
 Flow Column # \_\_\_\_\_  
 5L/min - 17mm  
 15L/min - 53mm  
 30L/min - 103mm

R-Value \_\_\_\_\_  
 Post Calibration Verification (L/s)  
 Flow Column # \_\_\_\_\_  
 32mm \_\_\_\_\_ (.139 - .169)  
 36mm \_\_\_\_\_ (.156 - .190)  
 53mm \_\_\_\_\_ (.228 - .278)  
 103mm \_\_\_\_\_ (.447 - .547)

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Suggested Service**

RECEIVED  
MAR 16 2017  
ALCOHOL TESTING PROGRAM

**Optical Bench Calibration** Performed By \_\_\_\_\_

Optical Bench Calibration N/A  
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # \_\_\_\_\_

| Simulator | Serial Number | Lot Number | Expiration |
|-----------|---------------|------------|------------|
| 0.000     |               | N/A        | N/A        |
| 0.040     |               |            |            |
| 0.100     |               |            |            |
| 0.200     |               |            |            |
| 0.400     |               |            |            |
| 0.080 DGS | N/A           |            |            |

Post Calibration Stability Checks

| Simulator | Serial Number | Lot Number | Expiration |
|-----------|---------------|------------|------------|
| 0.05      |               |            |            |
| 0.08      |               |            |            |
| 0.20      |               |            |            |
| 0.08 DGS  | N/A           |            |            |

**Department Inspection** Performed By Q

Barometric Pressure ID# 26932 Gauge 1028 Instrument 1026

Mouth Alcohol Solution Lot # 2016-A  
 Acetone Stock Solution Lot # 2017-A

| Simulator   | Serial Number |
|-------------|---------------|
| 0.00        | <u>SD1019</u> |
| Interferent | <u>SD1021</u> |
| 0.05        | <u>DR2035</u> |
| 0.08        | <u>SD3964</u> |
| 0.20        | <u>DR3855</u> |

**Attachments**

Form 41  
 Pre-Stability Tests  
 Flow Calibration

Optical Bench Cal  
 Post-Stability Tests  
 Other \_\_\_\_\_

Notes: QC KDS

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use

Brett Kvietland

Quality Control Review

Date

80-000810  
 State 1.4 checks  
 3/16/17

INTOXILYZER 8000  
 Instrument Initialization  
 07:26 03/16/2017

DIXIE COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000810  
 03/16/2017  
 Software: 8100.27

| Test               | 9/210L | Time  |
|--------------------|--------|-------|
| Air Blank          | 0.000  | 08:09 |
| Control Test       | 0.197  | 08:10 |
| Air Blank          | 0.000  | 08:10 |
| Control Test       | 0.196  | 08:11 |
| Air Blank          | 0.000  | 08:11 |
| Control Test       | 0.196  | 08:12 |
| Air Blank          | 0.000  | 08:13 |
| Control Test Stats |        |       |
| Average            | 0.1963 |       |
| Std Dev            | 0.0006 |       |
| Rel Std Dev(%)     | 0.2941 |       |

DIXIE COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000810  
 03/16/2017  
 Software: 8100.27

| Test               | 9/210L | Time  |
|--------------------|--------|-------|
| Air Blank          | 0.000  | 07:59 |
| Control Test       | 0.050  | 08:00 |
| Air Blank          | 0.000  | 08:00 |
| Control Test       | 0.049  | 08:01 |
| Air Blank          | 0.000  | 08:02 |
| Control Test       | 0.050  | 08:02 |
| Air Blank          | 0.000  | 08:03 |
| Control Test Stats |        |       |
| Average            | 0.0497 |       |
| Std Dev            | 0.0006 |       |
| Rel Std Dev(%)     | 1.1625 |       |

DIXIE COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000810  
 03/16/2017  
 Software: 8100.27

| Test               | 9/210L | Time  |
|--------------------|--------|-------|
| Air Blank          | 0.000  | 07:55 |
| Control Test       | 0.078  | 07:55 |
| Air Blank          | 0.000  | 07:56 |
| Control Test       | 0.078  | 07:57 |
| Air Blank          | 0.000  | 07:57 |
| Control Test       | 0.079  | 07:58 |
| Air Blank          | 0.000  | 07:58 |
| Control Test Stats |        |       |
| Average            | 0.0783 |       |
| Std Dev            | 0.0006 |       |
| Rel Std Dev(%)     | 0.7370 |       |

DIXIE COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000810  
 03/16/2017  
 Software: 8100.27

| Test               | 9/210L | Time  |
|--------------------|--------|-------|
| Air Blank          | 0.000  | 07:51 |
| Control Test       | 0.081  | 07:52 |
| Air Blank          | 0.000  | 07:52 |
| Control Test       | 0.081  | 07:52 |
| Air Blank          | 0.000  | 07:53 |
| Control Test       | 0.081  | 07:53 |
| Air Blank          | 0.000  | 07:54 |
| Control Test Stats |        |       |
| Average            | 0.0810 |       |
| Std Dev            | 0.0000 |       |
| Rel Std Dev(%)     | 0.0000 |       |

BK

DGS

Operator's Signature

Operator's Signature

Operator's Signature

Operator's Signature

DGS