



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Mason CountyS/N 80-000813Date In 7/13/17Date Out 7/26/17☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake		Quality Checks		Flow Calibration																
Performed By <u>[Signature]</u>		Performed By <u>[Signature]</u>		Performed By <u>[Signature]</u>																
☐ Registration ☒ Annual ☐ Return from CMI ☐ Return from Enforcement Electronics ☐ Other _____		☒ Lab Temp °C <u>24.7/20.4</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>448/58</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP102</u> 32mm <u>149</u> (.139 - .169) 36mm <u>164</u> (.156 - .190) 53mm <u>233</u> (.228 - .278) 103mm <u>492</u> (.447 - .547)		☒ Flow Calibration N/A ☐ Flow Calibration Complete Flow Column # <u>Aug 04 2017</u> ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)																
Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose		☒ Barometric Pressure Check Gauge ID # <u>26932</u> / <u>28421</u> ☒ Stability Checks		Maintenance Performed By <u>[Signature]</u> ☒ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____																
Other Equipment: ☐ Power cord ☐ Printer Cable ☒ Other <u>Static Bag</u>		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>G2078</td> <td>201603D 3/8/15</td> </tr> <tr> <td>0.08</td> <td>G8149</td> <td>201611B 11/15/18</td> </tr> <tr> <td>0.20</td> <td>G11621</td> <td>201702B 2/23/19</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG649605 7/14/18</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.05	G2078	201603D 3/8/15	0.08	G8149	201611B 11/15/18	0.20	G11621	201702B 2/23/19	0.08 DGS	N/A	AG649605 7/14/18	Quality Checks Cont. Performed By <u>[Signature]</u> Simulator Temperatures °C External Digital Therm. ID#: <u>300501</u> ☒ 34°C +/- .2 Serial #: <u>G2078</u> ☒ 34°C +/- .2 Serial #: <u>G8149</u> ☒ 34°C +/- .2 Serial #: <u>G11621 / G2408</u>	
Simulator	Serial #	Lot #/Exp																		
0.05	G2078	201603D 3/8/15																		
0.08	G8149	201611B 11/15/18																		
0.20	G11621	201702B 2/23/19																		
0.08 DGS	N/A	AG649605 7/14/18																		
Notes: _____																				

Calibration Adjustment		Department Inspection																													
Performed By _____		Performed By <u>[Signature]</u>																													
☒ Calibration Adjustment N/A ☐ Calibration Adjustment Complete Barometric Pressure Gauge _____ ID # _____		☒ Barometric Pressure ID# <u>26932</u> / <u>G28421</u> <u>1013</u> 1016 Gauge <u>1017</u> <u>1013</u> 1013 Instrument <u>1013</u>																													
<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Mouth Alcohol Solution Lot # <u>2016-C</u> Acetone Stock Solution Lot # <u>2016-B / 2017-A</u> 8/3/17	
Simulator	Serial Number	Lot Number	Expiration																												
0.000		N/A	N/A																												
0.040																															
0.100																															
0.200																															
0.300																															
0.080 DGS	N/A																														
☐ Post Calibration Adjustment Stability Checks		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A										
Simulator	Serial Number	Lot Number	Expiration																												
0.05																															
0.08																															
0.20																															
0.08 DGS	N/A																														
		<table border="1"> <thead> <tr> <th colspan="2">Attachments</th> </tr> </thead> <tbody> <tr> <td> ☒ Form 41 (2) ☒ Pre-Stability Tests ☐ Flow Calibration </td> <td> ☐ Calibration Adjustment ☐ Post-Stability Tests ☒ Other <u>Form 40</u> </td> </tr> </tbody> </table>		Attachments		☒ Form 41 (2) ☒ Pre-Stability Tests ☐ Flow Calibration	☐ Calibration Adjustment ☐ Post-Stability Tests ☒ Other <u>Form 40</u>																								
Attachments																															
☒ Form 41 (2) ☒ Pre-Stability Tests ☐ Flow Calibration	☐ Calibration Adjustment ☐ Post-Stability Tests ☒ Other <u>Form 40</u>																														

Notes/Suggested Service:

Got Int. Det. on Dry Gas during first D.I.
 Ran 100 stability checks over two days
 and found no issue. Found out they were fixing
 electric and air conditioning in building at time
 of inspection. Repeat inspection on a different

Quality Control Review

Date

day to show in compliance and no issue.
QA/QC [Signature] 8/1/17

- ☒ Instrument Complies with Chapter 11D-8, FAC
☐ Instrument Does Not Comply with Chapter 11D-8, FAC
☒ Return to/Place into Evidentiary Use
☐ Remain Out of Evidentiary Use
☒ Conduct an Agency Inspection Before Evidentiary Use

80-000813
Stab. 1.17 Cuckes
7/26/17

MARION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
07/26/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:21
Control Test	0.079	12:21
Air Blank	0.000	12:21
Control Test	0.079	12:22
Air Blank	0.000	12:22
Control Test	0.079	12:23
Air Blank	0.000	12:23
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DOS

Operator's Signature

MARION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
07/26/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:36
Control Test	0.049	12:36
Air Blank	0.000	12:37
Control Test	0.048	12:38
Air Blank	0.000	12:38
Control Test	0.049	12:39
Air Blank	0.000	12:39
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

Operator's Signature

MARION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
07/26/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:52
Control Test	0.079	12:53
Air Blank	0.000	12:53
Control Test	0.079	12:54
Air Blank	0.000	12:54
Control Test	0.078	12:55
Air Blank	0.000	12:56
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

Operator's Signature

MARION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
07/26/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:06
Control Test	0.199	13:06
Air Blank	0.000	13:07
Control Test	0.199	13:08
Air Blank	0.000	13:08
Control Test	0.198	13:09
Air Blank	0.000	13:09
Control Test Stats		
Average	0.1987	
Std Dev	0.0006	
Rel Std Dev(%)	0.2906	

Operator's Signature

8/1/17

MARION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000813
07/27/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:25
Control Test	0.080	08:25
Air Blank	0.000	08:26
Control Test	0.080	08:26
Air Blank	0.000	08:26
Control Test	0.080	08:27
Air Blank	0.000	08:27
Control Test	0.080	08:28
Air Blank	0.000	08:28
Control Test	0.080	08:28
Air Blank	0.000	08:29
Control Test	0.080	08:29
Air Blank	0.000	08:30
Control Test	0.080	08:30
Air Blank	0.000	08:31
Control Test	0.080	08:31
Air Blank	0.000	08:31
Control Test	0.080	08:32
Air Blank	0.000	08:32
Control Test	0.081	08:33
Air Blank	0.000	08:33
Control Test	0.081	08:33
Air Blank	0.000	08:34
Control Test	0.081	08:34
Air Blank	0.000	08:35
Control Test	0.080	08:35
Air Blank	0.000	08:36
Control Test	0.080	08:36
Air Blank	0.000	08:36
Control Test	0.080	08:37
Air Blank	0.000	08:37
Control Test	0.080	08:38
Air Blank	0.000	08:38
Control Test	0.080	08:38
Air Blank	0.000	08:39
Control Test	0.081	08:39
Air Blank	0.000	08:40
Control Test	0.080	08:40
Air Blank	0.000	08:41
Control Test	0.081	08:41
Air Blank	0.000	08:41
Control Test Stats		
Average	0.0803	
Std Dev	0.0004	
Rel Std Dev(%)	0.5536	

MARION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000813
07/27/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:11
Control Test	0.080	09:11
Air Blank	0.000	09:12
Control Test	0.080	09:12
Air Blank	0.000	09:12
Control Test	0.080	09:13
Air Blank	0.000	09:13
Control Test	0.081	09:14
Air Blank	0.000	09:14
Control Test	0.080	09:14
Air Blank	0.000	09:15
Control Test	0.081	09:15
Air Blank	0.000	09:16
Control Test	0.081	09:16
Air Blank	0.000	09:17
Control Test	0.080	09:17
Air Blank	0.000	09:17
Control Test	0.081	09:18
Air Blank	0.000	09:18
Control Test	0.081	09:19
Air Blank	0.000	09:19
Control Test	0.081	09:19
Air Blank	0.000	09:20
Control Test	0.080	09:20
Air Blank	0.000	09:21
Control Test	0.081	09:21
Air Blank	0.000	09:22
Control Test	0.081	09:22
Air Blank	0.000	09:22
Control Test	0.081	09:23
Air Blank	0.000	09:23
Control Test	0.081	09:24
Air Blank	0.000	09:24
Control Test	0.081	09:24
Air Blank	0.000	09:25
Control Test	0.081	09:25
Air Blank	0.000	09:26
Control Test	0.081	09:26
Air Blank	0.000	09:27
Control Test	0.081	09:27
Air Blank	0.000	09:27
Control Test Stats		
Average	0.0807	
Std Dev	0.0005	
Rel Std Dev(%)	0.5826	

9:11-9:27

MARION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000813
07/27/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:33
Control Test	0.080	11:34
Air Blank	0.000	11:34
Control Test	0.080	11:35
Air Blank	0.000	11:35
Control Test	0.080	11:35
Air Blank	0.000	11:36
Control Test	0.080	11:36
Air Blank	0.000	11:37
Control Test	0.080	11:37
Air Blank	0.000	11:38
Control Test	0.081	11:38
Air Blank	0.000	11:38
Control Test	0.081	11:39
Air Blank	0.000	11:39
Control Test	0.080	11:40
Air Blank	0.000	11:40
Control Test	0.080	11:40
Air Blank	0.000	11:41
Control Test	0.080	11:41
Air Blank	0.000	11:42
Control Test	0.080	11:42
Air Blank	0.000	11:43
Control Test	0.080	11:43
Air Blank	0.000	11:43
Control Test	0.080	11:44
Air Blank	0.000	11:44
Control Test	0.081	11:45
Air Blank	0.000	11:45
Control Test	0.080	11:45
Air Blank	0.000	11:46
Control Test	0.081	11:46
Air Blank	0.000	11:47
Control Test	0.080	11:47
Air Blank	0.000	11:48
Control Test	0.080	11:48
Air Blank	0.000	11:48
Control Test	0.080	11:49
Air Blank	0.000	11:49
Control Test	0.080	11:50
Air Blank	0.000	11:50
Control Test Stats		
Average	0.0802	
Std Dev	0.0004	
Rel Std Dev(%)	0.5117	

11:33-11:50

80-000813
DGS
Testing for
Int. Detect.

None found

(P)

AGG 19605

7/14/18

(P)

Operator Signature

Operator's Signature

Operator's Signature

8/4/17
22

INTOXILYZER 8000
Instrument Initialization
09:03 07/31/2017

MARION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000813
07/27/2017
Software: 8100.27

MARION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000813
07/31/2017
Software: 8100.27

MARION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000813
07/31/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:36
Control Test	0.080	13:36
Air Blank	0.000	13:37
Control Test	0.080	13:37
Air Blank	0.000	13:38
Control Test	0.080	13:38
Air Blank	0.000	13:38
Control Test	0.080	13:39
Air Blank	0.000	13:39
Control Test	0.080	13:40
Air Blank	0.000	13:40
Control Test	0.081	13:40
Air Blank	0.000	13:41
Control Test	0.080	13:41
Air Blank	0.000	13:42
Control Test	0.080	13:42
Air Blank	0.000	13:43
Control Test	0.081	13:43
Air Blank	0.000	13:43
Control Test	0.080	13:44
Air Blank	0.000	13:44
Control Test	0.080	13:45
Air Blank	0.000	13:45
Control Test	0.080	13:45
Air Blank	0.000	13:46
Control Test	0.080	13:46
Air Blank	0.000	13:47
Control Test	0.081	13:47
Air Blank	0.000	13:48
Control Test	0.080	13:48
Air Blank	0.000	13:48
Control Test	0.080	13:49
Air Blank	0.000	13:49
Control Test	0.080	13:50
Air Blank	0.000	13:50
Control Test	0.080	13:50
Air Blank	0.000	13:51
Control Test	0.080	13:51
Air Blank	0.000	13:51
Control Test	0.081	13:52
Air Blank	0.000	13:52
Control Test	0.080	13:53
Air Blank	0.000	13:53
Control Test Stats		
Average	0.0802	
Std Dev	0.0004	
Rel Std Dev(%)	0.4571	

Test	g/210L	Time
Air Blank	0.000	10:42
Control Test	0.080	10:43
Air Blank	0.000	10:43
Control Test	0.079	10:43
Air Blank	0.000	10:44
Control Test	0.080	10:44
Air Blank	0.000	10:45
Control Test	0.079	10:45
Air Blank	0.000	10:46
Control Test	0.079	10:46
Air Blank	0.000	10:46
Control Test	0.080	10:47
Air Blank	0.000	10:47
Control Test	0.080	10:48
Air Blank	0.000	10:48
Control Test	0.080	10:48
Air Blank	0.000	10:49
Control Test	0.080	10:49
Air Blank	0.000	10:50
Control Test	0.080	10:50
Air Blank	0.000	10:51
Control Test	0.080	10:51
Air Blank	0.000	10:51
Control Test	0.080	10:52
Air Blank	0.000	10:52
Control Test	0.080	10:53
Air Blank	0.000	10:53
Control Test	0.081	10:53
Air Blank	0.000	10:54
Control Test	0.080	10:54
Air Blank	0.000	10:55
Control Test	0.080	10:55
Air Blank	0.000	10:56
Control Test	0.080	10:56
Air Blank	0.000	10:56
Control Test	0.080	10:57
Air Blank	0.000	10:57
Control Test	0.081	10:58
Air Blank	0.000	10:58
Control Test	0.080	10:58
Air Blank	0.000	10:59
Control Test Stats		
Average	0.0799	
Std Dev	0.0005	
Rel Std Dev(%)	0.6384	

Test	g/210L	Time
Air Blank	0.000	15:04
Control Test	0.080	15:05
Air Blank	0.000	15:05
Control Test	0.080	15:05
Air Blank	0.000	15:06
Control Test	0.080	15:06
Air Blank	0.000	15:07
Control Test	0.080	15:07
Air Blank	0.000	15:08
Control Test	0.080	15:08
Air Blank	0.000	15:08
Control Test	0.080	15:09
Air Blank	0.000	15:09
Control Test	0.080	15:10
Air Blank	0.000	15:10
Control Test	0.079	15:10
Air Blank	0.000	15:11
Control Test	0.080	15:11
Air Blank	0.000	15:12
Control Test	0.080	15:12
Air Blank	0.000	15:13
Control Test	0.080	15:13
Air Blank	0.000	15:13
Control Test	0.080	15:14
Air Blank	0.000	15:14
Control Test	0.080	15:14
Air Blank	0.000	15:15
Control Test	0.080	15:15
Air Blank	0.000	15:16
Control Test	0.080	15:16
Air Blank	0.000	15:17
Control Test	0.080	15:17
Air Blank	0.000	15:18
Control Test	0.080	15:18
Air Blank	0.000	15:18
Control Test	0.080	15:19
Air Blank	0.000	15:19
Control Test	0.080	15:20
Air Blank	0.000	15:20
Control Test	0.080	15:21
Air Blank	0.000	15:21
Control Test Stats		
Average	0.0800	
Std Dev	0.0002	
Rel Std Dev(%)	0.2797	

80-000813
Testing for
Int. Detect
None
Found

AG-619605
7/11/18

ADB

13:36 → 13:53

10:42-10:59
7/31

15:04-15:21

Operator's Signature

Operator's Signature

Operator's Signature

8/4/17