



# INSTRUMENT PROCESSING SHEET

Agency Citrus CountyS/N 80-000819

Florida Department of Law Enforcement

Date In 10/31/17DI Completion Date 11/2/17 Ship  P/U  H/D  CMI  EE

**Intake** Performed By [Signature]

Annual  
 Registration  
 Return from CMI / EE

Visual Inspection:  
 Case  Handle  
 Keyboard  Dry Gas Shelf  
 Feet  Breath Tube  
 Ports  Screws Tight

Other Equipment/ Accessories:  
 Power cord  Printer Cable  
 Static Bag  12V DC Cable

Notes: \_\_\_\_\_

**Quality Checks** Performed By [Signature]

Breath Tube Screen  
 Replace External O-Rings  
 Instrument Set Up Verified  
 R-Value 127  
 Flow Verification (L/s)  
 Flow Column # ATP105  
 32 mm 144 (.139 - .169)  
 36 mm 158 (.156 - .190)  
 53 mm 236 (.228 - .278)  
 103 mm 500 (.447 - .547)  
 Barometric Pressure Check  
 Gauge ID # 28427  
 Stability Checks

**Flow Calibration** Performed By \_\_\_\_\_

Flow Column # \_\_\_\_\_  
 5L/min - 17mm  
 15L/min - 53mm  
 30L/min - 103mm  
 R-Value \_\_\_\_\_  
 Post Calibration Verification (L/s)  
 Flow Column # \_\_\_\_\_  
 32 mm \_\_\_\_\_ (.139 - .169)  
 36 mm \_\_\_\_\_ (.156 - .190)  
 53 mm \_\_\_\_\_ (.228 - .278)  
 103 mm \_\_\_\_\_ (.447 - .547)

Simulator	Serial #	Lot #/Exp
0.050	<u>G2835</u>	201707D 07/25/2019
0.080	<u>DR1279</u>	201707E 07/25/2019
0.200	<u>SD1025</u>	201707C 07/24/2019
0.080 DGS	N/A	<u>AG626605</u> 9/22/18

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Temperature Checks** Performed By [Signature]

Lab Temp °C 22.0  
 External Digital Therm. ID#: 300504  
 34°C +/- .2 Serial #: G2835  
 34°C +/- .2 Serial #: DR1279  
 34°C +/- .2 Serial #: SD1025

**Final Release Date**

**NOV 03 2017**

**Calibration Adjustment** Performed By \_\_\_\_\_

Barometric Pressure Gauge ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

**Department Inspection** Performed By [Signature]

Barometric Pressure ID# 28427  
 Gauge 1019 Instrument 1020  
 Mouth Alcohol Solution Lot # 2016-C  
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.000	<u>G2880</u>
Interferent	<u>G2840</u>
0.050	<u>G2835</u>
0.080	<u>DR1279</u>
0.200	<u>SD1025</u>

**Attachments**

Form 41  Flow Calibration  
 Stability Checks  Form 40  
 Calibration Adjustment  Other \_\_\_\_\_  
 Post-Stability Checks

Notes/Suggested Service: \_\_\_\_\_

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC  
 Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use

[Signature] 11/2/17 Tech Review / Date [Signature] 11/3/17 Admin Review / Date

80-000819  
 Stability Checks  
 11/2/17

INTOXILYZER 8000  
 Instrument Initialization  
 05:33 11/02/2017

CITRUS COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000819  
 11/02/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:14
Control Test	0.049	06:14
Air Blank	0.000	06:15
Control Test	0.049	06:16
Air Blank	0.000	06:16
Control Test	0.050	06:17
Air Blank	0.000	06:17
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

Operator's Signature

11/3/17

CITRUS COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000819  
 11/02/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:19
Control Test	0.081	06:19
Air Blank	0.000	06:20
Control Test	0.082	06:20
Air Blank	0.000	06:20
Control Test	0.082	06:21
Air Blank	0.000	06:21
Control Test Stats		
Average	0.0817	
Std Dev	0.0006	
Rel Std Dev(%)	0.7070	

DGS

Operator's Signature

CITRUS COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000819  
 11/02/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:25
Control Test	0.204	06:26
Air Blank	0.000	06:26
Control Test	0.204	06:27
Air Blank	0.000	06:27
Control Test	0.205	06:28
Air Blank	0.000	06:29
Control Test Stats		
Average	0.2043	
Std Dev	0.0006	
Rel Std Dev(%)	0.2826	

Operator's Signature

CITRUS COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000819  
 11/02/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:30
Control Test	0.080	06:31
Air Blank	0.000	06:31
Control Test	0.080	06:32
Air Blank	0.000	06:33
Control Test	0.081	06:33
Air Blank	0.000	06:34
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

DSM

Operator's Signature