



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Mount Dora PD

S/N 80-000825

Date In 8/3/17

Date Out 8/3/17

Ship P/U H/D CMI EE

Intake Performed By COB

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
 Case Handle
 Dry Gas Holder Feet
 Keyboard/Plug Back/Plugs
 Screws tight Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By EMS

Lab Temp °C 20.8
 Breath Tube Screen
 Replace External O-Rings
 Instrument Set Up Verified
 R-Value 214
 Flow Verification (L/s)
 Flow Column # API105
 32mm 0.156 (.139 - .169)
 36mm 0.167 (.156 - .190)
 53mm 0.238 (.228 - .278)
 103mm 0.507 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	G2078	201603D 3/8/18
0.08	G8149	201611B 11/5/18
0.20	G2408	201702B 2/23/19
0.08 DGS	N/A	AG1621604 9/22/18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 47mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Quality Checks Cont. Performed By EMS

Simulator Temperatures °C
 External Digital Therm. ID#: 300504
 34°C +/- .2 Serial #: G2078
 34°C +/- .2 Serial #: G8149
 34°C +/- .2 Serial #: G2408

Calibration Adjustment Performed By _____

Calibration Adjustment N/A
 Calibration Adjustment Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By EMS

Barometric Pressure 1017 Gauge
 ID# 28427 1016 Instrument

Mouth Alcohol Solution Lot # 2016C
 Acetone Stock Solution Lot # 2017A

Simulator	Serial Number
0.00	G2879
Interferent	G8144
0.05	G2078
0.08	G8149
0.20	G2408

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Calibration Adjustment
 Post-Stability Tests
 Other _____

Notes/Suggested Service:
Please change level 2 menu password to something unique. EMS

QC 8/3/17 JJ Graham 8/4/17
 Quality Control Review Date

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

RECEIVED
AUG 04 2017
FDLE
Alcohol Testing Program

Stability Checks #80-000825 Mount Dora P.D. 8/3/17 ~~RDB~~

500

MOUNT DORA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000825
08/03/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:11
Control Test	0.048	13:12
Air Blank	0.000	13:12
Control Test	0.048	13:13
Air Blank	0.000	13:13
Control Test	0.049	13:14
Air Blank	0.000	13:15
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	

~~RDB~~

Operator's Signature

8/4/17
ee

MOUNT DORA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000825
08/03/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:06
Control Test	0.077	13:06
Air Blank	0.000	13:07
Control Test	0.079	13:08
Air Blank	0.000	13:08
Control Test	0.079	13:09
Air Blank	0.000	13:09
Control Test Stats		
Average	0.0783	
Std Dev	0.0012	
Rel Std Dev(%)	1.4741	

~~RDB~~

Operator's Signature

MOUNT DORA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000825
08/03/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:00
Control Test	0.200	13:01
Air Blank	0.000	13:01
Control Test	0.199	13:02
Air Blank	0.000	13:03
Control Test	0.199	13:03
Air Blank	0.000	13:04
Control Test Stats		
Average	0.1933	
Std Dev	0.0005	
Rel Std Dev(%)	0.2896	

~~RDB~~

Operator's Signature

MOUNT DORA P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000825
08/03/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:18
Control Test	0.081	13:18
Air Blank	0.000	13:18
Control Test	0.082	13:19
Air Blank	0.000	13:19
Control Test	0.081	13:20
Air Blank	0.000	13:20
Control Test Stats		
Average	0.0813	
Std Dev	0.0006	
Rel Std Dev(%)	0.7099	

~~RDB~~

Operator's Signature