



INSTRUMENT PROCESSING SHEET

Agency Panama City Beach PDS/N 80-000838

Florida Department of Law Enforcement

Date In 11/14/17DI Completion Date 11/16/17 Ship P/U H/D CMI EE

Intake Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>185</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP105</u> 32 mm <u>1.164</u> (.139 - .169) 36 mm <u>1.171</u> (.156 - .190) 53 mm <u>1.257</u> (.228 - .278) 103 mm <u>1.515</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)
---	--	--

Final Release Date

NOV 21 2017

Simulator	Serial #	Lot #/Exp
0.050	G2835	201707D 7/25/19
0.080	DR1279	201707E 7/25/19
0.200	SD1025	201707C 7/24/19
0.080 DGS	N/A	AG619605 7/14/18

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Temperature Checks Performed By [Signature]

Lab Temp °C 21.8
 External Digital Therm. ID#: 300504
 34°C +/- .2 Serial #: G2835
 34°C +/- .2 Serial #: DR1279
 34°C +/- .2 Serial #: SD1025

Calibration Adjustment Performed By _____

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By [Signature]

Barometric Pressure ID# 28427
 Gauge 1015 Instrument 1015
 Mouth Alcohol Solution Lot # 2016-C
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.000	G2880
Interferent	G2840
0.050	G2835
0.080	DR1279
0.200	SD1025

Attachments

Form 41 Flow Calibration
 Stability Checks Form 40
 Calibration Adjustment Other _____
 Post-Stability Checks

Notes/Suggested Service: power supply problem reported but not observed.

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

[Signature] 11/20/17 [Signature] 11/21/17
 Tech Review / Date Admin Review / Date

PANAMA CITY BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000838
11/16/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:30
Control Test	0.050	08:31
Air Blank	0.000	08:32
Control Test	0.050	08:32
Air Blank	0.000	08:33
Control Test	0.050	08:34
Air Blank	0.000	08:34
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P Murphy

Operator's Signature

PANAMA CITY BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000838
11/16/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:38
Control Test	0.081	08:39
Air Blank	0.000	08:40
Control Test	0.081	08:40
Air Blank	0.000	08:41
Control Test	0.081	08:42
Air Blank	0.000	08:42
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P Murphy

Operator's Signature

PANAMA CITY BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000838
11/16/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:44
Control Test	0.203	08:45
Air Blank	0.000	08:45
Control Test	0.202	08:46
Air Blank	0.000	08:47
Control Test	0.201	08:47
Air Blank	0.000	08:48
Control Test Stats		
Average	0.2020	
Std Dev	0.0010	
Rel Std Dev(%)	0.4950	

P Murphy

Operator's Signature

DGS

PANAMA CITY BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000838
11/16/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:49
Control Test	0.080	08:50
Air Blank	0.000	08:50
Control Test	0.080	08:51
Air Blank	0.000	08:51
Control Test	0.081	08:51
Air Blank	0.000	08:52
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

DGS

P Murphy

Operator's Signature

11/21/17
08