



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Lynn Haven PD

S/N 80-000840

Date In 9/29/17

Date Out 10/4/2017

Ship P/U H/D CMI EE

Intake Performed By TDS

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By PGM

Lab Temp °C 22.4
 Breath Tube Screen
 Replace External O-Rings
 Instrument Set Up Verified
 R-Value 182
 Flow Verification (L/s)
 Flow Column # NTP102
 32mm .156 (.139 - .169)
 36mm .160 (.156 - .190)
 53mm .238 (.228 - .278)
 103mm .484 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 26932

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>5D3962</u>	<u>2017070</u> <u>7/25/19</u>
0.08	<u>5D1013</u>	<u>201707E</u> <u>7/25/19</u>
0.20	<u>DR3856</u>	<u>201707C</u> <u>7/24/19</u>
0.08 DGS	N/A	<u>AG708807</u> <u>3/29/19</u>

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Quality Checks Cont. Performed By PGM

Simulator Temperatures °C
 External Digital Therm. ID#: 300505
 34°C +/- .2 Serial #: 5D3962
 34°C +/- .2 Serial #: 5D1013
 34°C +/- .2 Serial #: DR3856

Calibration Adjustment Performed By _____

Calibration Adjustment N/A
 Calibration Adjustment Complete
 Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Department Inspection Performed By PGM

Barometric Pressure 1023 Gauge
 ID# 26932 1024 Instrument

Mouth Alcohol Solution Lot # 2016-C
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	<u>5D1019</u>
Interferent	<u>5D1021</u>
0.05	<u>5D3962</u>
0.08	<u>5D1013</u>
0.20	<u>DR3856</u>

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Calibration Adjustment
 Post-Stability Tests
 Other _____


Notes/Suggested Service:
QA/QC OK SP

JJ Durham 10/5/17
 Quality Control Review Date

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

LYNN HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000840
10/03/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:16
Control Test	0.049	12:17
Air Blank	0.000	12:17
Control Test	0.050	12:18
Air Blank	0.000	12:18
Control Test	0.049	12:19
Air Blank	0.000	12:20
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	



Operator's Signature

LYNN HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000840
10/03/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:30
Control Test	0.081	12:31
Air Blank	0.000	12:32
Control Test	0.082	12:32
Air Blank	0.000	12:33
Control Test	0.082	12:34
Air Blank	0.000	12:34
Control Test Stats		
Average	0.0817	
Std Dev	0.0006	
Rel Std Dev(%)	0.7070	



Operator's Signature

LYNN HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000840
10/03/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:51
Control Test	0.201	12:52
Air Blank	0.000	12:52
Control Test	0.201	12:53
Air Blank	0.000	12:53
Control Test	0.201	12:54
Air Blank	0.000	12:55
Control Test Stats		
Average	0.2010	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



Operator's Signature

LYNN HAVEN PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000840
10/03/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:11
Control Test	0.079	13:11
Air Blank	0.000	13:12
Control Test	0.080	13:12
Air Blank	0.000	13:13
Control Test	0.079	13:13
Air Blank	0.000	13:13
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

D65



Operator's Signature

SP

10/5/17
JP