

INSTRUMENT PROCESSING SHEET

Agency Miami Dade Police Department S/N 80-000880
 Date In 3/22/2017 Date Out 3/23/2017 Ship P/U H/D CMI EE

RECEIVED
 APR 03 2017
 FDLE
 Alcohol Testing Program

Intake Performed By DELL

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other ANTI STATIC BAG

Notes: _____

Quality Checks Performed By DELL

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 227
 Flow Verification (L/s)
 Flow Column # ATP101
 32mm 160 (.139 - .169)
 36mm 183 (.156 - .190)
 53mm 250 (.228 - .278)
 103mm 500 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 68639

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>503963</u>	<u>201603D</u> <u>03/08/2018</u>
0.08	<u>503948</u>	<u>201601F</u> <u>01/26/2018</u>
0.20	<u>503969</u>	<u>201604C</u> <u>04/05/2018</u>
0.08 DGS	N/A	<u>A6626605</u> <u>09/22/2018</u>

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete
 Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By DELL

Barometric Pressure 1020 Gauge
 ID# 28199 1020 Instrument

Mouth Alcohol Solution Lot # 2016A
 Acetone Stock Solution Lot # 2016B

Simulator	Serial Number
0.00	<u>SD 3965</u>
Interferent	<u>SD 3966</u>
0.05	<u>SD 3963</u>
0.08	<u>SD 3968</u>
0.20	<u>SD 3969</u>

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: **E-MAILED**
3/23/2017
QA/PC OK 3/30/2017 PJM

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Bruce Kirkland
 Quality Control Review

4/3/17
 Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-000880	Miami-Dade Police Department	3/23/2017	<i>WLL</i>

RL

<p>0.05g/210L <input checked="" type="checkbox"/></p> <p>0.047 to 0.053 <input checked="" type="checkbox"/></p> <p>Miami-Dade PD Intoxilyzer - Alcotest Analyzer Model 8000 SN 80-000880 03/23/2017 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 07:51 Control Test 0.049 07:52 Air Blank 0.000 07:52 Control Test 0.050 07:53 Air Blank 0.000 07:54 Control Test 0.051 07:54 Air Blank 0.000 07:55 Control Test Stats Average 0.0500 Std Dev 0.0010 Rel Std Dev(%) 2.0000</p> <p><i>WLL</i> Operator's Signature</p>	<p>0.08g/210L <input checked="" type="checkbox"/></p> <p>0.077 to 0.083 <input checked="" type="checkbox"/></p> <p>Miami-Dade PD Intoxilyzer - Alcotest Analyzer Model 8000 SN 80-000880 03/23/2017 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 07:56 Control Test 0.080 07:57 Air Blank 0.000 07:58 Control Test 0.080 07:58 Air Blank 0.000 07:59 Control Test 0.080 08:00 Air Blank 0.000 08:00 Control Test Stats Average 0.0800 Std Dev 0.0010 Rel Std Dev(%) 0.0000</p> <p><i>WLL</i> Operator's Signature</p>	<p>0.20g/210L <input checked="" type="checkbox"/></p> <p>0.194 to 0.206 <input checked="" type="checkbox"/></p> <p>Miami-Dade PD Intoxilyzer - Alcotest Analyzer Model 8000 SN 80-000880 03/23/2017 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 08:02 Control Test 0.198 08:02 Air Blank 0.000 08:03 Control Test 0.199 08:04 Air Blank 0.000 08:04 Control Test 0.197 08:05 Air Blank 0.000 08:05 Control Test Stats Average 0.1977 Std Dev 0.0006 Rel Std Dev(%) 0.2921</p> <p><i>WLL</i> Operator's Signature</p>	<p>DGS 0.08g/210L <input checked="" type="checkbox"/></p> <p>0.077 to 0.083 <input checked="" type="checkbox"/></p> <p>Miami-Dade PD Intoxilyzer - Alcotest Analyzer Model 8000 SN 80-000880 03/23/2017 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 08:07 Control Test 0.078 08:07 Air Blank 0.000 08:07 Control Test 0.079 08:08 Air Blank 0.000 08:08 Control Test 0.080 08:09 Air Blank 0.000 08:09 Control Test Stats Average 0.0790 Std Dev 0.0010 Rel Std Dev(%) 1.2658</p> <p><i>WLL</i> Operator's Signature</p>
---	---	---	---

ggm