



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Port St Joe PD S/N 80-000908
 Date In 10/10/17 Date Out 10/13/17 Ship P/U H/D CMI EE

Intake <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>Gatic Bag</u> Notes: _____ _____ _____	Quality Checks Performed By <u>PPM</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.2</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>251</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP105</u> 32mm <u>.156</u> (.139 - .169) 36mm <u>.167</u> (.156 - .190) 53mm <u>.246</u> (.228 - .278) 103mm <u>.530</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28421</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td><u>62835</u></td> <td><u>201707D</u> <u>7/25/19</u></td> </tr> <tr> <td>0.08</td> <td><u>DR1279</u></td> <td><u>201707E</u> <u>7/25/19</u></td> </tr> <tr> <td>0.20</td> <td><u>SD1025</u></td> <td><u>201707C</u> <u>7/24/19</u></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td><u>AG708807</u> <u>3/29/19</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	<u>62835</u>	<u>201707D</u> <u>7/25/19</u>	0.08	<u>DR1279</u>	<u>201707E</u> <u>7/25/19</u>	0.20	<u>SD1025</u>	<u>201707C</u> <u>7/24/19</u>	0.08 DGS	N/A	<u>AG708807</u> <u>3/29/19</u>	Flow Calibration Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547) Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Quality Checks Cont. Performed By <u>PPM</u> Simulator Temperatures °C External Digital Therm. ID#: <u>300504</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>62835</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>DR1279</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD1025</u>
Simulator	Serial #	Lot #/Exp															
0.05	<u>62835</u>	<u>201707D</u> <u>7/25/19</u>															
0.08	<u>DR1279</u>	<u>201707E</u> <u>7/25/19</u>															
0.20	<u>SD1025</u>	<u>201707C</u> <u>7/24/19</u>															
0.08 DGS	N/A	<u>AG708807</u> <u>3/29/19</u>															

RECEIVED
OCT 16 2017
FDLE
Alcohol Testing Program

Calibration Adjustment Performed By _____

Calibration Adjustment N/A
 Calibration Adjustment Complete
 Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By PPM

Barometric Pressure 1013 Gauge
 ID# 28421 1014 Instrument

Mouth Alcohol Solution Lot # 2016-C
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	<u>G2880</u>
Interferent	<u>G2840</u>
0.05	<u>G2835</u>
0.08	<u>DR1279</u>
0.20	<u>SD1025</u>

Attachments

<input checked="" type="checkbox"/> Form 41	<input type="checkbox"/> Calibration Adjustment
<input checked="" type="checkbox"/> Pre-Stability Tests	<input type="checkbox"/> Post-Stability Tests
<input type="checkbox"/> Flow Calibration	<input type="checkbox"/> Other _____

Notes/Suggested Service:

OK PPM 10/16/17
JJ Dehan 10/16/17

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Quality Control Review

Date

PORT ST JOE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000908
10/13/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:19
Control Test	0.048	08:20
Air Blank	0.000	08:21
Control Test	0.048	08:21
Air Blank	0.000	08:22
Control Test	0.049	08:23
Air Blank	0.000	08:23
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	



P Murphy

Operator's Signature

PORT ST JOE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000908
10/13/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:14
Control Test	0.078	08:14
Air Blank	0.000	08:15
Control Test	0.078	08:16
Air Blank	0.000	08:16
Control Test	0.078	08:17
Air Blank	0.000	08:17
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



P Murphy

Operator's Signature

PORT ST JOE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000908
10/13/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:30
Control Test	0.199	08:30
Air Blank	0.000	08:31
Control Test	0.200	08:32
Air Blank	0.000	08:32
Control Test	0.200	08:33
Air Blank	0.000	08:33
Control Test Stats		
Average	0.1997	
Std Dev	0.0006	
Rel Std Dev(%)	0.2892	



P Murphy

Operator's Signature

PORT ST JOE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000908
10/13/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:44
Control Test	0.080	08:44
Air Blank	0.000	08:45
Control Test	0.080	08:45
Air Blank	0.000	08:46
Control Test	0.080	08:46
Air Blank	0.000	08:47
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



DGS

P Murphy

Operator's Signature

10/14/17
JA