



# INSTRUMENT PROCESSING SHEET

Agency Charlotte CountyS/N 80-000945

Florida Department of Law Enforcement

Date In 11/6/17DI Completion Date 11/8/17 Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>SP</u>		<b>Quality Checks</b> Performed By <u>SP</u>		<b>Flow Calibration</b> Performed By _____																
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>1.21</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP102</u> 32 mm <u>.144</u> (.139 - .169) 36 mm <u>.147</u> (.156 - .190) 53 mm <u>.238</u> (.228 - .278) 103 mm <u>.507</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks		Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																
<b>Final Release Date</b>		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>SD3962</u></td> <td><u>201707D</u> <u>7-25-19</u></td> </tr> <tr> <td>0.080</td> <td><u>SD1013</u></td> <td><u>201707E</u> <u>7-25-19</u></td> </tr> <tr> <td>0.200</td> <td><u>SP 201707C</u> <u>DR3856</u></td> <td><u>201707C</u> <u>7-24-19</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG10210604</u> <u>9-22-18</u></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	<u>SD3962</u>	<u>201707D</u> <u>7-25-19</u>	0.080	<u>SD1013</u>	<u>201707E</u> <u>7-25-19</u>	0.200	<u>SP 201707C</u> <u>DR3856</u>	<u>201707C</u> <u>7-24-19</u>	0.080 DGS	N/A	<u>AG10210604</u> <u>9-22-18</u>	<b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____	
Simulator	Serial #	Lot #/Exp																		
0.050	<u>SD3962</u>	<u>201707D</u> <u>7-25-19</u>																		
0.080	<u>SD1013</u>	<u>201707E</u> <u>7-25-19</u>																		
0.200	<u>SP 201707C</u> <u>DR3856</u>	<u>201707C</u> <u>7-24-19</u>																		
0.080 DGS	N/A	<u>AG10210604</u> <u>9-22-18</u>																		
<b>NOV 14 2017</b>				<b>Temperature Checks</b> Performed By <u>SP</u>																
				<input checked="" type="checkbox"/> Lab Temp °C <u>22.3</u> External Digital Therm. ID#: <u>300504</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3962</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD1013</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>DR3856</u>																

<b>Calibration Adjustment</b> Performed By _____			
Barometric Pressure Gauge _____ ID # _____			
Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		
<input type="checkbox"/> Post Calibration Adjustment Stability Checks			
Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

<b>Department Inspection</b> Performed By <u>SP</u>	
Barometric Pressure ID# <u>26932</u>	
Gauge <u>1016</u>	Instrument <u>1016</u>
Mouth Alcohol Solution Lot # <u>2016-C</u>	
Acetone Stock Solution Lot # <u>2017-A</u>	
Simulator	Serial Number
0.000	<u>SD1019</u>
Interferent	<u>SD1021</u>
0.050	<u>SD3962</u>
0.080	<u>SD1013</u>
0.200	<u>DR3856</u>

<b>Attachments</b>	
<input checked="" type="checkbox"/> Form 41	<input type="checkbox"/> Flow Calibration
<input checked="" type="checkbox"/> Stability Checks	<input type="checkbox"/> Form 40
<input type="checkbox"/> Calibration Adjustment	<input type="checkbox"/> Other _____
<input type="checkbox"/> Post-Stability Checks	

Notes/Suggested Service: \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC
<input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC
<input checked="" type="checkbox"/> Return to/Place into Evidentiary Use
<input type="checkbox"/> Remain Out of Evidentiary Use
<input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use
<u>SP</u> 11/8/17 <u>JJ Durham</u> 11/14/17
Tech Review / Date      Admin Review / Date

# STABILITY CHECKS -- #80-000945

CHARLOTTE COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000945  
11/08/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:08
Control Test	0.050	09:09
Air Blank	0.000	09:10
Control Test	0.049	09:10
Air Blank	0.000	09:11
Control Test	0.049	09:12
Air Blank	0.000	09:12
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel. Std Dev(%)	1.1703	

SJP  
Operator's Signature

11/14/17  
DGS

CHARLOTTE COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000945  
11/08/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:03
Control Test	0.080	09:04
Air Blank	0.000	09:04
Control Test	0.080	09:05
Air Blank	0.000	09:05
Control Test	0.080	09:06
Air Blank	0.000	09:07
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel. Std Dev(%)	0.0000	

SJP  
Operator's Signature

CHARLOTTE COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000945  
11/08/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:19
Control Test	0.198	09:20
Air Blank	0.000	09:21
Control Test	0.197	09:21
Air Blank	0.000	09:22
Control Test	0.197	09:23
Air Blank	0.000	09:23
Control Test Stats		
Average	0.1973	
Std Dev	0.0006	
Rel. Std Dev(%)	0.2926	

SJP  
Operator's Signature

CHARLOTTE COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000945  
11/08/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:14
Control Test	0.081	09:14
Air Blank	0.000	09:15
Control Test	0.080	09:15
Air Blank	0.000	09:16
Control Test	0.082	09:16
Air Blank	0.000	09:16
Control Test Stats		
Average	0.0810	
Std Dev	0.0010	
Rel. Std Dev(%)	1.2346	

DGS

SJP  
Operator's Signature