



# INSTRUMENT PROCESSING SHEET

Agency Glades CountyS/N 80-000947

Florida Department of Law Enforcement

Date In 11/6/17DI Completion Date 11/8/17 Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>SP</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	<b>Quality Checks</b> Performed By <u>SP</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>144</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP102</u> 32 mm <u>.144</u> (.139 - .169) 36 mm <u>.156</u> (.156 - .190) 53 mm <u>.230</u> (.228 - .278) 103 mm <u>.507</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks	<b>Flow Calibration</b> Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)
--	---	--

**Final Release Date**

NOV 14 2017

Simulator	Serial #	Lot #/Exp
0.050	SD3962	201707D 7-25-19
0.080	SD1013	201707E 7-25-19
0.200	DR3856	201707C 7-24-19
0.080 DGS	N/A	AG1026604 9-22-18

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Temperature Checks** Performed By SP

Lab Temp °C 22.3  
 External Digital Therm. ID#: 300504  
 34°C +/-2 Serial #: SD3962  
 34°C +/-2 Serial #: SD1013  
 34°C +/-2 Serial #: DR3856

**Calibration Adjustment** Performed By \_\_\_\_\_

Barometric Pressure Gauge ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

**Department Inspection** Performed By SP

Barometric Pressure ID# 26932  
 Gauge 1016 Instrument 1016  
 Mouth Alcohol Solution Lot # 2016-C  
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.000	SD1019
Interferent	SD1021
0.050	SD3962
0.080	SD1013
0.200	DR3856

**Attachments**

Form 41  Flow Calibration  
 Stability Checks  Form 40  
 Calibration Adjustment  Other \_\_\_\_\_  
 Post-Stability Checks

Notes/Suggested Service: \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC  
 Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use

SP 11/8/17 J. Bahan 11/14/17  
 Tech Review / Date Admin Review / Date

# STABILITY CHECKS - #80-000947

GLADES COUNTY S.O.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000947  
 11/08/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:13
Control Test	0.049	09:14
Air Blank	0.000	09:14
Control Test	0.049	09:15
Air Blank	0.000	09:15
Control Test	0.049	09:16
Air Blank	0.000	09:17
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

11/14/17  
 DS

GLADES COUNTY S.O.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000947  
 11/08/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:07
Control Test	0.081	09:08
Air Blank	0.000	09:09
Control Test	0.080	09:09
Air Blank	0.000	09:10
Control Test	0.081	09:10
Air Blank	0.000	09:11
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

SP

Operator's Signature

GLADES COUNTY S.O.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000947  
 11/08/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:02
Control Test	0.197	09:03
Air Blank	0.000	09:04
Control Test	0.198	09:04
Air Blank	0.000	09:05
Control Test	0.198	09:05
Air Blank	0.000	09:06
Control Test Stats		
Average	0.1977	
Std Dev	0.0006	
Rel Std Dev(%)	0.2921	

SP

Operator's Signature

GLADES COUNTY S.O.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000947  
 11/08/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:18
Control Test	0.080	09:18
Air Blank	0.000	09:18
Control Test	0.080	09:19
Air Blank	0.000	09:19
Control Test	0.079	09:20
Air Blank	0.000	09:20
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

DGS

SP

Operator's Signature

ADDS