

INSTRUMENT PROCESSING SHEET

Agency Franklin County SO S/N 80-000952
 Date In 8/29/17 Date Out 9/1/2017 Ship P/U H/D CMI EE

Intake Performed By TP

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
 Case Handle
 Dry Gas Holder Feet
 Keyboard/Plug Back/Plugs
 Screws tight Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By PSM

Lab Temp °C 22.1
 Breath Tube Screen
 Replace External O-Rings
 Instrument Set Up Verified
 R-Value 224
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm .152 (.139 - .169)
 36mm .167 (.156 - .190)
 53mm .234 (.228 - .278)
 103mm .488 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 26932

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD3962	201707D 7/25/19
0.08	SD1013	201707E 7/25/19
0.20	SD3933	201707C 7/24/19
0.08 DGS	N/A	AG619605 7/14/18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Quality Checks Cont. Performed By PSM
Simulator Temperatures °C
 External Digital Therm. ID#: 300505
 34°C +/- .2 Serial #: SD3962
 34°C +/- .2 Serial #: SD1013
 34°C +/- .2 Serial #: SD3933

RECEIVED
 SEP 05 2017
 FDLE
 Alcohol Testing Program

Calibration Adjustment Performed By _____

Calibration Adjustment N/A
 Calibration Adjustment Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By PSM

Barometric Pressure 1014 Gauge
 ID# 26932 1013 Instrument

Mouth Alcohol Solution Lot # 2018-C
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	SD3962
0.08	SD1013
0.20	SD3933

Attachments

Form 41 Calibration Adjustment
 Pre-Stability Tests Post-Stability Tests
 Flow Calibration Other _____

Notes/Suggested Service:
QA/QC OK SP

[Signature] 9/5/17
 Quality Control Review Date

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

FRANKLIN COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000952
 08/31/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:23
Control Test	0.049	13:24
Air Blank	0.000	13:24
Control Test	0.050	13:25
Air Blank	0.000	13:26
Control Test	0.049	13:26
Air Blank	0.000	13:27
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

P Murphy

 Operator's Signature

FRANKLIN COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000952
 08/31/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:18
Control Test	0.081	13:18
Air Blank	0.000	13:19
Control Test	0.082	13:20
Air Blank	0.000	13:20
Control Test	0.082	13:21
Air Blank	0.000	13:21
Control Test Stats		
Average	0.0817	
Std Dev	0.0006	
Rel Std Dev(%)	0.7070	

P Murphy

 Operator's Signature

FRANKLIN COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000952
 08/31/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:28
Control Test	0.201	13:29
Air Blank	0.000	13:30
Control Test	0.201	13:30
Air Blank	0.000	13:31
Control Test	0.201	13:31
Air Blank	0.000	13:32
Control Test Stats		
Average	0.2010	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P Murphy

 Operator's Signature

FRANKLIN COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000952
 08/31/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:40
Control Test	0.081	13:40
Air Blank	0.000	13:41
Control Test	0.081	13:41
Air Blank	0.000	13:42
Control Test	0.080	13:42
Air Blank	0.000	13:42
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

DGS

P Murphy

 Operator's Signature

SP

9/5/17
[Signature]