



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Wakulla County S/N 80-000961

Date In 5/23/17 Date Out 5/24/2017 Ship P/U H/D CMI EE

- Intake** Performed By [Signature]
- Registration
 - Annual
 - Return from CMI
 - Return from Enforcement Electronics
 - Other _____

Visual Inspection:

OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

- Other Equipment:
- Power cord
 - Printer Cable
 - Other _____

Notes: _____

- Quality Checks** Performed By [Signature]
- Lab Temp °C 20.2
 - Breath Tube Screen
 - Replace External O-Rings
 - Instrument Set Up Verified
 - R-Value 187
 - Flow Verification (L/s)

Flow Column # ATP103

32mm	<u>.148</u>	(.139 - .169)
36mm	<u>.160</u>	(.156 - .190)
53mm	<u>.234</u>	(.228 - .278)
103mm	<u>.500</u>	(.447 - .547)

- Barometric Pressure Check Gauge ID # 26932
- Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>DR2035</u>	<u>201603D</u> <u>3/8/18</u>
0.08	<u>SD3964</u>	<u>201611B</u> <u>11/15/18</u>
0.20	<u>G11621</u>	<u>201604C</u> <u>4/5/18</u>
0.08 DGS	N/A	<u>AG626604</u> <u>9/22/18</u>

- Flow Calibration** Performed By _____

- Flow Calibration N/A
- Flow Calibration Complete
- Flow Column # _____
- 5L/min - 7mm
- 15L/min - 53mm
- 30L/min - 103mm

- R-Value _____
 - Post Calibration Verification (L/s)
- Flow Column # _____
- | | | |
|-------|-------|---------------|
| 32mm | _____ | (.139 - .169) |
| 36mm | _____ | (.156 - .190) |
| 53mm | _____ | (.228 - .278) |
| 103mm | _____ | (.447 - .547) |

- Maintenance** Performed By _____

- Battery Replacement
- Dry Gas Regulator Replacement
- Breath Tube Replacement
- Other _____

- Quality Checks Cont.** Performed By [Signature]
- Simulator Temperatures °C**

- External Digital Therm. ID#: 300505
- 34°C +- .2 Serial #: DR2035
 - 34°C +- .2 Serial #: SD3964
 - 34°C +- .2 Serial #: G11621

Calibration Adjustment Performed By _____

- Calibration Adjustment N/A
 - Calibration Adjustment Complete
- Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

- Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By [Signature]

- Barometric Pressure 1001 Gauge
- ID# 26932 1001 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	<u>G2879</u>
Interferent	<u>G8144</u>
0.05	<u>DR2035</u>
0.08	<u>SD3964</u>
0.20	<u>G11621</u>

Attachments

- Form 41
- Pre-Stability Tests
- Flow Calibration
- Calibration Adjustment
- Post-Stability Tests
- Other _____

Notes/Suggested Service:

QC [Signature] 5/30/17

[Signature] 5/30/17

- Instrument Complies with Chapter 11D-8, FAC
- Instrument Does Not Comply with Chapter 11D-8, FAC

- Return to/Place into Evidentiary Use
- Remain Out of Evidentiary Use
- Conduct an Agency Inspection Before Evidentiary Use

Quality Control Review _____ Date _____

WAKULLA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000961
05/24/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:16
Control Test	0.048	10:17
Air Blank	0.000	10:18
Control Test	0.048	10:18
Air Blank	0.000	10:19
Control Test	0.049	10:20
Air Blank	0.000	10:20
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	

P Murphy
Operator's Signature

WAKULLA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000961
05/24/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:26
Control Test	0.079	10:27
Air Blank	0.000	10:28
Control Test	0.079	10:28
Air Blank	0.000	10:29
Control Test	0.080	10:29
Air Blank	0.000	10:30
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

P Murphy
Operator's Signature

WAKULLA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000961
05/24/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:21
Control Test	0.195	10:22
Air Blank	0.000	10:22
Control Test	0.196	10:23
Air Blank	0.000	10:24
Control Test	0.197	10:24
Air Blank	0.000	10:25
Control Test Stats		
Average	0.1960	
Std Dev	0.0010	
Rel Std Dev(%)	0.5102	

P Murphy
Operator's Signature

WAKULLA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000961
05/24/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:31
Control Test	0.080	10:32
Air Blank	0.000	10:32
Control Test	0.080	10:33
Air Blank	0.000	10:33
Control Test	0.080	10:33
Air Blank	0.000	10:34
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P Murphy

DGS

P Murphy
Operator's Signature

JE