



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Orange County S/N 80-000963

Date In 1/17/17 Date Out 1/19/17 Ship P/U H/D CMI EE

Intake Performed By DD

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
 Case Handle
 Dry Gas Holder Feet
 Keyboard/Plug Back/Plugs
 Screws tight Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By DD

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 195
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm 158 (.139 - .169)
 36mm 179 (.156 - .190)
 53mm 260 (.228 - .278)
 103mm 509 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 26932

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	DR2035	201603D 3/8/18
0.08	SD1011	201601F 1/20/18
0.20	SD1025	201601C 4/5/18
0.08 DGS	N/A	AG619605 7/14/18

Flow Calibration Performed By DD

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
JAN 19 2017
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete
 Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By DD

Barometric Pressure 1016 Gauge
 ID# 26932 1016 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	DR2035
0.08	SD1011
0.20	SD1025

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: QC - DGS

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland

Quality Control Review


1/19/17
Date

80-000 963
 State 117 Cues
 1/19/17

INTOXILYZER 8000
 Instrument Initialization
 08:14 01/19/2017

ORANGE COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000963
 01/19/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:44
Control Test	0.197	08:44
Air Blank	0.000	08:45
Control Test	0.195	08:46
Air Blank	0.000	08:46
Control Test	0.196	08:47
Air Blank	0.000	08:47
Control Test Stats		
Average	0.1960	
Std Dev	0.0010	
Rel Std Dev(%)	0.5102	

Operator's Signature 

DS

ORANGE COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000963
 01/19/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:48
Control Test	0.048	08:49
Air Blank	0.000	08:49
Control Test	0.048	08:50
Air Blank	0.000	08:51
Control Test	0.048	08:51
Air Blank	0.000	08:52
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature 

ORANGE COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000963
 01/19/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:53
Control Test	0.078	08:53
Air Blank	0.000	08:54
Control Test	0.078	08:55
Air Blank	0.000	08:55
Control Test	0.078	08:56
Air Blank	0.000	08:56
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature 

BK

ORANGE COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-000963
 01/19/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:39
Control Test	0.079	08:39
Air Blank	0.000	08:40
Control Test	0.079	08:40
Air Blank	0.000	08:41
Control Test	0.080	08:41
Air Blank	0.000	08:41
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

Operator's Signature 

DS