

## INSTRUMENT PROCESSING SHEET

Agency Hood County

S/N 80-001002

Date In 3/20/17

Date Out 3/22/17

Ship  P/U  H/D  CMI  EE

**Intake** Performed By SP

Registration  
 Annual  
 Return from CMI  
 Return from Enforcement  
 Electronics  
 Other \_\_\_\_\_

Visual Inspection:  
OK Case OK Handle  
OK Dry Gas Holder OK Feet  
OK Keyboard/Plug OK Back/Plugs  
OK Screws tight OK Breath Hose

Other Equipment:  
 Power cord  
 Printer Cable  
 Other \_\_\_\_\_

Notes: \_\_\_\_\_

**Quality Checks** Performed By SP

Breath Tube Screen  
 Replace O-Rings  
 Instrument Set Up Verified  
 R-Value 159  
 Flow Verification (L/s)  
 Flow Column # ATP102  
 32mm .148 (.139 - .169)  
 36mm .160 (.156 - .190)  
 53mm .234 (.228 - .278)  
 103mm .511 (.447 - .547)

Barometric Pressure Check  
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	G3709	201603D 3-8-18
0.08	DR1279	201601F 1-26-18
0.20	DR3856	201604C 4-5-18
0.08 DGS	N/A	AG626604 9-22-18

**Flow Calibration** Performed By \_\_\_\_\_

Flow Calibration N/A  
 Flow Calibration Complete  
 Flow Column # \_\_\_\_\_  
 5L/min - 17mm  
 15L/min - 53mm  
 30L/min - 103mm

R-Value \_\_\_\_\_  
 Post Calibration Verification (L/s)  
 Flow Column # \_\_\_\_\_  
 32mm \_\_\_\_\_ (.139 - .169)  
 36mm \_\_\_\_\_ (.156 - .190)  
 53mm \_\_\_\_\_ (.228 - .278)  
 103mm \_\_\_\_\_ (.447 - .547)

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Suggested Service**

RECEIVED  
 MAR 28 2017  
 FDLE  
 Alcohol Testing Program

**Optical Bench Calibration** Performed By \_\_\_\_\_

Optical Bench Calibration N/A  
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

**Department Inspection** Performed By SP

Barometric Pressure  
 ID# 28427 1016 Gauge  
1015 Instrument

Mouth Alcohol Solution Lot # 2016-A  
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	G2880
Interferent	G2834
0.05	G3709
0.08	DR1279
0.20	DR3856

**Attachments**

Form 41 x3  
 Pre-Stability Tests  
 Flow Calibration  
 Optical Bench Cal  
 Post-Stability Tests  
 Other \_\_\_\_\_

Notes: Please change level 2 password to something unique SP

Reprinted inspection w/ correct form.  
loaded correct forms on to instrument. SP  
OCV [initials]

Brett Kirkland

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Quality Control Review

3/28/17  
 Date

STABILITY CHECKS - INSTRUMENT # 80-001002

HARDEE COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001002  
03/22/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:25
Control Test	0.050	10:26
Air Blank	0.000	10:27
Control Test	0.050	10:27
Air Blank	0.000	10:28
Control Test	0.050	10:29
Air Blank	0.000	10:29
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

HARDEE COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001002  
03/22/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:31
Control Test	0.080	10:32
Air Blank	0.000	10:32
Control Test	0.079	10:33
Air Blank	0.000	10:34
Control Test	0.079	10:34
Air Blank	0.000	10:35
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

SP

Operator's Signature

HARDEE COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001002  
03/22/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:37
Control Test	0.199	10:38
Air Blank	0.000	10:38
Control Test	0.199	10:39
Air Blank	0.000	10:40
Control Test	0.199	10:40
Air Blank	0.000	10:41
Control Test Stats		
Average	0.1990	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

HARDEE COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001002  
03/22/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:43
Control Test	0.080	10:43
Air Blank	0.000	10:44
Control Test	0.080	10:44
Air Blank	0.000	10:45
Control Test	0.081	10:45
Air Blank	0.000	10:46
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

DGS

DGS BK

SP

Operator's Signature