

## INSTRUMENT PROCESSING SHEET

Agency Coconut Creek Police Department S/N 80-001044  
 Date In 2/14/2017 Date Out 02/15/2017  Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>DELL</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Other _____ Notes: _____ _____ _____	<b>Quality Checks</b> Performed By <u>DELL</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>164</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP101</u> 32mm <u>144</u> (.139 - .169) 36mm <u>164</u> (.156 - .190) 53mm <u>234</u> (.228 - .278) 103mm <u>496</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28199</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD3967</td> <td>201603D 03/08/2018</td> </tr> <tr> <td>0.08</td> <td>SD3968</td> <td>201601F 01/26/2018</td> </tr> <tr> <td>0.20</td> <td>SD3969</td> <td>201604C 04/05/2018</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG 105301 02/22/2018</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	SD3967	201603D 03/08/2018	0.08	SD3968	201601F 01/26/2018	0.20	SD3969	201604C 04/05/2018	0.08 DGS	N/A	AG 105301 02/22/2018	<b>Flow Calibration</b> Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.05	SD3967	201603D 03/08/2018															
0.08	SD3968	201601F 01/26/2018															
0.20	SD3969	201604C 04/05/2018															
0.08 DGS	N/A	AG 105301 02/22/2018															
		<b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ <b>Suggested Service</b> _____ _____ _____															

RECEIVED  
 FEB 28 2017  
 FDLE  
 Alcohol Testing Program

<b>Optical Bench Calibration</b> Performed By _____ <input checked="" type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.400</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.400				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A		
Simulator	Serial Number	Lot Number	Expiration																																													
0.000		N/A	N/A																																													
0.040																																																
0.100																																																
0.200																																																
0.400																																																
0.080 DGS	N/A																																															
Simulator	Serial Number	Lot Number	Expiration																																													
0.05																																																
0.08																																																
0.20																																																
0.08 DGS	N/A																																															

<b>Department Inspection</b> Performed By <u>DELL</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1010</u> Gauge ID# <u>68639</u> <u>1012</u> Instrument Mouth Alcohol Solution Lot # <u>2046A</u> Acetone Stock Solution Lot # <u>2046B</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>SD 3965</td> </tr> <tr> <td>Interferent</td> <td>SD 3966</td> </tr> <tr> <td>0.05</td> <td>SD 3967</td> </tr> <tr> <td>0.08</td> <td>SD 3968</td> </tr> <tr> <td>0.20</td> <td>SD 3969</td> </tr> </tbody> </table>	Simulator	Serial Number	0.00	SD 3965	Interferent	SD 3966	0.05	SD 3967	0.08	SD 3968	0.20	SD 3969
Simulator	Serial Number											
0.00	SD 3965											
Interferent	SD 3966											
0.05	SD 3967											
0.08	SD 3968											
0.20	SD 3969											
<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Optical Bench Cal <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____												

Notes: **E-MAILED**   
02/15/2017  **APPROVED**  
QA/OC OK Pgm 2/27/2017

<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use
---

Brett Kirkland  
 Quality Control Review

2/28/17  
 Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-001044	Coconut Creek Police Department	02/15/2017	<i>WLL</i>

*WLL*

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>	0.194 to 0.206 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>

COCOONUT CREEK PD  
Intoxilyzer - Alcotest Analyzer  
Model 8000 SN 80-001044  
02/15/2017  
Software: 8100.27

COCOONUT CREEK PD  
Intoxilyzer - Alcotest Analyzer  
Model 8000 SN 80-001044  
02/15/2017  
Software: 8100.27

COCOONUT CREEK PD  
Intoxilyzer - Alcotest Analyzer  
Model 8000 SN 80-001044  
02/15/2017  
Software: 8100.27

COCOONUT CREEK PD  
Intoxilyzer - Alcotest Analyzer  
Model 8000 SN 80-001044  
02/15/2017  
Software: 8100.27

*WLL*

Test g/210L Time

Air Blank 0.000 09:00  
Control Test 0.052 09:01  
Air Blank 0.000 09:02  
Control Test 0.051 09:02  
Air Blank 0.000 09:03  
Control Test 0.051 09:03  
Air Blank 0.000 09:04

Control Test Stats  
Average 0.0513  
Std Dev 0.0006  
Rel Std Dev(%) 1.1247

Test g/210L Time

Air Blank 0.000 09:05  
Control Test 0.081 09:07  
Air Blank 0.000 09:07  
Control Test 0.081 09:08  
Air Blank 0.000 09:08  
Control Test 0.080 09:09  
Air Blank 0.000 09:10

Control Test Stats  
Average 0.0807  
Std Dev 0.0006  
Rel Std Dev(%) 0.7157

Test g/210L Time

Air Blank 0.000 09:17  
Control Test 0.202 09:18  
Air Blank 0.000 09:19  
Control Test 0.200 09:19  
Air Blank 0.000 09:20  
Control Test 0.201 09:21  
Air Blank 0.000 09:21

Control Test Stats  
Average 0.2010  
Std Dev 0.0010  
Rel Std Dev(%) 0.4975

Test g/210L Time

Air Blank 0.000 09:23  
Control Test 0.079 09:23  
Air Blank 0.000 09:24  
Control Test 0.079 09:24  
Air Blank 0.000 09:25  
Control Test 0.080 09:25  
Air Blank 0.000 09:25

Control Test Stats  
Average 0.0793  
Std Dev 0.0005  
Rel Std Dev(%) 0.7277

*WLL*  
Operator's Signature

*WLL*  
Operator's Signature

*WLL*  
Operator's Signature

*WLL*  
Operator's Signature