



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Jacksonville

S/N 80-001045

Date In 10/5/17

Date Out 10/18/17

Ship P/U H/D CMI EE

Intake Performed By DD

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
ok Case ok Handle
ok Dry Gas Holder ok Feet
ok Keyboard/Plug ok Back/Plugs
ok Screws tight ok Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By DD

Lab Temp °C 22.1
 Breath Tube Screen
 Replace External O-Rings
 Instrument Set Up Verified
 R-Value 182
 Flow Verification (L/s)
 Flow Column # ATP105
 32mm 156 (.139 - .169)
 36mm 179 (.156 - .190)
 53mm 244 (.228 - .278)
 103mm 519 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28421

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>G2835</u>	<u>201707D</u> <u>7/25/19</u>
0.08	<u>DR1279</u>	<u>201707E</u> <u>7/25/19</u>
0.20	<u>SD1025</u>	<u>201707C</u> <u>7/24/19</u>
0.08 DGS	N/A	<u>AG619605</u> <u>7/14/18</u>

Flow Calibration Performed By DD

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Quality Checks Cont. Performed By DD

Simulator Temperatures °C
 External Digital Therm. ID#: 300504
 34°C +- .2 Serial #: G2835
 34°C +- .2 Serial #: DR1279
 34°C +- .2 Serial #: SD1025

RECEIVED
OCT 19 2017
FDLE
Alcohol Testing Program

Calibration Adjustment Performed By _____

Calibration Adjustment N/A
 Calibration Adjustment Complete

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By DD

Barometric Pressure 1019 Gauge
 ID# 28421 1018 Instrument

Mouth Alcohol Solution Lot # 2016-C
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	<u>G2880</u>
Interferent	<u>G2840</u>
0.05	<u>G2835</u>
0.08	<u>DR1279</u>
0.20	<u>SD1025</u>

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Calibration Adjustment
 Post-Stability Tests
 Other _____

Notes/Suggested Service:
QA/QC v SP

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Quality Control Review _____ Date 10/19/17

80-001045
 Stability Checks
 10/19/17

INTOXILYZER 8000
 Instrument Initialization
 07:06 10/19/2017

JACKSONVILLE SO
 Intoxilyzer - Alcohol Analyzer SN 80-001045
 Model 8000
 10/19/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:38
Control Test	0.078	07:39
Air Blank	0.000	07:39
Control Test	0.079	07:40
Air Blank	0.000	07:41
Control Test	0.078	07:41
Air Blank	0.000	07:42
Control Test Stats		
Average	0.0783	
Std Dev	0.0016	
Rel Std Dev(%)	0.7370	


 Operator's Signature

10/19/17
 JS

JACKSONVILLE SO
 Intoxilyzer - Alcohol Analyzer SN 80-001045
 Model 8000
 10/19/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:43
Control Test	0.048	07:44
Air Blank	0.000	07:44
Control Test	0.048	07:45
Air Blank	0.000	07:45
Control Test	0.049	07:46
Air Blank	0.000	07:47
Control Test Stats		
Average	0.0483	
Std Dev	0.0016	
Rel Std Dev(%)	1.1945	


 Operator's Signature

JACKSONVILLE SO
 Intoxilyzer - Alcohol Analyzer SN 80-001045
 Model 8000
 10/19/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:48
Control Test	0.080	07:48
Air Blank	0.000	07:49
Control Test	0.080	07:49
Air Blank	0.000	07:50
Control Test	0.079	07:50
Air Blank	0.000	07:50
Control Test Stats		
Average	0.0797	
Std Dev	0.0016	
Rel Std Dev(%)	0.7247	

JS


 Operator's Signature

JACKSONVILLE SO
 Intoxilyzer - Alcohol Analyzer SN 80-001045
 Model 8000
 10/19/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:53
Control Test	0.199	07:54
Air Blank	0.000	07:54
Control Test	0.199	07:55
Air Blank	0.000	07:56
Control Test	0.200	07:56
Air Blank	0.000	07:57
Control Test Stats		
Average	0.1993	
Std Dev	0.0016	
Rel Std Dev(%)	0.2896	


 Operator's Signature

JS