

INSTRUMENT PROCESSING SHEET

Agency Clay County

S/N 1046

Date In 2/13/17

Date Out 2/9/17

Ship P/U H/D CMI EE

Intake Performed By DD

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
 Case Handle
 Dry Gas Holder Feet
 Keyboard/Plug Back/Plugs
 Screws tight Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Logic Bay

Notes: _____

Quality Checks Performed By BB

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 171
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm 154 (.139 - .169)
 36mm 178 (.156 - .190)
 53mm 259 (.228 - .278)
 103mm 502 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427 28421

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	G11739	201603D 3/8/18
0.08	G8149	201601F 1/26/18
0.20	G11621	201604C 4/5/18
0.08 DGS	N/A	AG626605 9/22/18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 10mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 FEB 10 2017
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By BB

Barometric Pressure
 ID# 28427 1016 Gauge
1014 Instrument

Mouth Alcohol Solution Lot # 2016-C
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	G2879
Interferent	G8144
0.05	G11739
0.08	G8149
0.20	G11621

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: QA/OC OK UGSM 2/9/2017

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Kurland
 Quality Control Review

2/10/17
 Date

80-001046
 Stability Checks
 2/9/17

INTOXILYZER 8000
 Instrument Initialization
 07:45 02/09/2017

CLAY COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001046
 02/09/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:09
Control Test	0.081	08:09
Air Blank	0.000	08:10
Control Test	0.081	08:10
Air Blank	0.000	08:11
Control Test	0.081	08:11
Air Blank	0.000	08:12
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DS

Operator's Signature

CLAY COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001046
 02/09/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:13
Control Test	0.049	08:14
Air Blank	0.000	08:14
Control Test	0.049	08:15
Air Blank	0.000	08:16
Control Test	0.050	08:16
Air Blank	0.000	08:17
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

Operator's Signature

CLAY COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001046
 02/09/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:21
Control Test	0.200	08:22
Air Blank	0.000	08:22
Control Test	0.199	08:23
Air Blank	0.000	08:23
Control Test	0.199	08:24
Air Blank	0.000	08:25
Control Test Stats		
Average	0.1993	
Std Dev	0.0006	
Rel Std Dev(%)	0.2896	

Operator's Signature

CLAY COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001046
 02/09/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:27
Control Test	0.079	08:28
Air Blank	0.000	08:28
Control Test	0.079	08:29
Air Blank	0.000	08:29
Control Test	0.080	08:30
Air Blank	0.000	08:31
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DS

BK

Operator's Signature