



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Polk CountyS/N 80-001055Date In 4/12/17Date Out 5/9/17☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake ☐ Registration ☒ Annual ☐ Return from CMI ☒ Return from Enforcement Electronics ☐ Other _____ Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: ☒ Power cord ☐ Printer Cable ☐ Other _____ Notes: _____ _____ _____ _____ _____	Quality Checks Performed By <u>SPS</u> ☒ Breath Tube Screen ☒ Replace O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>197</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP102</u> 32mm <u>0.140</u> (.139 - .169) 36mm <u>0.160</u> (.156 - .190) 53mm <u>0.230</u> (.228 - .278) 103mm <u>0.492</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28427</u> ☒ Stability Checks <u>4/12/17</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td><u>DR2035</u></td> <td><u>201603D</u> <u>3/8/18</u></td> </tr> <tr> <td>0.08</td> <td><u>SD3964</u></td> <td><u>201611B</u> <u>11/15/18</u></td> </tr> <tr> <td>0.20</td> <td><u>SD1025</u></td> <td><u>201604C</u> <u>4/5/18</u></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td><u>AG626004</u> <u>9/22/18</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	<u>DR2035</u>	<u>201603D</u> <u>3/8/18</u>	0.08	<u>SD3964</u>	<u>201611B</u> <u>11/15/18</u>	0.20	<u>SD1025</u>	<u>201604C</u> <u>4/5/18</u>	0.08 DGS	N/A	<u>AG626004</u> <u>9/22/18</u>	Flow Calibration Performed By _____ ☒ Flow Calibration N/A ☐ Flow Calibration Complete Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547) Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Suggested Service _____ _____ _____
Simulator	Serial #	Lot #/Exp															
0.05	<u>DR2035</u>	<u>201603D</u> <u>3/8/18</u>															
0.08	<u>SD3964</u>	<u>201611B</u> <u>11/15/18</u>															
0.20	<u>SD1025</u>	<u>201604C</u> <u>4/5/18</u>															
0.08 DGS	N/A	<u>AG626004</u> <u>9/22/18</u>															

RECEIVED
MAY 09 2017
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By <u>SPS</u> ☐ Optical Bench Calibration N/A ☒ Optical Bench Calibration Complete Barometric Pressure Gauge <u>1016</u> ID # <u>28421</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>SD1016</u></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td><u>SD1024</u></td> <td><u>16108</u></td> <td><u>12/8/18</u></td> </tr> <tr> <td>0.100</td> <td><u>G11739</u></td> <td><u>16001</u></td> <td><u>5/8/18</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1012</u></td> <td><u>16103</u></td> <td><u>6/14/18</u></td> </tr> <tr> <td>0.400</td> <td><u>G4444</u></td> <td><u>16102</u></td> <td><u>3/22/18</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>15615080A2</u></td> <td><u>7/5/17</u></td> </tr> </tbody> </table> ☒ Post Calibration Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td><u>G3709</u></td> <td><u>201603D</u></td> <td><u>3/8/18</u></td> </tr> <tr> <td>0.08</td> <td><u>DR1279</u></td> <td><u>201611B</u></td> <td><u>11/15/18</u></td> </tr> <tr> <td>0.20</td> <td><u>DR3856</u></td> <td><u>201604C</u></td> <td><u>4/5/18</u></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td><u>AG626004</u></td> <td><u>9/22/18</u></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	<u>SD1016</u>	N/A	N/A	0.040	<u>SD1024</u>	<u>16108</u>	<u>12/8/18</u>	0.100	<u>G11739</u>	<u>16001</u>	<u>5/8/18</u>	0.200	<u>SD1012</u>	<u>16103</u>	<u>6/14/18</u>	0.400	<u>G4444</u>	<u>16102</u>	<u>3/22/18</u>	0.080 DGS	N/A	<u>15615080A2</u>	<u>7/5/17</u>	Simulator	Serial Number	Lot Number	Expiration	0.05	<u>G3709</u>	<u>201603D</u>	<u>3/8/18</u>	0.08	<u>DR1279</u>	<u>201611B</u>	<u>11/15/18</u>	0.20	<u>DR3856</u>	<u>201604C</u>	<u>4/5/18</u>	0.08 DGS	N/A	<u>AG626004</u>	<u>9/22/18</u>	Department Inspection Performed By <u>SPS</u> ☒ Barometric Pressure <u>1017</u> Gauge ID# <u>26932</u> <u>1016</u> Instrument Mouth Alcohol Solution Lot # <u>2016-C</u> Acetone Stock Solution Lot # <u>2017-A</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td><u>G2880</u></td> </tr> <tr> <td>Interferent</td> <td><u>G2834</u></td> </tr> <tr> <td>0.05</td> <td><u>G3709</u></td> </tr> <tr> <td>0.08</td> <td><u>DR1279</u></td> </tr> <tr> <td>0.20</td> <td><u>DR3856</u></td> </tr> </tbody> </table>	Simulator	Serial Number	0.00	<u>G2880</u>	Interferent	<u>G2834</u>	0.05	<u>G3709</u>	0.08	<u>DR1279</u>	0.20	<u>DR3856</u>
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000	<u>SD1016</u>	N/A	N/A																																																										
0.040	<u>SD1024</u>	<u>16108</u>	<u>12/8/18</u>																																																										
0.100	<u>G11739</u>	<u>16001</u>	<u>5/8/18</u>																																																										
0.200	<u>SD1012</u>	<u>16103</u>	<u>6/14/18</u>																																																										
0.400	<u>G4444</u>	<u>16102</u>	<u>3/22/18</u>																																																										
0.080 DGS	N/A	<u>15615080A2</u>	<u>7/5/17</u>																																																										
Simulator	Serial Number	Lot Number	Expiration																																																										
0.05	<u>G3709</u>	<u>201603D</u>	<u>3/8/18</u>																																																										
0.08	<u>DR1279</u>	<u>201611B</u>	<u>11/15/18</u>																																																										
0.20	<u>DR3856</u>	<u>201604C</u>	<u>4/5/18</u>																																																										
0.08 DGS	N/A	<u>AG626004</u>	<u>9/22/18</u>																																																										
Simulator	Serial Number																																																												
0.00	<u>G2880</u>																																																												
Interferent	<u>G2834</u>																																																												
0.05	<u>G3709</u>																																																												
0.08	<u>DR1279</u>																																																												
0.20	<u>DR3856</u>																																																												

Notes: <u>Please change menu level 2 password to something unique. Optical bench calibration adjustment performed to bring values closer to nominal.</u> <u>QA/QC OK SP</u> <u>JJ Graham</u>	Attachments ☒ Form 41 ☒ Pre-Stability Tests ☐ Flow Calibration ☒ Optical Bench Cal ☒ Post-Stability Tests ☒ Other <u>Form 40</u> Instrument Compliance ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use <u>5/9/17</u>
--	--

Quality Control Review

Date

Revised November 2012

Stability Checks #80-001055 Polk County SO. 4/12/17 *DWS*

POLK COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001055
04/12/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:51
Control Test	0.049	14:51
Air Blank	0.000	14:52
Control Test	0.048	14:53
Air Blank	0.000	14:53
Control Test	0.048	14:54
Air Blank	0.000	14:55
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	

DWS
Operator's Signature

POLK COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001055
04/12/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:58
Control Test	0.079	14:59
Air Blank	0.000	15:00
Control Test	0.079	15:00
Air Blank	0.000	15:01
Control Test	0.078	15:02
Air Blank	0.000	15:02
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

DWS
Operator's Signature

POLK COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001055
04/12/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:04
Control Test	0.197	15:05
Air Blank	0.000	15:05
Control Test	0.193	15:06
Air Blank	0.000	15:07
Control Test	0.193	15:07
Air Blank	0.000	15:08
Control Test Stats		
Average	0.1943	
Std Dev	0.0023	
Rel Std Dev(%)	1.1884	

DWS
Operator's Signature

POLK COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001055
04/12/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:09
Control Test	0.193	15:10
Air Blank	0.000	15:10
Control Test	0.192	15:11
Air Blank	0.000	15:12
Control Test	0.192	15:12
Air Blank	0.000	15:13
Control Test Stats		
Average	0.1923	
Std Dev	0.0006	
Rel Std Dev(%)	0.3002	

DWS
Operator's Signature

SP

Stopped quality checks
due to room conditions. Will recheck
at a later date. *DWS*

Pre-cal adjustment

Stability Checks #80-001055 Polk County SO. 5/3/17 RWS

63

POLK COUNTY SO
Intoxilizer - Alcohol Analyzer
Model 8000
05/03/2017
Software: 8100.27

POLK COUNTY SO
Intoxilizer - Alcohol Analyzer
Model 8000
05/03/2017
Software: 8100.27

POLK COUNTY SO
Intoxilizer - Alcohol Analyzer
Model 8000
05/03/2017
Software: 8100.27

POLK COUNTY SO
Intoxilizer - Alcohol Analyzer
Model 8000
05/03/2017
Software: 8100.27

POLK COUNTY SO
Intoxilizer - Alcohol Analyzer
Model 8000
05/03/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:34
Control Test	0.048	14:35
Air Blank	0.000	14:36
Control Test	0.049	14:36
Air Blank	0.000	14:37
Control Test	0.048	14:38
Air Blank	0.000	14:38
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	

Test	g/210L	Time
Air Blank	0.000	14:40
Control Test	0.079	14:41
Air Blank	0.000	14:41
Control Test	0.078	14:42
Air Blank	0.000	14:42
Control Test	0.078	14:43
Air Blank	0.000	14:44
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

Test	g/210L	Time
Air Blank	0.000	14:29
Control Test	0.194	14:30
Air Blank	0.000	14:31
Control Test	0.194	14:31
Air Blank	0.000	14:32
Control Test	0.195	14:33
Air Blank	0.000	14:33
Control Test Stats		
Average	0.1943	
Std Dev	0.0006	
Rel Std Dev(%)	0.2971	

Test	g/210L	Time
Air Blank	0.000	14:49
Control Test	0.196	14:50
Air Blank	0.000	14:51
Control Test	0.195	14:51
Air Blank	0.000	14:52
Control Test	0.195	14:53
Air Blank	0.000	14:53
Control Test Stats		
Average	0.1953	
Std Dev	0.0006	
Rel Std Dev(%)	0.2956	

Test	g/210L	Time
Air Blank	0.000	14:57
Control Test	0.076	14:57
Air Blank	0.000	14:58
Control Test	0.077	14:58
Air Blank	0.000	14:58
Control Test	0.078	14:59
Air Blank	0.000	14:59
Control Test Stats		
Average	0.0770	
Std Dev	0.0010	
Rel Std Dev(%)	1.2987	

G3709

DR1279

DR3856

DR3856

SP

RWS

Operator's Signature

RWS

Operator's Signature

RWS

Operator's Signature

RWS

Operator's Signature

RWS

Operator's Signature

201603D
3/8/18

201611B
11/15/18

201604C
4/5/18
Refer to check
consistency. RWS

201604C
4/5/18

AG626605
9/22/18

Polk County S.O. 5/9/17 ~~ONS~~

POLK COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001055
05/09/2017 09:49:25

Auto Calibration
Max Power Res Value = 20
Auto Range Res Value = 12

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%%

Samples Taken = 4, Discarded = 1
Sum IO = 12869, 9um IO = 13938

<<<< CHANNEL ; >>>>

Sample #	% Abs	(% Abs Ref)
Sample #1	= 0.0880	(-0.0250)
Sample #2	= 0.0680	(-0.0050)
Sample #3	= 0.0970	(0.0010)
Sample #4	= 0.0770	(0.0160)
Avg % Abs	= 0.0807	(0.0037)
STD DEV	= 0.0148	(0.0112)
REL STD DEV	= 18.40	(305.540)

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	0.1350	(-0.0020)
Sample #2 =	0.1210	(0.0000)
Sample #3 =	0.1190	(0.0140)
Sample #4 =	0.1300	(0.0060)
Avg. % Abs =	0.1233	(0.0067)
STD DEV =	0.0059	(0.0070)
REL STD DEV =	4.751	(105.357)

Sol Value = 0.040 a/210L ***

Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12867, Sum Io = 13936

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	0.7740	(-0.0130)
Sample #2 =	0.7600	(0.0030)
Sample #3 =	0.7730	(0.0240)
Sample #4 =	0.7630	(0.0140)
Avg % Abs =	0.7653	(0.0137)
STD DEV =	0.0068	(0.0105)
REL STD DEV =	0.889	(76.855)

```
<<<< CHANNEL 2 >>>>
Sample      % Abs      (% Abs Ref)
Sample #1 = 1.5400   (-0.0150)
Sample #2 = 1.5210   (0.0000)
Sample #3 = 1.5390   (0.0050)
Sample #4 = 1.5280   (0.0000)
Avg % Abs = 1.5293   (0.0017)
STD DEV = 0.0091   (0.0029)
REL STD DEV = 0.593   (173.205)
```

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12861, Sum Io = 13932

```

<<<< CHANNEL 1 >>>>
Sample      % Abs      (% Abs Ref)
Sample #1 = 1.8660   [-0.0210]
Sample #2 = 1.8540   [0.0190]
Sample #3 = 1.8430   [0.0230]
Sample #4 = 1.8560   [0.0350]
Avg % Abs = 1.8543  [0.0257]
STD DEV = 0.0115   [0.0083]
REF STD DEV = 0.620 [32.442]

```

```

<<<< CHANNEL 2 >>>>
Sample      % Abs      (% Abs Ref)
Sample #1 = 3.7200   (-0.0240)
Sample #2 = 3.7290   (0.0070)
Sample #3 = 3.7000   (0.0050)
Sample #4 = 3.7030   (0.0290)
Avg % Abs = 3.7107   (0.0137)
STD DEV = 0.0159   (0.0133)
REL STD DEV = 0.430   (97.43%)

```

```
Sol Value = 0.200 g/210L ***
Fit value = 0.9524 ng/l 22%%
Samples Taken = 4, Discarded = 1
Sum Io = 12855, Num Io = 13931

<<<< CHANNEL 1 >>>>

Sample      % Abs      (% Abs Ref)
Sample #1 = 3.5630      (-0.0160)
Sample #2 = 3.5400      (0.0220)
Sample #3 = 3.5580      (0.0040)
Sample #4 = 3.5490      (0.0340)
Avg % Abs = 3.5490      (0.0200)
STD DEV = 0.0090      (0.0151)
REL STD DEV = 0.254      (75.498)
```

```

<<<< CHANNEL 2 >>>>
Sample      % Abs      (% Abs Ref)
Sample #1 = 7.1120   (-0.0110)
Sample #2 = 7.0720   (0.0360)
Sample #3 = 7.0630   (0.0570)
Sample #4 = 7.0710   (0.0560)
Aug % Abs = 7.0687  (0.0497)
STD DEV = 0.0049  (0.0118)
REL STD DEV = 0.070  (23.851)

```

```
Sol value = 0.400 g/210L ***
Fit value = 1.9048 mg/l 222%
Samples Taken = 4, Discarded = 1
Sum Io = 12852, Num Io = 13930
<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 6.8580 (-0.0110)
Sample #2 = 6.8150 (0.0320)
Sample #3 = 6.8010 (0.0610)
Sample #4 = 6.8100 (0.0570)
Avg % Abs = 6.8087 (0.0500)
STD DEV = 0.0071 (0.0157)
REL STD DEV = 0.104 (31.432)
```

```

<<<< CHANNEL 2 >>>>
Sample      (% Abs Ref)
% Abs
Sample #1 = 13.3500 (0.0020)
Sample #2 = 13.2720 (0.1160)
Sample #3 = 13.2470 (0.1410)
Sample #4 = 13.2450 (0.1500)
Sample #5 = 13.2547 (0.1357)

BIU DEV = 0.0000 0.0000
REL STD DEV = 0.113 (12.985)

```

```

*****
<<<<< CHANNEL 1 >>>>>
Sol Val = 0.0000 ng/1 or 0.000 g/210L
% Abs = 0.081
Std Dev = 0.01 Rel Std Dev = 18.40
Sol Val = 0.1905 ng/1 or 0.040 g/210L
% Abs = 0.765
Std Dev = 0.01 Rel Std Dev = 0.69
Sol Val = 0.4762 ng/1 or 0.100 g/210L
% Abs = 1.854
Std Dev = 0.01 Rel Std Dev = 0.62
Sol Val = 0.9524 ng/1 or 0.200 g/210L
% Abs = 3.549
Std Dev = 0.01 Rel Std Dev = 0.25
Sol Val = 1.9048 ng/1 or 0.400 g/210L
% Abs = 6.809
Std Dev = 0.01 Rel Std Dev = 0.10
Zero Order Coef = -184.53
First Order Coef = 2634.64
Second Order Coef = 27.90
Standard Deviation = 36.273617

```

```

<<<<< CHANNEL 2 >>>>>
  1.001 Ual = 0.0000 mg/l or 0.000 g/210L
    % Abs = 0.123
  1.000 Std Dev = 0.01 Rel Std Dev = 4.75
  0.001 Ual = 0.1905 mg/l or 0.040 g/210L
    % Abs = 1.529
  0.000 Std Dev = 0.01 Rel Std Dev = 0.59
  0.001 Ual = 0.4762 mg/l or 0.100 g/210L
    % Abs = 3.711
  0.000 Std Dev = 0.02 Rel Std Dev = 0.43
  0.001 Ual = 0.9524 mg/l or 0.200 g/210L
    % Abs = 7.069
  0.000 Std Dev = 0.00 Rel Std Dev = 0.07
  0.001 Ual = 1.9048 mg/l or 0.400 g/210L
    % Abs = 13.255
  0.000 Std Dev = 0.02 Rel Std Dev = 0.11
  Zero Order Coef = -126.50
  First Order Coef = 1274.09
  Second Order Coef = 13.00
  Standard Deviation = 31.865787

```

Solution Stats Quadratic Fit Chan 1		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.001	-0.0006
0.040	0.039	0.0012
0.100	0.101	-0.0007
0.200	0.200	0.0001
0.400	0.400	0.0000

Solution Stats Quadratic Fit Chan 2			
Act	Fit	Residual	
g/210L	g/210L	g/210L	
0.000	0.001	-0.0006	
0.040	0.039	0.0011	
0.100	0.100	-0.0004	
0.200	0.200	-0.0001	
0.400	0.400	0.0001	

```

501 value = 0.080 g/210L ***
:it value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
***** CHANNEL 1 *****
Sample #1 = 3158.00
Sample #2 = 3144.00
Sample #3 = 3172.00
Sample #4 = 3187.00
Average Result = 3167.6667
STD DEV = 21.8251
REL STD DEV = 0.689
*****
***** CHANNEL 2 *****
Sample #1 = 3350.00
Sample #2 = 3357.00
Sample #3 = 3363.00
Sample #4 = 3373.00
Average Result = 3364.3333
STD DEV = 8.0829
REL STD DEV = 0.240

```

```

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1016
3 um H2O Adjust (mg/1*10,000) = 642
9 um H2O Adjust (mg/1*10,000) = 445
*** AUTO CAL PASS

```

Post-Cal Adjust
Stability Checks #80-001055 Polk County S.O. 5/9/17 ~~06~~

DS

POLK COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001055
05/09/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:08
Control Test	0.049	11:09
Air Blank	0.000	11:10
Control Test	0.050	11:10
Air Blank	0.000	11:11
Control Test	0.049	11:11
Air Blank	0.000	11:12
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	



Operator's Signature

POLK COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001055
05/09/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:58
Control Test	0.081	10:58
Air Blank	0.000	10:59
Control Test	0.080	11:00
Air Blank	0.000	11:00
Control Test	0.079	11:01
Air Blank	0.000	11:01
Control Test Stats		
Average	0.0800	
Std Dev	0.0010	
Rel Std Dev(%)	1.2500	



Operator's Signature

POLK COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001055
05/09/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:03
Control Test	0.199	11:04
Air Blank	0.000	11:04
Control Test	0.197	11:05
Air Blank	0.000	11:06
Control Test	0.198	11:06
Air Blank	0.000	11:07
Control Test Stats		
Average	0.1980	
Std Dev	0.0010	
Rel Std Dev(%)	0.5051	



Operator's Signature

POLK COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001055
05/09/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:54
Control Test	0.080	10:55
Air Blank	0.000	10:55
Control Test	0.081	10:55
Air Blank	0.000	10:56
Control Test	0.081	10:56
Air Blank	0.000	10:57
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	



Operator's Signature

SP

RECEIVED

Florida Department of Law Enforcement

Alcohol Testing Program

MAY 09 2017
FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: POLK COUNTY SO
Time of Inspection: 14:49

Date of Inspection: 04/12/2017

Serial Number: 80-001055
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____

Number of Simulators Used: _____

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT. NOT A COMPLIANCE CHECK.

N/A Compliance Not
Determined

JO

SP

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Danielle M Bell

DANIELLE M BELL

Signature and Printed Name

04/12/2017
Date