



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Gulfport P.D. S/N 80-001061
Date In 6/2/17 Date Out 6/5/17 Ship P/U H/D CMI EE

Intake Performed By [Signature]

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By [Signature]

Lab Temp °C 23.8
 Breath Tube Screen
 Replace External O-Rings
 Instrument Set Up Verified
 R-Value 138
 Flow Verification (L/s)
 Flow Column # ATP-102
 32mm 0.156 (.139 - .169)
 36mm 0.167 (.156 - .190)
 53mm 0.238 (.228 - .278)
 103mm 0.511 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

| Simulator | Serial # | Lot #/Exp |
|-----------|---------------|-----------------------------------|
| 0.05 | <u>SD3962</u> | <u>201603D</u> <u>3/8/18</u> |
| 0.08 | <u>SD3964</u> | <u>201611B</u> <u>1/5/18</u> |
| 0.20 | <u>SD3933</u> | <u>201604C</u> <u>4/5/18</u> |
| 0.08 DGS | N/A | <u>AG626604</u> <u>9/22/18</u> |

Flow Calibration Performed By [Signature]

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Quality Checks Cont. Performed By [Signature]

Simulator Temperatures °C
 External Digital Therm. ID#: 300505
 34°C +/- .2 Serial #: SD3962
 34°C +/- .2 Serial #: SD3964
 34°C +/- .2 Serial #: SD3933

RECEIVED JUN 05 2017 FDLE Alcohol Testing Program

Calibration Adjustment Performed By _____

Calibration Adjustment N/A
 Calibration Adjustment Complete

Barometric Pressure Gauge _____ ID # _____

| Simulator | Serial Number | Lot Number | Expiration |
|-----------|---------------|------------|------------|
| 0.000 | | N/A | N/A |
| 0.040 | | | |
| 0.100 | | | |
| 0.200 | | | |
| 0.300 | | | |
| 0.080 DGS | N/A | | |

Post Calibration Adjustment Stability Checks

| Simulator | Serial Number | Lot Number | Expiration |
|-----------|---------------|------------|------------|
| 0.05 | | | |
| 0.08 | | | |
| 0.20 | | | |
| 0.08 DGS | N/A | | |

Department Inspection Performed By [Signature]

Barometric Pressure 1009 Gauge
 ID# 28427 1005 Instrument

Mouth Alcohol Solution Lot # 2016-C
 Acetone Stock Solution Lot # 2017-A

| Simulator | Serial Number |
|-------------|---------------|
| 0.00 | <u>G2880</u> |
| Interferent | <u>G2834</u> |
| 0.05 | <u>SD3962</u> |
| 0.08 | <u>SD3964</u> |
| 0.20 | <u>SD3933</u> |

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Calibration Adjustment
 Post-Stability Tests
 Other _____

Notes/Suggested Service:
Please change level 2 menu password to something unique. [Signature]

[Signature] 6/5/17
[Signature] 6/5/17

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Quality Control Review

Date

Stability Checks # 80-001061 Gulfport P.D. 6/5/17 *DMS*

DMS

GULFPORT PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001061
06/05/2017
Software: 8100.27

| Test | 9/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 10:26 |
| Control Test | 0.081 | 10:26 |
| Air Blank | 0.000 | 10:27 |
| Control Test | 0.080 | 10:27 |
| Air Blank | 0.000 | 10:27 |
| Control Test | 0.080 | 10:28 |
| Air Blank | 0.000 | 10:28 |
| Control Test | 0.080 | 10:28 |
| Control Test Stats | | |
| Average | 0.0803 | |
| Std Dev | 0.0006 | |
| Rel Std Dev(%) | 0.7187 | |

DMS
Operator's Signature

GULFPORT PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001061
06/05/2017
Software: 8100.27

| Test | 9/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 10:21 |
| Control Test | 0.196 | 10:21 |
| Air Blank | 0.000 | 10:22 |
| Control Test | 0.196 | 10:22 |
| Air Blank | 0.000 | 10:23 |
| Control Test | 0.196 | 10:24 |
| Air Blank | 0.000 | 10:24 |
| Control Test | 0.000 | 10:24 |
| Control Test Stats | | |
| Average | 0.1960 | |
| Std Dev | 0.0000 | |
| Rel Std Dev(%) | 0.0000 | |

DMS
Operator's Signature

GULFPORT PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001061
06/05/2017
Software: 8100.27

| Test | 9/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 10:16 |
| Control Test | 0.078 | 10:16 |
| Air Blank | 0.000 | 10:17 |
| Control Test | 0.078 | 10:17 |
| Air Blank | 0.000 | 10:18 |
| Control Test | 0.078 | 10:19 |
| Air Blank | 0.000 | 10:19 |
| Control Test | 0.000 | 10:19 |
| Control Test Stats | | |
| Average | 0.0780 | |
| Std Dev | 0.0000 | |
| Rel Std Dev(%) | 0.0000 | |

DMS
Operator's Signature

GULFPORT PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001061
06/05/2017
Software: 8100.27

| Test | 9/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 10:11 |
| Control Test | 0.048 | 10:11 |
| Air Blank | 0.000 | 10:12 |
| Control Test | 0.048 | 10:13 |
| Air Blank | 0.000 | 10:13 |
| Control Test | 0.049 | 10:14 |
| Air Blank | 0.000 | 10:14 |
| Control Test | 0.000 | 10:14 |
| Control Test Stats | | |
| Average | 0.0483 | |
| Std Dev | 0.0006 | |
| Rel Std Dev(%) | 1.1945 | |

DMS
Operator's Signature

JF