



Alcohol Testing Program

### INSTRUMENT PROCESSING SHEET

Agency Clearwater PD S/N 80-001068  
 Date In 6/17/17 Date Out 6/5/17  
 Ship  P/U  H/D  CMI  EE

**Intake** Performed By [Signature]

Registration  
 Annual  
 Return from CMI  
 Return from Enforcement Electronics  
 Other \_\_\_\_\_

Visual Inspection:  
 Case OK  Handle OK  
 Dry Gas Holder OK  Feet OK  
 Keyboard/Plug OK  Back/Plugs OK  
 Screws tight OK  Breath Hose OK

Other Equipment:  
 Power cord  
 Printer Cable  
 Other \_\_\_\_\_

Notes: \_\_\_\_\_

**Quality Checks** Performed By [Signature]

Lab Temp °C 23.8  
 Breath Tube Screen  
 Replace External O-Rings  
 Instrument Set Up Verified  
 R-Value 218  
 Flow Verification (L/s)  
 Flow Column # ATP-102  
 32mm 0.144 (.139 - .169)  
 36mm 0.160 (.156 - .190)  
 53mm 0.234 (.228 - .278)  
 103mm 0.488 (.447 - .547)

Barometric Pressure Check  
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>SD39162</u>	<u>201603D</u> <u>3/8/18</u>
0.08	<u>SD3964</u>	<u>201601B</u> <u>11/5/18</u>
0.20	<u>SD3933</u>	<u>201604C</u> <u>4/5/18</u>
0.08 DGS	N/A	<u>AG12604</u> <u>9/22/18</u>

**Flow Calibration** Performed By [Signature]

Flow Calibration N/A  
 Flow Calibration Complete  
 Flow Column # \_\_\_\_\_  
 5L/min - 17mm  
 15L/min - 53mm  
 30L/min - 103mm  
 R-Value \_\_\_\_\_  
 Post Calibration Verification (L/s)  
 Flow Column # \_\_\_\_\_  
 32mm \_\_\_\_\_ (.139 - .169)  
 36mm \_\_\_\_\_ (.156 - .190)  
 53mm \_\_\_\_\_ (.228 - .278)  
 103mm \_\_\_\_\_ (.447 - .547)

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Quality Checks Cont.** Performed By [Signature]  
**Simulator Temperatures °C**  
 External Digital Therm. ID#: 300505  
 34°C +/- .2 Serial #: SD39162  
 34°C +/- .2 Serial #: SD3964  
 34°C +/- .2 Serial #: SD3933

RECEIVED  
 JUN 05 2017  
 FDLE  
 Alcohol Testing Program

**Calibration Adjustment** Performed By \_\_\_\_\_

Calibration Adjustment N/A  
 Calibration Adjustment Complete  
 Barometric Pressure Gauge \_\_\_\_\_ ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

**Department Inspection** Performed By [Signature]

Barometric Pressure 1009 Gauge  
 ID# 28427 1004 Instrument

Mouth Alcohol Solution Lot # 2016-C  
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	<u>G2880</u>
Interferent	<u>G2834</u>
0.05	<u>SD39162</u>
0.08	<u>SD3964</u>
0.20	<u>SD3933</u>

**Attachments**

Form 41  
 Pre-Stability Tests  
 Flow Calibration

Calibration Adjustment  
 Post-Stability Tests  
 Other \_\_\_\_\_

Notes/Suggested Service:  
QC [Signature] 6/5/17

[Signature] 6/5/17  
 Quality Control Review Date

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC  
 Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use

Stability Checks #80-001068 Clearwater P.D. 6/5/17 *RMS*

*RMS*

CLEARWATER PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001068  
06/05/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:05
Control Test	0.043	10:06
Air Blank	0.000	10:06
Control Test	0.050	10:07
Air Blank	0.000	10:07
Control Test	0.049	10:08
Air Blank	0.000	10:09
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

*RMS*

Operator's Signature

CLEARWATER PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001068  
06/05/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:10
Control Test	0.080	10:11
Air Blank	0.000	10:12
Control Test	0.080	10:12
Air Blank	0.000	10:13
Control Test	0.079	10:14
Air Blank	0.000	10:14
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

*RMS*

*RMS*

Operator's Signature

CLEARWATER PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001068  
06/05/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:15
Control Test	0.201	10:16
Air Blank	0.000	10:17
Control Test	0.200	10:17
Air Blank	0.000	10:18
Control Test	0.200	10:18
Air Blank	0.000	10:19
Control Test Stats		
Average	0.2003	
Std Dev	0.0006	
Rel Std Dev(%)	0.2882	

*RMS*

Operator's Signature

CLEARWATER PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001068  
06/05/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:20
Control Test	0.078	10:20
Air Blank	0.000	10:21
Control Test	0.078	10:21
Air Blank	0.000	10:22
Control Test	0.079	10:22
Air Blank	0.000	10:23
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

*RMS*

Operator's Signature

*JP*