

## INSTRUMENT PROCESSING SHEET

Agency Jacksonville SO S/N 80-001069  
 Date In 7/28/17 Date Out 8/30/17  Ship  P/U  H/D  CMI  EE

**Intake** Performed By [Signature]

Registration  
 Annual  
 Return from CMI  
 Return from Enforcement Electronics  
 Other \_\_\_\_\_

Visual Inspection:  
 Case ok Handle ok  
 Dry Gas Holder ok Feet ok  
 Keyboard/Plug ok Back/Plugs ok  
 Screws tight ok Breath Hose ok

Other Equipment:  
 Power cord  
 Printer Cable  
 Other Static Bag

Notes: \_\_\_\_\_

**Quality Checks** Performed By [Signature]

Lab Temp °C 21.9  
 Breath Tube Screen  
 Replace External O-Rings  
 Instrument Set Up Verified  
 R-Value 203  
 Flow Verification (L/s)  
 Flow Column # ATP102  
 32mm 156 (.139 - .169)  
 36mm 179 (.156 - .190)  
 53mm 242 (.228 - .278)  
 103mm 500 (.447 - .547)

Barometric Pressure Check  
 Gauge ID # 26932

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD3962	201707D 7/25/19
0.08	SD1013	201707E 7/25/19
0.20	SD3933	201707C 7/24/19
0.08 DGS	N/A	AK619605 7/14/18

**Flow Calibration** Performed By [Signature]

Flow Calibration N/A  
 Flow Calibration Complete  
 Flow Column # \_\_\_\_\_  
 5L/min - 17mm  
 15L/min - 53mm  
 30L/min - 103mm  
 R-Value \_\_\_\_\_  
 Post Calibration Verification (L/s)  
 Flow Column # \_\_\_\_\_  
 32mm \_\_\_\_\_ (.139 - .169)  
 36mm \_\_\_\_\_ (.156 - .190)  
 53mm \_\_\_\_\_ (.228 - .278)  
 103mm \_\_\_\_\_ (.447 - .547)

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Quality Checks Cont.** Performed By [Signature]

**Simulator Temperatures °C**  
 External Digital Therm. ID#: 300505  
 34°C +/- .2 Serial #: SD3962  
 34°C +/- .2 Serial #: SD1013  
 34°C +/- .2 Serial #: SD3933

**Calibration Adjustment** Performed By \_\_\_\_\_

Calibration Adjustment N/A  
 Calibration Adjustment Complete

Barometric Pressure Gauge \_\_\_\_\_ ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

**Department Inspection** Performed By [Signature]

Barometric Pressure 1013 Gauge  
 ID# 26932 1012 Instrument

Mouth Alcohol Solution Lot # 2016-C  
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	SD3962
0.08	SD1013
0.20	SD3933

**Attachments**

Form 41  
 Pre-Stability Tests  
 Flow Calibration  
 Calibration Adjustment  
 Post-Stability Tests  
 Other \_\_\_\_\_

Notes/Suggested Service:  
QA/PC OK QSPM 8/30/2017

[Signature] 8/30/17  
 Date

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use

Quality Control Review

RECEIVED  
 AUG 30 2017  
 FDLE  
 Alcohol Testing Program

80-001069  
 Stability Checks  
 8/30/17

*RAM*

INTOXILYZER 8000  
 Instrument Initialization  
 07:43 08/30/2017

JACKSONVILLE SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001069  
 08/30/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:52
Control Test	0.083	08:52
Air Blank	0.000	08:53
Control Test	0.083	08:54
Air Blank	0.000	08:54
Control Test	0.083	08:55
Air Blank	0.000	08:55
Control Test Stats		
Average	0.0830	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator Signature

JACKSONVILLE SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001069  
 08/30/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:01
Control Test	0.051	09:02
Air Blank	0.000	09:02
Control Test	0.051	09:03
Air Blank	0.000	09:03
Control Test	0.051	09:04
Air Blank	0.000	09:05
Control Test Stats		
Average	0.0510	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator Signature

JACKSONVILLE SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001069  
 08/30/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:09
Control Test	0.202	09:10
Air Blank	0.000	09:10
Control Test	0.201	09:11
Air Blank	0.000	09:11
Control Test	0.202	09:12
Air Blank	0.000	09:13
Control Test Stats		
Average	0.2017	
Std Dev	0.0006	
Rel Std Dev(%)	0.2863	

Operator Signature

*DCS*

Operator Signature

8/30/17  
*JA*