



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency F4P-Ocala

S/N 80-001070

Date In 1/13/17

Date Out 2/9/17

Ship P/U H/D CMI EE

Intake Performed By (Signature)

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By (Signature)

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 202
 Flow Verification (L/s)
 Flow Column # A78102
 32mm 160 (.139 - .169)
 36mm 18188 (.156 - .190)
 53mm 264 (.228 - .278)
 103mm 509 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427 28421

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	G11739	201603D 3/8/18
0.08	G8149	201601F 1/26/18
0.20	G11621	201604C 4/5/18
0.08 DGS	N/A	A8626605 9/22/18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
FEB 10 2017
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By (Signature)

Barometric Pressure 1016 Gauge
 ID# 28427 1017 Instrument

Mouth Alcohol Solution Lot # 2016-c
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	G2879
Interferent	G8144
0.05	G11739
0.08	G8149
0.20	G11621

Attachments

Form 41 Optical Bench Cal
 Pre-Stability Tests Post-Stability Tests
 Flow Calibration Other _____

Notes: QA/QC OK PGM 2/9/2017

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland

2/10/17

Quality Control Review

Date

8U-001070
 Stability Checks
 2/2/17

FL HIGHWAY PATROL
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001070
 02/09/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:11
Control Test	0.078	08:11
Air Blank	0.000	08:11
Control Test	0.077	08:12
Air Blank	0.000	08:12
Control Test	0.077	08:13
Air Blank	0.000	08:13
Control Test Stats		
Average	0.0773	
Std Dev	0.0006	
Rel Std Dev(%)	0.7466	

DOS

Operator's Signature

FL HIGHWAY PATROL
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001070
 02/09/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:19
Control Test	0.048	08:20
Air Blank	0.000	08:20
Control Test	0.049	08:21
Air Blank	0.000	08:22
Control Test	0.048	08:22
Air Blank	0.000	08:23
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	

Operator's Signature

FL HIGHWAY PATROL
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001070
 02/09/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:25
Control Test	0.194	08:25
Air Blank	0.000	08:26
Control Test	0.194	08:27
Air Blank	0.000	08:27
Control Test	0.195	08:28
Air Blank	0.000	08:28
Control Test Stats		
Average	0.1943	
Std Dev	0.0006	
Rel Std Dev(%)	0.2971	

Operator's Signature

FL HIGHWAY PATROL
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001070
 02/09/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:30
Control Test	0.078	08:30
Air Blank	0.000	08:31
Control Test	0.078	08:32
Air Blank	0.000	08:32
Control Test	0.078	08:33
Air Blank	0.000	08:33
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

QPM

MSK

Operator's Signature