



# Stability Checks #80-001074 Pasco County SO. 4/13/17 ~~RMS~~

*DOS*

PASCO COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001074  
06/13/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:32
Control Test	0.050	09:32
Air Blank	0.000	09:33
Control Test	0.049	09:33
Air Blank	0.000	09:34
Control Test	0.049	09:35
Air Blank	0.000	09:35
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

*RMS*  
Operator's Signature

*SD*

PASCO COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001074  
06/13/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:37
Control Test	0.079	09:38
Air Blank	0.000	09:38
Control Test	0.079	09:39
Air Blank	0.000	09:39
Control Test	0.080	09:40
Air Blank	0.000	09:40
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

*RMS*  
Operator's Signature

PASCO COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001074  
06/13/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:42
Control Test	0.200	09:43
Air Blank	0.000	09:44
Control Test	0.200	09:44
Air Blank	0.000	09:45
Control Test	0.199	09:46
Air Blank	0.000	09:46
Control Test Stats		
Average	0.1997	
Std Dev	0.0006	
Rel Std Dev(%)	0.2892	

*RMS*  
Operator's Signature

PASCO COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001074  
06/13/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:48
Control Test	0.078	09:48
Air Blank	0.000	09:49
Control Test	0.078	09:49
Air Blank	0.000	09:50
Control Test	0.078	09:50
Air Blank	0.000	09:51
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

*SP*

*RMS*  
Operator's Signature