



INSTRUMENT PROCESSING SHEET

Agency Florida Highway Patrol Miami

S/N 80-001075

Florida Department of Law Enforcement

Date In 11/30/2017 DI Completion Date 11/30/2017

Ship P/U H/D CMI EE

Intake Performed By <u>DELL</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>DELL</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>217</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>.152</u> (.139 - .169) 36 mm <u>.171</u> (.156 - .190) 53 mm <u>.242</u> (.228 - .278) 103 mm <u>.507</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28663</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD3967</td> <td>201707D 07/25/2019</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> <td>201707E 07/25/2019</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> <td>201707C 07/24/2019</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG626604 09/22/2018</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	SD3967	201707D 07/25/2019	0.080	SD3968	201707E 07/25/2019	0.200	SD3969	201707C 07/24/2019	0.080 DGS	N/A	AG626604 09/22/2018	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.050	SD3967	201707D 07/25/2019															
0.080	SD3968	201707E 07/25/2019															
0.200	SD3969	201707C 07/24/2019															
0.080 DGS	N/A	AG626604 09/22/2018															

Final Release Date

FDLE

DEC 08 2017

Alcohol Testing Program

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Temperature Checks Performed By DELL

Lab Temp °C 23.33c
 External Digital Therm. ID#: 300918
 34°C +- .2 Serial #: SD3967
 34°C +- .2 Serial #: SD3968
 34°C +- .2 Serial #: SD3969

Calibration Adjustment Performed By _____

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By DELL

Barometric Pressure ID# 68639
 Gauge 1018 Instrument 1019
 Mouth Alcohol Solution Lot # 2016-C
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.000	SD3965
Interferent	SD3966
0.050	SD3967
0.080	SD3968
0.200	SD3969

Attachments

Form 41 Flow Calibration
 Stability Checks Form 40
 Calibration Adjustment Other _____
 Post-Stability Checks

Notes/Suggested Service: Emailed **APPROVED**

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Pom 12/6/17 JJ Decker 12/8/17
 Tech Review / Date Admin Review / Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-001075	Florida Highway Patrol Miami	11/30/2017	<i>ML</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
SN: SD3967 Temp: 34.05c	SN: SD3968 Temp: 33.99c	SN: SD3969 Temp: 34.07c	Lot AG626604
0.047 to 0.053 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>	0.194 to 0.206 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>

Test	g/210L	Time	Test	g/210L	Time	Test	g/210L	Time
FHP TROOP E Intoxilyzer - Alconol Analyzer Model: 8000 11/30/2017 Software: 8100.27			FHP TROOP E Intoxilyzer - Alconol Analyzer Model: 8000 11/30/2017 Software: 8100.27			FHP TROOP E Intoxilyzer - Alconol Analyzer Model: 8000 11/30/2017 Software: 8100.27		
Air Blank	0.000	10:20	Air Blank	0.000	10:24	Air Blank	0.000	10:29
Control Test	0.049	10:20	Control Test	0.080	10:25	Control Test	0.198	10:30
Air Blank	0.000	10:21	Air Blank	0.000	10:25	Air Blank	0.000	10:30
Control Test	0.049	10:22	Control Test	0.081	10:26	Control Test	0.200	10:31
Air Blank	0.000	10:22	Air Blank	0.000	10:27	Air Blank	0.000	10:31
Control Test	0.049	10:23	Control Test	0.081	10:27	Control Test	0.200	10:32
Air Blank	0.000	10:23	Air Blank	0.000	10:28	Air Blank	0.000	10:33
Control Test Stats			Control Test Stats			Control Test Stats		
Average	0.0490		Average	0.0807		Average	0.1993	
Std Dev	0.0000		Std Dev	0.0006		Std Dev	0.0012	
Rel Std Dev(%)	0.0000		Rel Std Dev(%)	0.7157		Rel Std Dev(%)	0.5793	
<i>ML</i> Operator's Signature			<i>ML</i> Operator's Signature			<i>ML</i> Operator's Signature		

FBM

11/30/17
ML