

INSTRUMENT PROCESSING SHEET

Agency FHP - Parma City S/N 80-001111
 Date In 6/12/17 Date Out 6/19/2017 Ship P/U H/D CMI EE

Intake Performed By TDG

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By Q9M

Lab Temp °C 21.6
 Breath Tube Screen
 Replace External O-Rings
 Instrument Set Up Verified
 R-Value 188
 Flow Verification (L/s)
 Flow Column # NTP102
 32mm .144 (.139 - .169)
 36mm .160 (.156 - .190)
 53mm .234 (.228 - .278)
 103mm .467 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 26932

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>G11739</u>	<u>201603D</u> <u>3/8/18</u>
0.08	<u>G8149</u>	<u>201611B</u> <u>11/15/18</u>
0.20	<u>G11621</u>	<u>201604C</u> <u>4/15/18</u>
0.08 DGS	N/A	<u>AG626604</u> <u>9/22/18</u>

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (105)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Quality Checks Cont. Performed By Q9M
Simulator Temperatures °C
 External Digital Therm. ID#: 300505
 34°C +/- .2 Serial #: G11739
 34°C +/- .2 Serial #: G8149
 34°C +/- .2 Serial #: G11621

Calibration Adjustment Performed By _____

Calibration Adjustment N/A
 Calibration Adjustment Complete
 Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By Q9M

Barometric Pressure 1013 Gauge
 ID# 26932 1013 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	<u>G2879</u>
Interferent	<u>G8144</u>
0.05	<u>G11739</u>
0.08	<u>G8149</u>
0.20	<u>G11621</u>

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Calibration Adjustment
 Post-Stability Tests
 Other _____

Notes/Suggested Service: _____

Quality Control/Review 6/20/17 Date 6/20/17

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001111
06/19/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:18
Control Test	0.050	08:19
Air Blank	0.000	08:20
Control Test	0.049	08:20
Air Blank	0.000	08:21
Control Test	0.050	08:21
Air Blank	0.000	08:22
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

P Murphy

Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001111
06/19/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:13
Control Test	0.079	08:14
Air Blank	0.000	08:14
Control Test	0.079	08:15
Air Blank	0.000	08:15
Control Test	0.080	08:16
Air Blank	0.000	08:17
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

P Murphy

Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001111
06/19/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:24
Control Test	0.198	08:24
Air Blank	0.000	08:25
Control Test	0.197	08:26
Air Blank	0.000	08:26
Control Test	0.198	08:27
Air Blank	0.000	08:27
Control Test Stats		
Average	0.1977	
Std Dev	0.0006	
Rel Std Dev(%)	0.2921	

P Murphy

Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001111
06/19/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:38
Control Test	0.079	08:38
Air Blank	0.000	08:39
Control Test	0.079	08:39
Air Blank	0.000	08:40
Control Test	0.080	08:40
Air Blank	0.000	08:41
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DGS

P Murphy

Operator's Signature

JD