



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency FHP - CocoaS/N 80-001112Date In 2/6/17Date Out 2/22/17☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	Performed By <u>JP</u>	Quality Checks	Performed By <u>SP</u>	Flow Calibration	Performed By <u>SP</u>													
☐ Registration ☒ Annual ☐ Return from CMI ☐ Return from Enforcement Electronics ☐ Other _____ Visual Inspection: <u>ok</u> Case <u>ok</u> Handle <u>ok</u> Dry Gas Holder <u>ok</u> Feet <u>ok</u> Keyboard/Plug <u>ok</u> Back/Plugs <u>ok</u> Screws tight <u>ok</u> Breath Hose Other Equipment: ☐ Power cord ☐ Printer Cable ☒ Other <u>Static Bag</u> Notes: _____ _____ _____ _____	☒ Breath Tube Screen ☒ Replace O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>130</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP103</u> 32mm <u>.144</u> (.139 - .169) 36mm <u>.160</u> (.156 - .190) 53mm <u>.230</u> (.228 - .278) 103mm <u>.496</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28427</u> ☒ Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td><u>G3709</u></td> <td><u>201603D</u> <u>3-8-18</u></td> </tr> <tr> <td>0.08</td> <td><u>SD3964</u></td> <td><u>201601F</u> <u>1-26-18</u></td> </tr> <tr> <td>0.20</td> <td><u>DR3855</u></td> <td><u>201604C</u> <u>4-5-18</u></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td><u>AG626605</u> <u>9-22-18</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	<u>G3709</u>	<u>201603D</u> <u>3-8-18</u>	0.08	<u>SD3964</u>	<u>201601F</u> <u>1-26-18</u>	0.20	<u>DR3855</u>	<u>201604C</u> <u>4-5-18</u>	0.08 DGS	N/A	<u>AG626605</u> <u>9-22-18</u>	☒ Flow Calibration N/A ☐ Flow Calibration Complete Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)	Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Suggested Service _____ _____ _____
Simulator	Serial #	Lot #/Exp																
0.05	<u>G3709</u>	<u>201603D</u> <u>3-8-18</u>																
0.08	<u>SD3964</u>	<u>201601F</u> <u>1-26-18</u>																
0.20	<u>DR3855</u>	<u>201604C</u> <u>4-5-18</u>																
0.08 DGS	N/A	<u>AG626605</u> <u>9-22-18</u>																

Optical Bench Calibration Performed By SP
☐ Optical Bench Calibration N/A
☒ Optical Bench Calibration Complete
 Barometric Pressure Gauge 1004 ID # 28421

Simulator	Serial Number	Lot Number	Expiration
0.000	<u>SD1016</u>	N/A	N/A
0.040	<u>SD1024</u>	<u>16101</u>	<u>2-2-18</u>
0.100	<u>SD1013</u>	<u>16001</u>	<u>5-8-18</u>
0.200	<u>DR3856</u>	<u>16103</u>	<u>6-14-18</u>
0.400	<u>G2403</u>	<u>16102</u>	<u>3-22-18</u>
0.080 DGS	N/A	<u>03415080A2</u>	<u>3-5-17</u>

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05	<u>G3709</u>	<u>201603D</u>	<u>3-8-18</u>
0.08	<u>SD3964</u>	<u>201601F</u>	<u>1-26-18</u>
0.20	<u>DR3855</u>	<u>201604C</u>	<u>4-5-18</u>
0.08 DGS	N/A	<u>AG626605</u>	<u>9-22-18</u>

 Notes: PERFORMED OPTICAL BENCH CAL
TO BRING VALUES CLOSER TO NOMINAL
SP
QC: JDS
Department Inspection Performed By SP
☒ Barometric Pressure 1005 Gauge
 ID# 28427 1004 Instrument

 Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	<u>G2880</u>
Interferent	<u>G2834</u>
0.05	<u>G3709</u>
0.08	<u>SD3964</u>
0.20	<u>DR3855</u>

Attachments

☒ Form 41	☒ Optical Bench Cal
☒ Pre-Stability Tests	☒ Post-Stability Tests
☐ Flow Calibration	☐ Other _____

☒ Instrument Complies with Chapter 11D-8, FAC
☐ Instrument Does Not Comply with Chapter 11D-8, FAC
☒ Return to/Place into Evidentiary Use
☐ Remain Out of Evidentiary Use
☒ Conduct an Agency Inspection Before Evidentiary Use

Quality Control Review

Date

Revised November 2012

STABILITY CHECKS - INSTRUMENT # 80-001112

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000
02/16/2017
Software: 8100.27
SN 80-001112

Test	g/210L	Time
Air Blank	0.000	10:05
Control Test	0.051	10:05
Air Blank	0.000	10:06
Control Test	0.050	10:07
Air Blank	0.000	10:07
Control Test	0.050	10:08
Air Blank	0.000	10:08
Control Test Stats		
Average	0.0503	
Std Dev	0.0006	
Rel Std Dev(%)	1.1471	

SP

Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000
02/16/2017
Software: 8100.27
SN 80-001112

Test	g/210L	Time
Air Blank	0.000	09:59
Control Test	0.080	10:00
Air Blank	0.000	10:01
Control Test	0.080	10:01
Air Blank	0.000	10:02
Control Test	0.080	10:03
Air Blank	0.000	10:03
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000
02/16/2017
Software: 8100.27
SN 80-001112

Test	g/210L	Time
Air Blank	0.000	09:54
Control Test	0.201	09:55
Air Blank	0.000	09:56
Control Test	0.200	09:56
Air Blank	0.000	09:57
Control Test	0.199	09:58
Air Blank	0.000	09:58
Control Test Stats		
Average	0.2000	
Std Dev	0.0010	
Rel Std Dev(%)	0.5000	

SP

Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000
02/16/2017
Software: 8100.27
SN 80-001112

Test	g/210L	Time
Air Blank	0.000	09:51
Control Test	0.083	09:51
Air Blank	0.000	09:52
Control Test	0.083	09:52
Air Blank	0.000	09:53
Control Test	0.083	09:53
Air Blank	0.000	09:53
Control Test Stats		
Average	0.0830	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

73K

SP

Operator's Signature

FL HIGHWAY PATROL

Intoxilyzer - Alcohol Analyzer

Model 8000
02/22/2017
SN 80-001112
09:19:54

Auto Calibration

Max Power Res Value = 18
Auto Range Res Value = 12

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %
Samples Taken = 4, Discarded = 1
Sum Io = 12924, Sum Io = 14017

Channel 1

Sample % Abs (% Abs Ref)
Sample #1 = 0.0650 (-0.0040)
Sample #2 = 0.1010 (0.0120)
Sample #3 = 0.0740 (0.0650)
Sample #4 = 0.0910 (0.0580)
Avg % Abs = 0.0887 (0.0450)
STD DEV = 0.0137 (0.0288)
REL STD DEV = 15.395 (63.983)

Channel 2

Sample % Abs (% Abs Ref)
Sample #1 = 0.0610 (-0.0170)
Sample #2 = 0.1020 (-0.0560)
Sample #3 = 0.0770 (-0.0280)
Sample #4 = 0.0800 (-0.0230)
Avg % Abs = 0.0863 (-0.0357)
STD DEV = 0.0137 (0.0178)
REL STD DEV = 15.811 (49.867)

Sol Value

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %
Samples Taken = 4, Discarded = 1
Sum Io = 12909, Sum Io = 14016

Channel 1

Sample % Abs (% Abs Ref)
Sample #1 = 0.8180 (-0.0210)
Sample #2 = 0.8200 (0.0320)
Sample #3 = 0.8260 (0.0430)
Sample #4 = 0.8220 (0.0610)
Avg % Abs = 0.8227 (0.0453)
STD DEV = 0.0031 (0.0146)
REL STD DEV = 0.371 (32.294)

Channel 2

Sample % Abs (% Abs Ref)
Sample #1 = 1.5450 (-0.0100)
Sample #2 = 1.5130 (0.0280)
Sample #3 = 1.5050 (0.0330)
Sample #4 = 1.4980 (0.0520)
Avg % Abs = 1.5053 (0.0377)
STD DEV = 0.0075 (0.0127)
REL STD DEV = 0.499 (33.617)

Sol Value

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %
Samples Taken = 4, Discarded = 1
Sum Io = 12892, Sum Io = 14005

Channel 1

Sample % Abs (% Abs Ref)
Sample #1 = 1.8930 (-0.0220)
Sample #2 = 1.8860 (0.0290)
Sample #3 = 1.8940 (0.0370)
Sample #4 = 1.9130 (0.0260)
Avg % Abs = 1.8977 (0.0307)
STD DEV = 0.0139 (0.0057)
REL STD DEV = 0.731 (18.542)

Channel 2

Sample % Abs (% Abs Ref)
Sample #1 = 3.6020 (-0.0140)
Sample #2 = 3.6020 (0.0320)
Sample #3 = 3.6030 (0.0400)
Sample #4 = 3.6260 (0.0330)
Avg % Abs = 3.6103 (0.0350)
STD DEV = 0.0136 (0.0044)
REL STD DEV = 0.376 (12.454)

Sol Value

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %
Samples Taken = 4, Discarded = 1
Sum Io = 12896, Sum Io = 14005

Channel 1

Sample % Abs (% Abs Ref)
Sample #1 = 3.6740 (-0.0250)
Sample #2 = 3.6430 (0.0170)
Sample #3 = 3.6420 (0.0510)
Sample #4 = 3.6380 (0.0680)
Avg % Abs = 3.6410 (0.0453)
STD DEV = 0.0026 (0.0260)
REL STD DEV = 0.073 (57.282)

Channel 2

Sample % Abs (% Abs Ref)
Sample #1 = 7.0090 (-0.0260)
Sample #2 = 6.9450 (0.0610)
Sample #3 = 6.9470 (0.0880)
Sample #4 = 6.9390 (0.0940)
Avg % Abs = 6.9437 (0.0810)
STD DEV = 0.0042 (0.0176)
REL STD DEV = 0.060 (121.702)

Sol Value

Sol Value = 0.400 g/210L ***
Fit value = 1.9048 mg/l %
Samples Taken = 4, Discarded = 1
Sum Io = 12891, Sum Io = 14000

Channel 1

Sample % Abs (% Abs Ref)
Sample #1 = 7.0700 (-0.0070)
Sample #2 = 6.9900 (0.1150)
Sample #3 = 7.0060 (0.1200)
Sample #4 = 6.9660 (0.1310)
Avg % Abs = 6.9873 (0.1220)
STD DEV = 0.0201 (0.0082)
REL STD DEV = 0.288 (6.709)

Channel 2

Sample % Abs (% Abs Ref)
Sample #1 = 13.2750 (-0.0070)
Sample #2 = 13.1400 (0.1760)
Sample #3 = 13.1180 (0.1930)
Sample #4 = 13.1060 (0.1940)
Avg % Abs = 13.1213 (0.1877)
STD DEV = 0.0172 (0.0101)
REL STD DEV = 0.131 (5.390)

Sol Value

Sol Value = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.086
Std Dev = 0.01 Rel Std Dev = 15.81
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.505
Std Dev = 0.01 Rel Std Dev = 0.50
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.610
Std Dev = 0.01 Rel Std Dev = 0.38
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.944
Std Dev = 0.00 Rel Std Dev = 0.06
Sol Val = 1.9048 mg/l or 0.400 g/210L
% Abs = 13.121
Std Dev = 0.02 Rel Std Dev = 0.13
Zero Order Coef = -102.62
First Order Coef = 1305.63
Second Order Coef = 11.71
Standard Deviation = 9.56828

Auto Calibration

Max Power Res Value = 18
Auto Range Res Value = 12
Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %
Samples Taken = 4, Discarded = 1
Sum Io = 12924, Sum Io = 14017

Channel 1

Sample % Abs (% Abs Ref)
Sample #1 = 0.0650 (-0.0040)
Sample #2 = 0.1010 (0.0120)
Sample #3 = 0.0740 (0.0650)
Sample #4 = 0.0910 (0.0580)
Avg % Abs = 0.0887 (0.0450)
STD DEV = 0.0137 (0.0288)
REL STD DEV = 15.395 (63.983)

Channel 2

Sample % Abs (% Abs Ref)
Sample #1 = 0.0610 (-0.0170)
Sample #2 = 0.1020 (-0.0560)
Sample #3 = 0.0770 (-0.0280)
Sample #4 = 0.0800 (-0.0230)
Avg % Abs = 0.0863 (-0.0357)
STD DEV = 0.0137 (0.0178)
REL STD DEV = 15.811 (49.867)

Sol Value

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %
Samples Taken = 4, Discarded = 1
Sum Io = 12909, Sum Io = 14016

Channel 1

Sample % Abs (% Abs Ref)
Sample #1 = 0.8180 (-0.0210)
Sample #2 = 0.8200 (0.0320)
Sample #3 = 0.8260 (0.0430)
Sample #4 = 0.8220 (0.0610)
Avg % Abs = 0.8227 (0.0453)
STD DEV = 0.0031 (0.0146)
REL STD DEV = 0.371 (32.294)

Solution Stats Quadratic Fit Chan 1

Act Fit Residual
g/210L g/210L g/210L
0.000 -0.000 0.0001
0.040 0.040 -0.0001
0.100 0.100 -0.0001
0.200 0.200 0.0001
0.400 0.400 -0.0000

Solution Stats Quadratic Fit Chan 2

Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0002
0.040 0.040 0.0003
0.100 0.100 -0.0000
0.200 0.200 -0.0001
0.400 0.400 0.0001

Sol Value = 0.080 g/210L ***

Fit value = 0.3810 mg/l %

Samples Taken = 4, Discarded = 1

Channel 1

Sample #1 = 3038.00

Sample #2 = 3092.00

Sample #3 = 3041.00

Sample #4 = 3211.00

Average Result = 3114.6667

STD DEV = 87.2372

REL STD DEV = 2.801

Channel 2

Sample #1 = 3424.00

Sample #2 = 3457.00

Sample #3 = 3477.00

Sample #4 = 3524.00

Average Result = 3486.0000

STD DEV = 34.3948

REL STD DEV = 0.987

Channel 1

Dry Gas H2O Adjust Results

Barometric Pressure = 1004

3 um H2O Adjust (mg/l*10,000) = 695

9 um H2O Adjust (mg/l*10,000) = 323

Auto CAL PASS

OPTICAL BENCH CALIBRATION

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001112
02/22/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:12
Control Test	0.051	10:12
Air Blank	0.000	10:13
Control Test	0.051	10:14
Air Blank	0.000	10:14
Control Test	0.051	10:15
Air Blank	0.000	10:15
Control Test Stats		
Average	0.0510	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001112
02/22/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:08
Control Test	0.079	10:09
Air Blank	0.000	10:09
Control Test	0.080	10:09
Air Blank	0.000	10:10
Control Test	0.079	10:10
Air Blank	0.000	10:11
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DGS

SP

Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001112
02/22/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:04
Control Test	0.081	10:04
Air Blank	0.000	10:05
Control Test	0.080	10:06
Air Blank	0.000	10:06
Control Test	0.080	10:07
Air Blank	0.000	10:08
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

SP

Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001112
02/22/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:59
Control Test	0.201	09:59
Air Blank	0.000	10:00
Control Test	0.199	10:01
Air Blank	0.000	10:01
Control Test	0.198	10:02
Air Blank	0.000	10:03
Control Test Stats		
Average	0.1993	
Std Dev	0.0015	
Rel Std Dev(%)	0.7663	

SP

Operator's Signature

POST CALIBRATION STABILITY CHECKS
INSTRUMENT # 80-001112

RWS

BK