

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001115
12/01/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:14
Control Test	0.048	09:15
Air Blank	0.000	09:15
Control Test	0.048	09:16
Air Blank	0.000	09:17
Control Test	0.048	09:17
Air Blank	0.000	09:18
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001115
12/01/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:09
Control Test	0.079	09:09
Air Blank	0.000	09:10
Control Test	0.079	09:11
Air Blank	0.000	09:11
Control Test	0.079	09:12
Air Blank	0.000	09:12
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001115
12/01/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:03
Control Test	0.199	09:03
Air Blank	0.000	09:04
Control Test	0.199	09:04
Air Blank	0.000	09:05
Control Test	0.199	09:06
Air Blank	0.000	09:06
Control Test Stats		
Average	0.1990	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001115
12/01/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:33
Control Test	0.079	09:33
Air Blank	0.000	09:33
Control Test	0.078	09:34
Air Blank	0.000	09:34
Control Test	0.078	09:35
Air Blank	0.000	09:35
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

DGS



Operator's Signature

SP

12/5/17
JA