



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency FHP - Troop D S/N 80-001123

Date In 7/31/17 Date Out 11/17/17 Ship P/U H/D CMI EE

Intake Performed By [Signature]

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____
NOV 14 2017

Quality Checks Performed By [Signature]

Lab Temp °C 22.2
 Breath Tube Screen
 Replace External O-Rings
 Instrument Set Up Verified
 R-Value 209
 Flow Verification (L/s)
 Flow Column # ATP105
 32mm 152 (.139 - .169)
 36mm 171 (.156 - .190)
 53mm 238 (.228 - .278)
 103mm 515 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	G2835	201707D 7/25/19
0.08	DE1279	201707E 7/25/19
0.20	SD1025	201707C 7/24/19
0.08 DGS	N/A	AG16266015 9/22/18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By [Signature]

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Quality Checks Cont. Performed By [Signature]

Simulator Temperatures °C

External Digital Therm. ID#: 30504
 34°C +/- .2 Serial #: G2835
 34°C +/- .2 Serial #: DE1279
 34°C +/- .2 Serial #: SD1025

Calibration Adjustment Performed By _____

Calibration Adjustment N/A
 Calibration Adjustment Complete

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By [Signature]

Barometric Pressure _____ Gauge
 ID# 28427 _____ Instrument
1017
1014

Mouth Alcohol Solution Lot # 2016-C
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	G2580
Interferent	G2840
0.05	G2835
0.08	DE1279
0.20	SD1025

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Calibration Adjustment
 Post-Stability Tests
 Other Form 40
CMA Invoice for regulator

Notes/Suggested Service:
BA/OC's Regs 11/14/17

[Signature]
 Quality Control Review

11/14/17
 Date

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

80-001123
 Stability checks
 11/7/17

INTOXILYZER 8000
 Instrument Initialization
 08:57 11/07/2017

FHP TROOP D ORLANDO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001123
 11/07/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:38
Control Test	0.197	09:39
Air Blank	0.000	09:39
Control Test	0.198	09:40
Air Blank	0.000	09:41
Control Test	0.199	09:41
Air Blank	0.000	09:42
Control Test Stats		
Average	0.1980	
Std Dev	0.0010	
Rel Std Dev(%)	0.5051	

Operator's Signature

FHP TROOP D ORLANDO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001123
 11/07/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:33
Control Test	0.079	09:33
Air Blank	0.000	09:34
Control Test	0.079	09:34
Air Blank	0.000	09:35
Control Test	0.080	09:35
Air Blank	0.000	09:35
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DGS

Operator's Signature

11/7/17

FHP TROOP D ORLANDO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001123
 11/07/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:43
Control Test	0.049	09:44
Air Blank	0.000	09:44
Control Test	0.050	09:45
Air Blank	0.000	09:45
Control Test	0.050	09:46
Air Blank	0.000	09:47
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

Operator's Signature

FHP TROOP D ORLANDO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001123
 11/07/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:55
Control Test	0.080	09:55
Air Blank	0.000	09:56
Control Test	0.080	09:57
Air Blank	0.000	09:57
Control Test	0.080	09:58
Air Blank	0.000	09:58
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature