

INSTRUMENT PROCESSING SHEET

Agency Valusia County

S/N 80-001131

Date In 2/9/17

Date Out 2/17/17

Ship P/U H/D CMI EE

Intake Performed By DP

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By PPM

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 140
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm .148 (.139 - .169)
 36mm .160 (.156 - .190)
 53mm .228 (.228 - .278)
 103mm .503 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 26932

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>DR2035</u>	<u>201603D</u> <u>03/8/18</u>
0.08	<u>SD1011</u>	<u>201601F</u> <u>1/26/18</u>
0.20	<u>SD1025</u>	<u>201604C</u> <u>4/5/18</u>
0.08 DGS	<u>N/A</u>	<u>03415030A1</u> <u>03415080A1</u>

PPM AG626604 2/15/17 PPM
9/22/18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 FEB 21 2017
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	<u>N/A</u>		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	<u>N/A</u>		

Department Inspection Performed By PPM

Barometric Pressure ID# 26932 Gauge 1018 Instrument 1018

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	<u>SD1019</u>
Interferent	<u>SD1021</u>
0.05	<u>DR2035</u>
0.08	<u>SD1011</u>
0.20	<u>SD1025</u>

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: OK 2/20/17 PP

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland

Quality Control Review

2/21/17

Date

VOLUSIA COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8800 SN 80-001131
02/17/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:52
Control Test	0.049	09:52
Air Blank	0.000	09:53
Control Test	0.049	09:54
Air Blank	0.000	09:54
Control Test	0.049	09:55
Air Blank	0.000	09:55
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P Murphy

Operator's Signature

VOLUSIA COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8800 SN 80-001131
02/17/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:46
Control Test	0.081	09:47
Air Blank	0.000	09:48
Control Test	0.079	09:48
Air Blank	0.000	09:49
Control Test	0.080	09:50
Air Blank	0.000	09:50
Control Test Stats		
Average	0.0800	
Std Dev	0.0010	
Rel Std Dev(%)	1.2500	

P Murphy

Operator's Signature

VOLUSIA COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8800 SN 80-001131
02/17/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:57
Control Test	0.198	09:57
Air Blank	0.000	09:58
Control Test	0.199	09:59
Air Blank	0.000	09:59
Control Test	0.198	10:00
Air Blank	0.000	10:00
Control Test Stats		
Average	0.1983	
Std Dev	0.0006	
Rel Std Dev(%)	0.2911	

P Murphy

Operator's Signature

VOLUSIA COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8800 SN 80-001131
02/17/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:12
Control Test	0.080	10:12
Air Blank	0.000	10:13
Control Test	0.080	10:13
Air Blank	0.000	10:14
Control Test	0.080	10:14
Air Blank	0.000	10:14
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

J

BK

D65

P Murphy

Operator's Signature