



INSTRUMENT PROCESSING SHEET

Agency Lake Helen PDS/N 80-001149Florida Department of
Law EnforcementDate In 11/3/17DI Completion Date 11/9/17☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By <u>[Signature]</u> ☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>[Signature]</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>183</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP105</u> 32 mm <u>.160</u> (.139 - .169) 36 mm <u>.179</u> (.156 - .190) 53 mm <u>.253</u> (.228 - .278) 103 mm <u>.520</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28421</u> ☒ Stability Checks	Flow Calibration Performed By _____ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																											
Final Release Date NOV 14 2017	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>SD3962</u></td> <td><u>201767D</u> <u>7/25/19</u></td> </tr> <tr> <td>0.080</td> <td><u>SD1013</u></td> <td><u>201707E</u> <u>7/25/19</u></td> </tr> <tr> <td>0.200</td> <td><u>DR3856</u></td> <td><u>201767C</u> <u>7/24/19</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG626605</u> <u>9/22/18</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	<u>SD3962</u>	<u>201767D</u> <u>7/25/19</u>	0.080	<u>SD1013</u>	<u>201707E</u> <u>7/25/19</u>	0.200	<u>DR3856</u>	<u>201767C</u> <u>7/24/19</u>	0.080 DGS	N/A	<u>AG626605</u> <u>9/22/18</u>	Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Temperature Checks Performed By <u>[Signature]</u> ☒ Lab Temp °C <u>21.7</u> External Digital Therm. ID#: <u>300504</u> ☒ 34°C +/- .2 Serial #: <u>G2835</u> ☒ 34°C +/- .2 Serial #: <u>DR1279</u> ☒ 34°C +/- .2 Serial #: <u>SD1025</u>																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	<u>SD3962</u>	<u>201767D</u> <u>7/25/19</u>																																																											
0.080	<u>SD1013</u>	<u>201707E</u> <u>7/25/19</u>																																																											
0.200	<u>DR3856</u>	<u>201767C</u> <u>7/24/19</u>																																																											
0.080 DGS	N/A	<u>AG626605</u> <u>9/22/18</u>																																																											
Calibration Adjustment Performed By <u>[Signature]</u> Barometric Pressure Gauge <u>1015</u> ID # <u>28427</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>SD1016</u></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td><u>SD1018</u></td> <td><u>16320</u></td> <td><u>10/12/18</u></td> </tr> <tr> <td>0.100</td> <td><u>G2834</u></td> <td><u>17060</u></td> <td><u>2/14/19</u></td> </tr> <tr> <td>0.200</td> <td><u>G2408</u></td> <td><u>17090</u></td> <td><u>2/24/19</u></td> </tr> <tr> <td>0.300</td> <td><u>DR1275</u></td> <td><u>16410</u></td> <td><u>12/19/18</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>17817680A2</u></td> <td><u>8/5/19</u></td> </tr> </tbody> </table> ☒ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>G2835</u></td> <td><u>201707D</u></td> <td><u>7/25/19</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1279</u></td> <td><u>201767E</u></td> <td><u>7/25/19</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1025</u></td> <td><u>201707C</u></td> <td><u>7/24/19</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG626605</u></td> <td><u>9/22/18</u></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	<u>SD1016</u>	N/A	N/A	0.040	<u>SD1018</u>	<u>16320</u>	<u>10/12/18</u>	0.100	<u>G2834</u>	<u>17060</u>	<u>2/14/19</u>	0.200	<u>G2408</u>	<u>17090</u>	<u>2/24/19</u>	0.300	<u>DR1275</u>	<u>16410</u>	<u>12/19/18</u>	0.080 DGS	N/A	<u>17817680A2</u>	<u>8/5/19</u>	Simulator	Serial Number	Lot Number	Expiration	0.050	<u>G2835</u>	<u>201707D</u>	<u>7/25/19</u>	0.080	<u>DR1279</u>	<u>201767E</u>	<u>7/25/19</u>	0.200	<u>SD1025</u>	<u>201707C</u>	<u>7/24/19</u>	0.080 DGS	N/A	<u>AG626605</u>	<u>9/22/18</u>	Department Inspection Performed By <u>[Signature]</u> Barometric Pressure ID# <u>28421</u> Gauge <u>1014</u> Instrument <u>1015</u> Mouth Alcohol Solution Lot # <u>2016-C</u> Acetone Stock Solution Lot # <u>2017-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>G2880</u></td> </tr> <tr> <td>Interferent</td> <td><u>G2840</u></td> </tr> <tr> <td>0.050</td> <td><u>G2835</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1279</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1025</u></td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	<u>G2880</u>	Interferent	<u>G2840</u>	0.050	<u>G2835</u>	0.080	<u>DR1279</u>	0.200	<u>SD1025</u>
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000	<u>SD1016</u>	N/A	N/A																																																										
0.040	<u>SD1018</u>	<u>16320</u>	<u>10/12/18</u>																																																										
0.100	<u>G2834</u>	<u>17060</u>	<u>2/14/19</u>																																																										
0.200	<u>G2408</u>	<u>17090</u>	<u>2/24/19</u>																																																										
0.300	<u>DR1275</u>	<u>16410</u>	<u>12/19/18</u>																																																										
0.080 DGS	N/A	<u>17817680A2</u>	<u>8/5/19</u>																																																										
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050	<u>G2835</u>	<u>201707D</u>	<u>7/25/19</u>																																																										
0.080	<u>DR1279</u>	<u>201767E</u>	<u>7/25/19</u>																																																										
0.200	<u>SD1025</u>	<u>201707C</u>	<u>7/24/19</u>																																																										
0.080 DGS	N/A	<u>AG626605</u>	<u>9/22/18</u>																																																										
Simulator	Serial Number																																																												
0.000	<u>G2880</u>																																																												
Interferent	<u>G2840</u>																																																												
0.050	<u>G2835</u>																																																												
0.080	<u>DR1279</u>																																																												
0.200	<u>SD1025</u>																																																												
Notes/Suggested Service: _____ _____ _____ _____ _____ _____	Attachments ☒ Form 41 ☐ Flow Calibration ☒ Stability Checks ☐ Form 40 ☒ Calibration Adjustment ☐ Other _____ ☒ Post-Stability Checks _____																																																												
☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use <u>SP 11/14/17</u> <u>[Signature]</u> <u>11/14/17</u> Tech Review / Date Admin Review / Date																																																													

LAKE HELEN P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001149
11/08/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:56
Control Test	0.053	14:57
Air Blank	0.000	14:58
Control Test	0.052	14:58
Air Blank	0.000	14:59
Control Test	0.052	15:00
Air Blank	0.000	15:00
Control Test Stats		
Average	0.0523	
Std Dev	0.0006	
Rel Std Dev(%)	1.1032	

P Murphy
Operator's Signature

LAKE HELEN P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001149
11/08/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:51
Control Test	0.084	14:52
Air Blank	0.000	14:52
Control Test	0.085	14:53
Air Blank	0.000	14:54
Control Test	0.083	14:54
Air Blank	0.000	14:55
Control Test Stats		
Average	0.0840	
Std Dev	0.0010	
Rel Std Dev(%)	1.1905	

P Murphy
Operator's Signature

LAKE HELEN P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001149
11/08/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:46
Control Test	0.203	14:47
Air Blank	0.000	14:47
Control Test	0.205	14:48
Air Blank	0.000	14:48
Control Test	0.206	14:49
Air Blank	0.000	14:50
Control Test Stats		
Average	0.2047	
Std Dev	0.0015	
Rel Std Dev(%)	0.7463	

P Murphy
Operator's Signature

LAKE HELEN P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001149
11/08/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:16
Control Test	0.080	15:16
Air Blank	0.000	15:16
Control Test	0.080	15:17
Air Blank	0.000	15:17
Control Test	0.081	15:18
Air Blank	0.000	15:18
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

DGS

P Murphy
Operator's Signature

SP

11/14/17
20

LAKE HELEN P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001149
11/09/2017 10:54:34

Auto Calibration
Max Power Res Value = 33
Auto Range Res Value = 20

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 13008, Sum Io = 14034

Channel 1

Sample	% Abs	(% Abs Ref)
Sample #1	0.1380	(-0.0180)
Sample #2	0.1380	(0.0400)
Sample #3	0.1200	(0.0330)
Sample #4	0.0940	(0.0790)
Avg % Abs	0.1173	(0.0507)
STD DEV	0.0221	(0.0248)
REL STD DEV	18.853	(48.919)

Channel 2

Sample	% Abs	(% Abs Ref)
Sample #1	0.1260	(-0.0100)
Sample #2	0.1090	(0.0120)
Sample #3	0.1180	(-0.0070)
Sample #4	0.1110	(0.0100)
Avg % Abs	0.1127	(0.0050)
STD DEV	0.0047	(0.0104)
REL STD DEV	4.195	(208.806)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12999, Sum Io = 14031

Channel 1

Sample	% Abs	(% Abs Ref)
Sample #1	0.9040	(-0.0120)
Sample #2	0.9010	(0.0060)
Sample #3	0.9080	(0.0300)
Sample #4	0.9300	(0.0080)
Avg % Abs	0.9130	(0.0147)
STD DEV	0.0151	(0.0133)
REL STD DEV	1.657	(90.795)

Channel 2

Sample	% Abs	(% Abs Ref)
Sample #1	1.5780	(-0.0060)
Sample #2	1.5730	(0.0060)
Sample #3	1.5990	(0.0080)
Sample #4	1.5940	(0.0110)
Avg % Abs	1.5887	(0.0083)
STD DEV	0.0138	(0.0025)
REL STD DEV	0.868	(30.199)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12997, Sum Io = 14030

Channel 1

Sample	% Abs	(% Abs Ref)
Sample #1	2.0750	(-0.0110)
Sample #2	2.0710	(0.0440)
Sample #3	2.0590	(0.0300)
Sample #4	2.0730	(0.0420)
Avg % Abs	2.0677	(0.0387)
STD DEV	0.0076	(0.0076)
REL STD DEV	0.366	(19.582)

Channel 2

Sample	% Abs	(% Abs Ref)
Sample #1	3.7740	(-0.0040)
Sample #2	3.7820	(0.0330)
Sample #3	3.7780	(0.0250)
Sample #4	3.7640	(0.0420)
Avg % Abs	3.7747	(0.0333)
STD DEV	0.0095	(0.0085)
REL STD DEV	0.250	(25.515)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12995, Sum Io = 14028

Channel 1

Sample	% Abs	(% Abs Ref)
Sample #1	4.0150	(-0.0250)
Sample #2	3.9580	(0.0200)
Sample #3	3.9510	(0.0580)
Sample #4	3.9310	(0.0670)
Avg % Abs	3.9467	(0.0483)
STD DEV	0.0140	(0.0249)
REL STD DEV	0.355	(51.614)

Channel 2

Sample	% Abs	(% Abs Ref)
Sample #1	7.2430	(0.0130)
Sample #2	7.1850	(0.0610)
Sample #3	7.1700	(0.0790)
Sample #4	7.1910	(0.0800)
Avg % Abs	7.1820	(0.0733)
STD DEV	0.0108	(0.0107)
REL STD DEV	0.151	(14.581)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12994, Sum Io = 14022

Channel 1

Sample	% Abs	(% Abs Ref)
Sample #1	5.7500	(-0.0090)
Sample #2	5.7820	(0.0590)
Sample #3	5.7720	(0.0780)
Sample #4	5.7590	(0.0890)
Avg % Abs	5.7710	(0.0753)
STD DEV	0.0115	(0.0152)
REL STD DEV	0.200	(20.146)

Channel 2

Sample	% Abs	(% Abs Ref)
Sample #1	10.4040	(-0.0210)
Sample #2	10.4120	(0.0610)
Sample #3	10.4290	(0.0750)
Sample #4	10.4290	(0.0830)
Avg % Abs	10.4233	(0.0730)
STD DEV	0.0098	(0.0111)
REL STD DEV	0.094	(15.254)

AUTO CAL DATA

Channel 1

Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.117
Std Dev = 0.02 Rel Std Dev = 18.85
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.913
Std Dev = 0.02 Rel Std Dev = 1.66
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 2.068
Std Dev = 0.01 Rel Std Dev = 0.37
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.947
Std Dev = 0.01 Rel Std Dev = 0.36
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.771
Std Dev = 0.01 Rel Std Dev = 0.20
Zero Order Coef = -289.49
First Order Coef = 2396.12
Second Order Coef = 22.51
Standard Deviation = 7.989819

Channel 2

Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.113
Std Dev = 0.00 Rel Std Dev = 4.19
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.589
Std Dev = 0.01 Rel Std Dev = 0.87
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.775
Std Dev = 0.01 Rel Std Dev = 0.25
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 7.182
Std Dev = 0.01 Rel Std Dev = 0.15
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 10.423
Std Dev = 0.01 Rel Std Dev = 0.09
Zero Order Coef = -136.11
First Order Coef = 1255.33
Second Order Coef = 12.32
Standard Deviation = 12.257858

Solution Stats Quadratic Fit Chan 1		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.000	0.0002
0.040	0.040	-0.0003
0.100	0.100	0.0000
0.200	0.200	0.0001
0.300	0.300	-0.0001

SP

11/14/17

80-061149

Solution Stats Quadratic Fit Chan 2		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0001
0.040	0.040	0.0003
0.100	0.100	-0.0003
0.200	0.200	0.0002
0.300	0.300	-0.0000

Sol Value = 0.080 g/210L ***
 Fit value = 0.3810 mg/l %%%
 Samples Taken = 4, Discarded = 1

***** CHANNEL 1

Sample #1 = 3115.00
 Sample #2 = 3073.00
 Sample #3 = 3103.00
 Sample #4 = 3053.00
 Average Result = 3076.3333
 STD DEV = 25.1661
 REL STD DEV = 0.818

***** CHANNEL 2

Sample #1 = 3395.00
 Sample #2 = 3377.00
 Sample #3 = 3390.00
 Sample #4 = 3387.00
 Average Result = 3384.6667
 STD DEV = 6.8069
 REL STD DEV = 0.201

Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1014
 3 um H2O Adjust (mg/l*10,000) = 733
 9 um H2O Adjust (mg/l*10,000) = 425
 **** AUTO CAL PASS

LAKE HELEN P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001149
 11/09/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:37
Control Test	0.049	11:38
Air Blank	0.000	11:39
Control Test	0.049	11:39
Air Blank	0.000	11:40
Control Test	0.049	11:41
Air Blank	0.000	11:41
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P Murphy
 Operator's Signature

LAKE HELEN P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001149
 11/09/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:48
Control Test	0.201	11:48
Air Blank	0.000	11:49
Control Test	0.199	11:50
Air Blank	0.000	11:50
Control Test	0.199	11:51
Air Blank	0.000	11:51
Control Test Stats		
Average	0.1997	
Std Dev	0.0012	
Rel Std Dev(%)	0.5783	

P Murphy
 Operator's Signature

LAKE HELEN P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001149
 11/09/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:42
Control Test	0.080	11:43
Air Blank	0.000	11:44
Control Test	0.079	11:44
Air Blank	0.000	11:45
Control Test	0.080	11:46
Air Blank	0.000	11:46
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

P Murphy
 Operator's Signature

LAKE HELEN P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001149
 11/09/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:54
Control Test	0.079	11:54
Air Blank	0.000	11:55
Control Test	0.079	11:55
Air Blank	0.000	11:56
Control Test	0.080	11:56
Air Blank	0.000	11:57
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DGS

P Murphy
 Operator's Signature

POST
 CALIBRATION
 ADJUST.
 STABILITIES

11/14/17
SP