

## INSTRUMENT PROCESSING SHEET

Agency Haines Cnty PD S/N 80-001151  
 Date In 6/23/17 Date Out 6/27/17  Ship  P/U  H/D  CMI  EE

<b>Intake</b> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____  Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose  Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Other _____  Notes: _____ _____ _____	<b>Quality Checks</b> Performed By <u>KDB</u> <input checked="" type="checkbox"/> Lab Temp °C <u>20.0</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>228 224</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP103</u> 32mm <u>0.148</u> (.139 - .169) 36mm <u>0.164</u> (.156 - .190) 53mm <u>0.234</u> (.228 - .278) 103mm <u>0.500</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td><u>SD3962</u></td> <td><u>2016D3D</u> <u>3/8/18</u></td> </tr> <tr> <td>0.08</td> <td><u>SD3964</u></td> <td><u>201611B</u> <u>11/15/18</u></td> </tr> <tr> <td>0.20</td> <td><u>SD3933</u></td> <td><u>2016D4C</u> <u>4/15/18</u></td> </tr> <tr> <td>0.08 DGS</td> <td><u>N/A</u></td> <td><u>AG626604</u> <u>9/22/18</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	<u>SD3962</u>	<u>2016D3D</u> <u>3/8/18</u>	0.08	<u>SD3964</u>	<u>201611B</u> <u>11/15/18</u>	0.20	<u>SD3933</u>	<u>2016D4C</u> <u>4/15/18</u>	0.08 DGS	<u>N/A</u>	<u>AG626604</u> <u>9/22/18</u>	<b>Flow Calibration</b> Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)  <b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____  <b>Quality Checks Cont.</b> Performed By <u>KDB</u> <b>Simulator Temperatures °C</b> External Digital Therm. ID#: <u>300504</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>SD3962</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>SD3964</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>SD3933</u>
Simulator	Serial #	Lot #/Exp															
0.05	<u>SD3962</u>	<u>2016D3D</u> <u>3/8/18</u>															
0.08	<u>SD3964</u>	<u>201611B</u> <u>11/15/18</u>															
0.20	<u>SD3933</u>	<u>2016D4C</u> <u>4/15/18</u>															
0.08 DGS	<u>N/A</u>	<u>AG626604</u> <u>9/22/18</u>															


RECEIVED  
 JUN 29 2017  
 FDLE  
 Alcohol Testing Program

<b>Calibration Adjustment</b> Performed By _____ <input checked="" type="checkbox"/> Calibration Adjustment N/A <input type="checkbox"/> Calibration Adjustment Complete Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td><u>N/A</u></td> <td><u>N/A</u></td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td><u>N/A</u></td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td><u>N/A</u></td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		<u>N/A</u>	<u>N/A</u>	0.040				0.100				0.200				0.300				0.080 DGS	<u>N/A</u>			Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	<u>N/A</u>			<b>Department Inspection</b> Performed By <u>KDB</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1014</u> Gauge ID# <u>28427</u> <u>1014</u> Instrument  Mouth Alcohol Solution Lot # <u>2016-C</u> Acetone Stock Solution Lot # <u>2017-A</u> <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td><u>G2880</u></td> </tr> <tr> <td>Interferent</td> <td><u>G2834</u></td> </tr> <tr> <td>0.05</td> <td><u>SD3962</u></td> </tr> <tr> <td>0.08</td> <td><u>SD3964</u></td> </tr> <tr> <td>0.20</td> <td><u>SD3933</u></td> </tr> </tbody> </table>	Simulator	Serial Number	0.00	<u>G2880</u>	Interferent	<u>G2834</u>	0.05	<u>SD3962</u>	0.08	<u>SD3964</u>	0.20	<u>SD3933</u>
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		<u>N/A</u>	<u>N/A</u>																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	<u>N/A</u>																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.05																																																													
0.08																																																													
0.20																																																													
0.08 DGS	<u>N/A</u>																																																												
Simulator	Serial Number																																																												
0.00	<u>G2880</u>																																																												
Interferent	<u>G2834</u>																																																												
0.05	<u>SD3962</u>																																																												
0.08	<u>SD3964</u>																																																												
0.20	<u>SD3933</u>																																																												
<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____																																																													

Notes/Suggested Service:  
QA/QC OK SP

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC  
 Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use

Brett Kuehler 6/28/17  
 Quality Control Review Date

Stability Checks #80-001151 Haines City P.D. 6/27/17 

HAINES CITY PD  
Intoxilyzer - Alcohol Analyzer  
Model: 8000 SN 80-001151  
06/27/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:41
Control Test	0.050	09:41
Air Blank	0.000	09:42
Control Test	0.050	09:42
Air Blank	0.000	09:43
Control Test	0.049	09:44
Air Blank	0.000	09:44
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

  
Operator's Signature

HAINES CITY PD  
Intoxilyzer - Alcohol Analyzer  
Model: 8000 SN 80-001151  
06/27/2017  
Software: 8100.27


Test	g/210L	Time
Air Blank	0.000	09:46
Control Test	0.080	09:46
Air Blank	0.000	09:47
Control Test	0.079	09:47
Air Blank	0.000	09:48
Control Test	0.079	09:49
Air Blank	0.000	09:49
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

  
Operator's Signature

HAINES CITY PD  
Intoxilyzer - Alcohol Analyzer  
Model: 8000 SN 80-001151  
06/27/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:53
Control Test	0.200	09:54
Air Blank	0.000	09:54
Control Test	0.199	09:55
Air Blank	0.000	09:55
Control Test	0.199	09:56
Air Blank	0.000	09:57
Control Test Stats		
Average	0.1993	
Std Dev	0.0006	
Rel Std Dev(%)	0.2896	

  
Operator's Signature

  
HAINES CITY PD  
Intoxilyzer - Alcohol Analyzer  
Model: 8000 SN 80-001151  
06/27/2017  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:58
Control Test	0.082	09:58
Air Blank	0.000	09:59
Control Test	0.082	09:59
Air Blank	0.000	09:59
Control Test	0.082	09:59
Air Blank	0.000	10:00
Control Test	0.082	10:00
Control Test Stats		
Average	0.0820	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP  
ASK

  
Operator's Signature