

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-001178	Miami Police Department	11/13/2017	<i>MLL</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
SN: SD3967 Temp: 34.06c	SN: SD3968 Temp: 34.00c	SN: SD3969 Temp: 34.06c	Lot AG626604
0.047 to 0.053 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>	0.194 to 0.206 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>

<p>MIAMI PD Intoxilyzer - Alcohol Analyzer Model 8000 11/13/2017 SN 80-001178 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 06:12 Control Test 0.050 06:12 Air Blank 0.000 06:13 Control Test 0.049 06:14 Air Blank 0.000 06:14 Control Test 0.049 06:15 Air Blank 0.000 06:15</p> <p>Control Test Stats Average 0.0493 Std Dev 0.0006 Rel. Std Dev(%) 1.1703</p> <p>Operator's Signature <i>MLL</i></p>	<p>MIAMI PD Intoxilyzer - Alcohol Analyzer Model 8000 11/13/2017 SN 80-001178 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 06:16 Control Test 0.080 06:17 Air Blank 0.000 06:18 Control Test 0.080 06:18 Air Blank 0.000 06:19 Control Test 0.080 06:20 Air Blank 0.000 06:20</p> <p>Control Test Stats Average 0.0800 Std Dev 0.0000 Rel. Std Dev(%) 0.0000</p> <p>Operator's Signature <i>DELL</i></p>	<p>MIAMI PD Intoxilyzer - Alcohol Analyzer Model 8000 11/13/2017 SN 80-001178 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 06:21 Control Test 0.198 06:22 Air Blank 0.000 06:23 Control Test 0.198 06:23 Air Blank 0.000 06:24 Control Test 0.198 06:25 Air Blank 0.000 06:25</p> <p>Control Test Stats Average 0.1980 Std Dev 0.0000 Rel. Std Dev(%) 0.0000</p> <p>Operator's Signature <i>DELL</i></p>	<p>MIAMI PD Intoxilyzer - Alcohol Analyzer Model 8000 11/13/2017 SN 80-001178 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 06:26 Control Test 0.078 06:26 Air Blank 0.000 06:27 Control Test 0.078 06:27 Air Blank 0.000 06:28 Control Test 0.078 06:28 Air Blank 0.000 06:29</p> <p>Control Test Stats Average 0.0780 Std Dev 0.0000 Rel. Std Dev(%) 0.0000</p> <p>Operator's Signature <i>MLL</i></p>
---	--	--	---

11/21/17
DELL