



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Panama City Beach PD S/N 80-001192

Date In 10/20/17 Date Out 10/26/17

Ship P/U H/D CMI EE

**Intake** Performed By TD

Registration  
 Annual  
 Return from CMI  
 Return from Enforcement Electronics  
 Other \_\_\_\_\_

Visual Inspection:  
 Case  Handle  
 Dry Gas Holder  Feet  
 Keyboard/Plug  Back/Plugs  
 Screws tight  Breath Hose

Other Equipment:  
 Power cord  
 Printer Cable  
 Other Static Bag

Notes: \_\_\_\_\_

**Quality Checks** Performed By RCM

Lab Temp °C 22.2  
 Breath Tube Screen  
 Replace External O-Rings  
 Instrument Set Up Verified  
 R-Value 157  
 Flow Verification (L/s)  
 Flow Column # ATP105  
 32mm .156 (.139 - .169)  
 36mm .167 (.156 - .190)  
 53mm .246 (.228 - .278)  
 103mm .519 (.447 - .547)

Barometric Pressure Check  
 Gauge ID # 28421

Stability Checks

| Simulator | Serial #      | Lot #/Exp                         |
|-----------|---------------|-----------------------------------|
| 0.05      | <u>62835</u>  | <u>2017070</u><br><u>7/25/19</u>  |
| 0.08      | <u>DR1279</u> | <u>201707E</u><br><u>7/25/19</u>  |
| 0.20      | <u>SD1025</u> | <u>201707C</u><br><u>7/24/19</u>  |
| 0.08 DGS  | N/A           | <u>AG626605</u><br><u>9/22/18</u> |

**Flow Calibration** Performed By \_\_\_\_\_

Flow Calibration N/A  
 Flow Calibration Complete  
 Flow Column # \_\_\_\_\_  
 5L/min - 17mm  
 15L/min - 30mm  
 30L/min - 103mm

R-Value \_\_\_\_\_  
 Post Calibration Verification (L/s)  
 Flow Column # \_\_\_\_\_  
 32mm \_\_\_\_\_ (.139 - .169)  
 36mm \_\_\_\_\_ (.156 - .190)  
 53mm \_\_\_\_\_ (.228 - .278)  
 103mm \_\_\_\_\_ (.447 - .547)

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Quality Checks Cont.** Performed By RCM

**Simulator Temperatures °C**  
 External Digital Therm. ID#: 300504  
 34°C +/- .2 Serial #: 62835  
 34°C +/- .2 Serial #: DR1279  
 34°C +/- .2 Serial #: SD1025

RECEIVED  
OCT 26 2017  
FDLE  
Alcohol Testing Program

**Calibration Adjustment** Performed By \_\_\_\_\_

Calibration Adjustment N/A  
 Calibration Adjustment Complete

Barometric Pressure Gauge \_\_\_\_\_ ID # \_\_\_\_\_

| Simulator | Serial Number | Lot Number | Expiration |
|-----------|---------------|------------|------------|
| 0.000     |               | N/A        | N/A        |
| 0.040     |               |            |            |
| 0.100     |               |            |            |
| 0.200     |               |            |            |
| 0.300     |               |            |            |
| 0.080 DGS | N/A           |            |            |

Post Calibration Adjustment Stability Checks

| Simulator | Serial Number | Lot Number | Expiration |
|-----------|---------------|------------|------------|
| 0.05      |               |            |            |
| 0.08      |               |            |            |
| 0.20      |               |            |            |
| 0.08 DGS  | N/A           |            |            |

**Department Inspection** Performed By RCM

Barometric Pressure 1011 Gauge  
 ID# 28421 1008 Instrument

Mouth Alcohol Solution Lot # 2016-C  
 Acetone Stock Solution Lot # 2017-A

| Simulator   | Serial Number |
|-------------|---------------|
| 0.00        | <u>62880</u>  |
| Interferent | <u>62840</u>  |
| 0.05        | <u>62835</u>  |
| 0.08        | <u>DR1279</u> |
| 0.20        | <u>SD1025</u> |

**Attachments**

|   |   |
|---|---|
| <input checked="" type="checkbox"/> Form 41             | <input type="checkbox"/> Calibration Adjustment |
| <input checked="" type="checkbox"/> Pre-Stability Tests | <input type="checkbox"/> Post-Stability Tests   |
| <input type="checkbox"/> Flow Calibration               | <input type="checkbox"/> Other _____            |

Notes/Suggested Service:  
QA/DC 8/25/17

\_\_\_\_\_  
10/26/17

Instrument Complies with Chapter 11D-8, FAC

Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use

Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Quality Control Review

Date

PANAMA CITY BEACH PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001192  
10/26/2017  
Software: 8100.27

| Test               | g/210L | Time  |
|--------------------|--------|-------|
| Air Blank          | 0.000  | 06:59 |
| Control Test       | 0.050  | 07:00 |
| Air Blank          | 0.000  | 07:00 |
| Control Test       | 0.050  | 07:01 |
| Air Blank          | 0.000  | 07:02 |
| Control Test       | 0.050  | 07:02 |
| Air Blank          | 0.000  | 07:03 |
| Control Test Stats |        |       |
| Average            | 0.0500 |       |
| Std Dev            | 0.0000 |       |
| Rel Std Dev(%)     | 0.0000 |       |

*P Murphy*  
-----  
Operator's Signature

PANAMA CITY BEACH PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001192  
10/26/2017  
Software: 8100.27

| Test               | g/210L | Time  |
|--------------------|--------|-------|
| Air Blank          | 0.000  | 06:54 |
| Control Test       | 0.080  | 06:55 |
| Air Blank          | 0.000  | 06:55 |
| Control Test       | 0.080  | 06:56 |
| Air Blank          | 0.000  | 06:56 |
| Control Test       | 0.081  | 06:57 |
| Air Blank          | 0.000  | 06:58 |
| Control Test Stats |        |       |
| Average            | 0.0803 |       |
| Std Dev            | 0.0006 |       |
| Rel Std Dev(%)     | 0.7187 |       |

*P Murphy*  
-----  
Operator's Signature

PANAMA CITY BEACH PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001192  
10/26/2017  
Software: 8100.27

| Test               | g/210L | Time  |
|--------------------|--------|-------|
| Air Blank          | 0.000  | 07:05 |
| Control Test       | 0.201  | 07:06 |
| Air Blank          | 0.000  | 07:06 |
| Control Test       | 0.203  | 07:07 |
| Air Blank          | 0.000  | 07:07 |
| Control Test       | 0.202  | 07:08 |
| Air Blank          | 0.000  | 07:09 |
| Control Test Stats |        |       |
| Average            | 0.2020 |       |
| Std Dev            | 0.0010 |       |
| Rel Std Dev(%)     | 0.4950 |       |

*P Murphy*  
-----  
Operator's Signature

PANAMA CITY BEACH PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001192  
10/26/2017  
Software: 8100.27

| Test               | g/210L | Time  |
|--------------------|--------|-------|
| Air Blank          | 0.000  | 07:11 |
| Control Test       | 0.080  | 07:12 |
| Air Blank          | 0.000  | 07:12 |
| Control Test       | 0.080  | 07:12 |
| Air Blank          | 0.000  | 07:13 |
| Control Test       | 0.080  | 07:13 |
| Air Blank          | 0.000  | 07:14 |
| Control Test Stats |        |       |
| Average            | 0.0800 |       |
| Std Dev            | 0.0000 |       |
| Rel Std Dev(%)     | 0.0000 |       |

*MS*

DGS

*P Murphy*  
-----  
Operator's Signature

10/26/17  
*JD*