



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency NAS - Pensacola S/N 80-001205
 Date In 10/18/17 Date Out 10/19/17 Ship P/U H/D CMI EE

Intake Performed By TP

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
 Case Handle
 Dry Gas Holder Feet
 Keyboard/Plug Back/Plugs
 Screws tight Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By PGM

Lab Temp °C 22.1
 Breath Tube Screen
 Replace External O-Rings
 Instrument Set Up Verified
 R-Value 221
 Flow Verification (L/s)
 Flow Column # ATP103
 32mm .148 (.139 - .169)
 36mm .164 (.156 - .190)
 53mm .238 (.228 - .278)
 103mm .492 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28421

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>G2835</u>	<u>201707D</u> <u>7/25/19</u>
0.08	<u>DR1279</u>	<u>201707E</u> <u>7/25/19</u>
0.20	<u>SD1025</u>	<u>201707C</u> <u>7/25/19</u>
0.08 DGS	N/A	<u>A6626605</u> <u>9/22/18</u>

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Quality Checks Cont. Performed By TP

Simulator Temperatures °C
 External Digital Therm. ID#: 300504
 34°C +/- .2 Serial #: G2835
 34°C +/- .2 Serial #: DR1279
 34°C +/- .2 Serial #: SD1025

RECEIVED
 OCT 23 2017
 FDLE
 Alcohol Testing Program

Calibration Adjustment Performed By _____

Calibration Adjustment N/A
 Calibration Adjustment Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Department Inspection Performed By PGM

Barometric Pressure 1020 Gauge
 ID# 28421 1021 Instrument

Mouth Alcohol Solution Lot # 2012-C
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	<u>G2880</u>
Interferent	<u>G2840</u>
0.05	<u>G2835</u>
0.08	<u>DR1279</u>
0.20	<u>SD1025</u>

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Calibration Adjustment
 Post-Stability Tests
 Other _____

Notes/Suggested Service:
QA/QC ✓ SP

JJ Graham
 Quality Control Review

10/23/17
 Date

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

NAS PENSACOLA PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001205
10/19/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:35
Control Test	0.051	10:35
Air Blank	0.000	10:36
Control Test	0.051	10:37
Air Blank	0.000	10:37
Control Test	0.050	10:38
Air Blank	0.000	10:38
Control Test Stats		
Average	0.0507	
Std Dev	0.0006	
Rel Std Dev(%)	1.1395	


Operator's Signature

NAS PENSACOLA PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001205
10/19/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:51
Control Test	0.081	10:51
Air Blank	0.000	10:52
Control Test	0.080	10:52
Air Blank	0.000	10:53
Control Test	0.080	10:54
Air Blank	0.000	10:54
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	


Operator's Signature

NAS PENSACOLA PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001205
10/19/2017
Software: 8100.27


Test	g/210L	Time
Air Blank	0.000	10:45
Control Test	0.200	10:46
Air Blank	0.000	10:47
Control Test	0.201	10:47
Air Blank	0.000	10:48
Control Test	0.200	10:49
Air Blank	0.000	10:49
Control Test Stats		
Average	0.2003	
Std Dev	0.0006	
Rel Std Dev(%)	0.2882	


Operator's Signature

NAS PENSACOLA PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001205
10/19/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:56
Control Test	0.080	10:56
Air Blank	0.000	10:57
Control Test	0.080	10:57
Air Blank	0.000	10:58
Control Test	0.080	10:58
Air Blank	0.000	10:58
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS


Operator's Signature

SP

10/23/17
JP