



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency HENDRY COUNTY SO S/N 80-001207
Date In 5/30/17 Date Out 6/1/17

Intake Performed By [Signature]
Registration
Annual
Return from CMI
Return from Enforcement Electronics
Other
Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose
Other Equipment:
Power cord
Printer Cable
Other STATIC BAG
Notes:

Quality Checks Performed By SP
Lab Temp 21.0
Breath Tube Screen
Replace External O-Rings
Instrument Set Up Verified
R-Value 210
Flow Verification (L/s)
Flow Column # ATP103
32mm .152 (.139 - .169)
36mm .167 (.156 - .190)
53mm .234 (.228 - .278)
103mm .492 (.447 - .547)
Barometric Pressure Check
Gauge ID # 28427
Stability Checks
Simulator Serial # Lot #/Exp
0.05 DR2035 201603D 3-8-18
0.08 SD3964 201611B 11-15-18
0.20 G11621 201604C 4-5-18
0.08 DGS N/A AG1026604 9-22-18

Flow Calibration Performed By
Flow Calibration N/A
Flow Calibration Complete
Flow Column #
5L/min - 17mm
15L/min - 53mm
30L/min - 103mm
R-Value
Post Calibration Verification (L/s)
Flow Column #
32mm (.139 - .169)
36mm (.156 - .190)
53mm (.228 - .278)
103mm (.447 - .547)
Maintenance Performed By
Battery Replacement
Dry Gas Regulator Replacement
Breath Tube Replacement
Other
Quality Checks Cont. Performed By [Signature]
Simulator Temperatures C
External Digital Therm. ID#: 300505
34C +/- .2 Serial #: DR2035
34C +/- .2 Serial #: SD3964
34C +/- .2 Serial #: G11621

Calibration Adjustment Performed By
Calibration Adjustment N/A
Calibration Adjustment Complete
Barometric Pressure Gauge ID #
Simulator Serial Number Lot Number Expiration
0.000 N/A N/A
0.040
0.100
0.200
0.300
0.080 DGS N/A
Post Calibration Adjustment Stability Checks
Simulator Serial Number Lot Number Expiration
0.05
0.08
0.20
0.08 DGS N/A

Department Inspection Performed By SP
Barometric Pressure 1016 Gauge
ID# 28427 1017 Instrument
Mouth Alcohol Solution Lot # 2016-A
Acetone Stock Solution Lot # 2017-A
Simulator Serial Number
0.00 G12879
Interferent G8144
0.05 DR2035
0.08 SD3964
0.20 G11621
Attachments
Form 41
Pre-Stability Tests
Flow Calibration
Calibration Adjustment
Post-Stability Tests
Other

Notes/Suggested Service:
QAQC - DB
[Signature] 6/1/17

Instrument Complies with Chapter 11D-8, FAC
Instrument Does Not Comply with Chapter 11D-8, FAC
Return to/Place into Evidentiary Use
Remain Out of Evidentiary Use
Conduct an Agency Inspection Before Evidentiary Use

Quality Control Review Date

STABILITY CHECKS - #80-001207- HENDRY COUNTY SO - 6/1/17

HENDRY COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001207
 06/01/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:47
Control Test	0.048	09:48
Air Blank	0.000	09:48
Control Test	0.048	09:49
Air Blank	0.000	09:50
Control Test	0.048	09:50
Air Blank	0.000	09:51
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

JD

HENDRY COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001207
 06/01/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:42
Control Test	0.079	09:43
Air Blank	0.000	09:43
Control Test	0.079	09:44
Air Blank	0.000	09:44
Control Test	0.079	09:45
Air Blank	0.000	09:46
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

HENDRY COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001207
 06/01/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:56
Control Test	0.194	09:57
Air Blank	0.000	09:58
Control Test	0.196	09:58
Air Blank	0.000	09:59
Control Test	0.194	10:00
Air Blank	0.000	10:00
Control Test Stats		
Average	0.1947	
Std Dev	0.0012	
Rel Std Dev(%)	0.5932	

SP

Operator's Signature

HENDRY COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001207
 06/01/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:52
Control Test	0.079	09:52
Air Blank	0.000	09:53
Control Test	0.080	09:53
Air Blank	0.000	09:53
Control Test	0.080	09:54
Air Blank	0.000	09:54
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

DGAS

SP

Operator's Signature

SP