



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Port Orange PD

S/N 80-00243

Date In 10/11/17

Date Out 10/19/17

Ship P/U H/D CMI EE

Intake <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Dry Gas Holder <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Keyboard/Plug <input checked="" type="checkbox"/> Back/Plugs <input checked="" type="checkbox"/> Screws tight <input checked="" type="checkbox"/> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>Static Bag</u> Notes: _____ _____ _____	Quality Checks Performed By <u>ASB</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.1</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>180</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP103</u> 32mm <u>.160</u> (.139 - .169) 36mm <u>.171</u> (.156 - .190) 53mm <u>.246</u> (.228 - .278) 103mm <u>.492</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28421</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)															
	<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td><u>G2835</u></td> <td><u>2017070</u> <u>7/25/19</u></td> </tr> <tr> <td>0.08</td> <td><u>DR1279</u></td> <td><u>201707E</u> <u>7/25/19</u></td> </tr> <tr> <td>0.20</td> <td><u>501025</u></td> <td><u>201707C</u> <u>7/24/19</u></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td><u>REF26605</u> <u>7/22/18</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	<u>G2835</u>	<u>2017070</u> <u>7/25/19</u>	0.08	<u>DR1279</u>	<u>201707E</u> <u>7/25/19</u>	0.20	<u>501025</u>	<u>201707C</u> <u>7/24/19</u>	0.08 DGS	N/A	<u>REF26605</u> <u>7/22/18</u>	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Quality Checks Cont. Performed By <u>ASB</u> Simulator Temperatures °C External Digital Therm. ID#: <u>300504</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>G2835</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>DR1279</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>501025</u>
Simulator	Serial #	Lot #/Exp															
0.05	<u>G2835</u>	<u>2017070</u> <u>7/25/19</u>															
0.08	<u>DR1279</u>	<u>201707E</u> <u>7/25/19</u>															
0.20	<u>501025</u>	<u>201707C</u> <u>7/24/19</u>															
0.08 DGS	N/A	<u>REF26605</u> <u>7/22/18</u>															

RECEIVED
OCT 23 2017
FDLE
Alcohol Testing Program

Calibration Adjustment Performed By _____

- Calibration Adjustment N/A
- Calibration Adjustment Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By ASB

- Barometric Pressure 1020 Gauge ID# 28421 1020 Instrument

Mouth Alcohol Solution Lot # 2016-C
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	<u>G2880</u>
Interferent	<u>G2840</u>
0.05	<u>G2835</u>
0.08	<u>DR1279</u>
0.20	<u>501025</u>

Attachments

- Form 41
- Pre-Stability Tests
- Flow Calibration
- Calibration Adjustment
- Post-Stability Tests
- Other form 40

Notes/Suggested Service:
QA/QC ✓ SP

J. Bohan
 Quality Control Review Date 10/23/17

- Instrument Complies with Chapter 11D-8, FAC
- Instrument Does Not Comply with Chapter 11D-8, FAC
- Return to/Place into Evidentiary Use
- Remain Out of Evidentiary Use
- Conduct an Agency Inspection Before Evidentiary Use

PORT ORANGE P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001245
10/19/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:57
Control Test	0.050	11:57
Air Blank	0.000	11:58
Control Test	0.050	11:59
Air Blank	0.000	11:59
Control Test	0.050	12:00
Air Blank	0.000	12:00
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P Murphy

Operator's Signature

PORT ORANGE P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001245
10/19/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:52
Control Test	0.081	11:52
Air Blank	0.000	11:53
Control Test	0.081	11:54
Air Blank	0.000	11:54
Control Test	0.080	11:55
Air Blank	0.000	11:55
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

P Murphy

Operator's Signature

PORT ORANGE P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001245
10/19/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:47
Control Test	0.203	11:47
Air Blank	0.000	11:48
Control Test	0.203	11:49
Air Blank	0.000	11:49
Control Test	0.202	11:50
Air Blank	0.000	11:50
Control Test Stats		
Average	0.2027	
Std Dev	0.0006	
Rel Std Dev(%)	0.2849	

P Murphy

Operator's Signature

PORT ORANGE P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001245
10/19/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:15
Control Test	0.080	12:15
Air Blank	0.000	12:15
Control Test	0.079	12:16
Air Blank	0.000	12:16
Control Test	0.080	12:17
Air Blank	0.000	12:17
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

DB5

P Murphy

Operator's Signature

SP

10/23/17
JL