

INSTRUMENT PROCESSING SHEET

Agency MONROE County Sheriff's Office S/N 80-001254

Date In 01/25/2017 Date Out 01/25/2017 Ship P/U H/D CMI EE

RECEIVED
FEB 10 2017
FDLE
Alcohol Testing Program

Intake Performed By DEAR

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other BATTERY REPLACEMENT

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By DEAR

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 205
 Flow Verification (L/s)
 Flow Column # ATD101
 32mm 164 (.139 - .169)
 36mm 179 (.156 - .190)
 53mm 250 (.228 - .278)
 103mm 530 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28663

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>503967</u>	<u>201603D</u> <u>03/08/2018</u>
0.08	<u>503968</u>	<u>201601F</u> <u>01/26/2018</u>
0.20	<u>503969</u>	<u>201604C</u> <u>04/05/2018</u>
0.08 DGS	N/A	<u>A6605301</u> <u>02/22/2018</u>

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By DEAR

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By DEAR

Barometric Pressure 1016 Gauge
 ID# 68639 1015 Instrument

Mouth Alcohol Solution Lot # 2016 A
 Acetone Stock Solution Lot # 2016 B

Simulator	Serial Number
0.00	<u>503965</u>
Interferent	<u>503966</u>
0.05	<u>503967</u>
0.08	<u>503968</u>
0.20	<u>503969</u>

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other FORM 40

Notes: 01/25/2017 APPROVED

OK/AC OK QSM 2/8/2017

Brett Haveland

Quality Control Review

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

2/10/17
Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-001254	Monroe County Sheriff's Office	01/25/2017	<i>AW</i>

0.05g/210L <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>	0.20g/210L <input checked="" type="checkbox"/>	DGS 0.08g/210L <input checked="" type="checkbox"/>
0.047 to 0.053 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>	0.194 to 0.206 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>

MONROE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001254
01/25/2017
Software: 8100.27

MONROE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001254
01/25/2017
Software: 8100.27

MONROE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001254
01/25/2017
Software: 8100.27

MONROE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001254
01/25/2017
Software: 8100.27

Test g/210L Time
Air Blank 0.000 09:40
Control Test 0.050 09:40
Air Blank 0.000 09:41
Control Test 0.049 09:42
Air Blank 0.000 09:42
Control Test 0.050 09:43
Air Blank 0.000 09:43
Control Test Stats
Average 0.0497
Std Dev 0.0006
Rel Std Dev(%) 1.1625

Test g/210L Time
Air Blank 0.000 09:45
Control Test 0.080 09:46
Air Blank 0.000 09:47
Control Test 0.079 09:47
Air Blank 0.000 09:48
Control Test 0.079 09:49
Air Blank 0.000 09:49
Control Test Stats
Average 0.0793
Std Dev 0.0006
Rel Std Dev(%) 0.7277

Test g/210L Time
Air Blank 0.000 09:50
Control Test 0.198 09:51
Air Blank 0.000 09:52
Control Test 0.197 09:52
Air Blank 0.000 09:53
Control Test 0.197 09:54
Air Blank 0.000 09:54
Control Test Stats
Average 0.1973
Std Dev 0.0006
Rel Std Dev(%) 0.2926

Test g/210L Time
Air Blank 0.000 09:56
Control Test 0.080 09:56
Air Blank 0.000 09:57
Control Test 0.079 09:57
Air Blank 0.000 09:58
Control Test 0.079 09:58
Air Blank 0.000 09:58
Control Test Stats
Average 0.0793
Std Dev 0.0006
Rel Std Dev(%) 0.7277

AW
Operator's Signature

AW
Operator's Signature

AW
Operator's Signature

AW
Operator's Signature

AW
SK