

INSTRUMENT PROCESSING SHEET

Agency Palm Bay PD S/N 80-001265
 Date In 3/15/17 Date Out 3/22/17 Ship P/U H/D CMI EE

Intake Performed By TD

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
 Case Handle
 Dry Gas Holder Feet
 Keyboard/Plug Back/Plugs
 Screws tight Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By SP

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 179
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm .140 (.139 - .169)
 36mm .167 (.156 - .190)
 53mm .234 (.228 - .278)
 103mm .496 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>G3709</u>	<u>201603D</u> <u>3-8-18</u>
0.08	<u>DR1279</u>	<u>201601F</u> <u>1-26-18</u>
0.20	<u>DR3856</u>	<u>201604C</u> <u>4-5-18</u>
0.08 DGS	N/A	<u>AG626604</u> <u>9-22-18</u>

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By SP

Barometric Pressure
 ID# 28427 1018 Gauge
1018 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	<u>G2880</u>
Interferent	<u>G2834</u>
0.05	<u>G3709</u>
0.08	<u>DR1279</u>
0.20	<u>DR3856</u>

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: QA/OC OK GSPM 3/23/2017

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

RECEIVED
 MAR 23 2017
 FDLE
 Alcohol Testing Program

Brett Hickland

Quality Control Review

3/23/17
Date

PALM BAY P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001265
 03/22/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:10
Control Test	0.051	10:11
Air Blank	0.000	10:11
Control Test	0.049	10:12
Air Blank	0.000	10:12
Control Test	0.051	10:13
Air Blank	0.000	10:14
Control Test Stats		
Average	0.0503	
Std Dev	0.0012	
Rel Std Dev(%)	2.2941	

SP

Operator's Signature

PALM BAY P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001265
 03/22/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:19
Control Test	0.079	10:20
Air Blank	0.000	10:21
Control Test	0.080	10:21
Air Blank	0.000	10:22
Control Test	0.080	10:23
Air Blank	0.000	10:23
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

SP

Operator's Signature

PALM BAY P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001265
 03/22/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:15
Control Test	0.199	10:15
Air Blank	0.000	10:16
Control Test	0.199	10:16
Air Blank	0.000	10:17
Control Test	0.199	10:18
Air Blank	0.000	10:18
Control Test Stats		
Average	0.1990	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

PALM BAY P.D.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001265
 03/22/2017
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:24
Control Test	0.080	10:25
Air Blank	0.000	10:25
Control Test	0.079	10:25
Air Blank	0.000	10:26
Control Test	0.079	10:26
Air Blank	0.000	10:27
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DGS

SP

Operator's Signature

STABILITY CHECKS
 INSTRUMENT # 80-001265

BK

DGS