

INSTRUMENT PROCESSING SHEET

Agency Pinalas County

S/N 80-001273

Date In 2/7/17

Date Out 2/14/17

Ship P/U H/D CMI EE

Intake Performed By [Signature]

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
 Case Handle
 Dry Gas Holder Jet
 Keyboard/Plug Back/Plugs
 Screws tight Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By RMS

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 194
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm 0.144 (.139 - .169)
 36mm 0.164 (.156 - .190)
 53mm 0.234 (.228 - .278)
 103mm 0.503 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 26932

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	G11739	201603D 3/8/18
0.08	G8149	201601F 1/26/18
0.20	G11621	201604C 4/5/18
0.08 DGS	N/A	761626605 9/18/18 22805

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 FEB 14 2017
 FDLE Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By [Signature]

Barometric Pressure
 ID# 28421 Gauge 1013
 Instrument 1005

Mouth Alcohol Solution Lot # 2016-C
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	G2879
Interferent	G8144
0.05	G11739
0.08	G8149
0.20	G11621

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: QA/OC SP 2/14/17

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland
 Quality Control Review

2/14/17
 Date

Stability Checks # 80-001273 Pinellas County S.O. 2/14/17 *RMS*

RMS

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001273
02/14/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:38
Control Test	0.081	11:39
Air Blank	0.000	11:39
Control Test	0.081	11:40
Air Blank	0.000	11:40
Control Test	0.082	11:40
Air Blank	0.000	11:41
Control Test Stats		
Average	0.0813	
Std Dev	0.0006	
Rel Std Dev(%)	0.7099	

BSK SR

RMS
Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001273
02/14/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:34
Control Test	0.196	11:35
Air Blank	0.000	11:35
Control Test	0.196	11:36
Air Blank	0.000	11:37
Control Test	0.196	11:37
Air Blank	0.000	11:38
Control Test Stats		
Average	0.1960	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

RMS
Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001273
02/14/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:29
Control Test	0.080	11:30
Air Blank	0.000	11:30
Control Test	0.079	11:31
Air Blank	0.000	11:31
Control Test	0.079	11:32
Air Blank	0.000	11:33
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

RMS
Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001273
02/14/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:24
Control Test	0.050	11:25
Air Blank	0.000	11:25
Control Test	0.050	11:26
Air Blank	0.000	11:26
Control Test	0.051	11:27
Air Blank	0.000	11:28
Control Test Stats		
Average	0.0503	
Std Dev	0.0006	
Rel Std Dev(%)	1.1471	

RMS
Operator's Signature