



# INSTRUMENT PROCESSING SHEET

Agency Gilchrist County SOS/N 80-001285

Florida Department of Law Enforcement

Date In 11/1/2017DI Completion Date 11/13/17 Ship  P/U  H/D  CMI  EE

**Intake** Performed By TP

Annual  
 Registration  
 Return from CMI / EE

Visual Inspection:  
 Case  Handle  
 Keyboard  Dry Gas Shelf  
 Feet  Breath Tube  
 Ports  Screws Tight

Other Equipment/ Accessories:  
 Power cord  Printer Cable  
 Static Bag  12V DC Cable

Notes: \_\_\_\_\_

**Quality Checks** Performed By Q

Breath Tube Screen  
 Replace External O-Rings  
 Instrument Set Up Verified  
 R-Value 171  
 Flow Verification (L/s)  
 Flow Column # ATP105  
 32 mm 162 (.139 - .169)  
 36 mm 181 (.156 - .190)  
 53 mm 250 (.228 - .278)  
 103 mm 519 (.447 - .547)  
 Barometric Pressure Check  
 Gauge ID # 28421  
 Stability Checks

**Flow Calibration** Performed By \_\_\_\_\_

Flow Column # \_\_\_\_\_  
 5L/min - 17mm  
 15L/min - 53mm  
 30L/min - 103mm  
 R-Value \_\_\_\_\_  
 Post Calibration Verification (L/s)  
 Flow Column # \_\_\_\_\_  
 32 mm \_\_\_\_\_ (.139 - .169)  
 36 mm \_\_\_\_\_ (.156 - .190)  
 53 mm \_\_\_\_\_ (.228 - .278)  
 103 mm \_\_\_\_\_ (.447 - .547)

**Final Release Date**

**NOV 14 2017**

Simulator	Serial #	Lot #/Exp
0.050	<u>G2835</u>	201707D 07/25/2019
0.080	<u>DR1279</u>	201707E 07/25/2019
0.200	<u>SD1025</u>	201707C 07/24/2019
0.080 DGS	N/A	<u>AG619605</u> 7/14/18

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Temperature Checks** Performed By Q

Lab Temp °C 22.3  
 External Digital Therm. ID#: 300504  
 34°C +/-2 Serial #: G2835  
 34°C +/-2 Serial #: DR1279  
 34°C +/-2 Serial #: SD1025

**Calibration Adjustment** Performed By \_\_\_\_\_

Barometric Pressure Gauge ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

**Department Inspection** Performed By Q

Barometric Pressure ID# 28421  
 Gauge 1022 Instrument 1018  
 Mouth Alcohol Solution Lot # 2016-C  
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.000	<u>G2880</u>
Interferent	<u>G2870</u>
0.050	<u>G2835</u>
0.080	<u>DR1279</u>
0.200	<u>SD1025</u>

**Attachments**

Form 41  
 Stability Checks  
 Calibration Adjustment  
 Post-Stability Checks

Flow Calibration  
 Form 40  
 Other \_\_\_\_\_

Notes/Suggested Service: \_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use

SP 11/13/17 JJ Dahan 11/14/17  
 Tech Review / Date Admin Review / Date

80-001285  
 Stability Checks  
 11/13/17

INTOXILYZER 8000  
 Instrument Initialization  
 08:30 11/13/2017

GILCHRIST COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001285  
 11/13/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:06
Control Test	0.079	09:07
Air Blank	0.000	09:07
Control Test	0.079	09:08
Air Blank	0.000	09:08
Control Test	0.080	09:09
Air Blank	0.000	09:10
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

Operator's Signature

GILCHRIST COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001285  
 11/13/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:10
Control Test	0.049	09:11
Air Blank	0.000	09:12
Control Test	0.049	09:12
Air Blank	0.000	09:13
Control Test	0.049	09:14
Air Blank	0.000	09:14
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

GILCHRIST COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001285  
 11/13/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:15
Control Test	0.200	09:16
Air Blank	0.000	09:16
Control Test	0.199	09:17
Air Blank	0.000	09:18
Control Test	0.198	09:18
Air Blank	0.000	09:19
Control Test Stats		
Average	0.1990	
Std Dev	0.0010	
Rel Std Dev(%)	0.5025	

Operator's Signature

DGS

Operator's Signature

11/14/17  
 SP

SP